



ALCO-SENSOR IV WITH PRINTER MAINTENANCE REPORT

REPORT #7

Complete this report in duplicate at the time of the regular monthly preventative maintenance check, and whenever instrument is repaired. Send copy to Department of Health and Senior Services; retain original in department file.

ALCO SENSOR IV SN 108269	PRINTER SN 099.3586.809	DATE OF INSPECTION 03/04/2020
-----------------------------	----------------------------	----------------------------------

LOCATION OF INSTRUMENT (STREET AND CITY) Zone D-14 / Cassville, MO	TIME OF INSPECTION 1:25 pm
---	-------------------------------

CHECKLIST: Place a mark in the box by each item if found to be satisfactory or if operating within established limits. (Write in observed values where determined.) Unmarked items must be corrected before using instrument.

DIGITAL READOUT (ALL ELEMENTS OPERATIONAL)

TEMPERATURE OF ALCO SENSOR (10°C - 40°C)

PRINTER WORKING PROPERLY

TIME AND DATE DISPLAYING PROPERLY

BREATH ALCOHOL ACCURACY STANDARDS

SIMULATOR SOLUTION COMPRESSED ETHANOL-GAS MIXTURE

STANDARD SUPPLIER RepCo Marketing, Inc. LOT # 19001 EXP. DATE 03/12/2021

SIMULATOR TEMPERATURE (34°C ± 0.2°C) 34.02 SIMULATOR SN MP2467 SIMULATOR EXP DATE 01/10/2020

CALIBRATION CHECK - (ONLY ONE STANDARD IS TO BE USED PER MAINTENANCE REPORT)

Run three tests using a standard solution. All three tests must be within ±5% of the standard value and must have a spread of .005 or less. Check the box corresponding to the standard solution being used. (PRINTOUT ATTACHED)

- 0.100% STANDARD - MUST READ BETWEEN 0.095% and 0.105% INCLUSIVE
- 0.080% STANDARD - MUST READ BETWEEN 0.076% and 0.084% INCLUSIVE
- 0.040% STANDARD - MUST READ BETWEEN 0.038% and 0.042% INCLUSIVE

TEST 1 ← 0.101%

TEST 2 ← 0.101%

TEST 3 ← 101%

RFI DETECTOR OPERATING

**INDICATE THE NUMBER OF BREATH TESTS IN THE FOLLOWING RANGES SINCE THE LAST MAINTENANCE REPORT:
(DO NOT INCLUDE SELF-ADMINISTERED TESTS)**

REFUSALS	0	(0-.04)	0	(.05-.09)	0	(.10-.14)	0	(.15-.19)	0	(OVER .19)	0
----------	---	---------	---	-----------	---	-----------	---	-----------	---	------------	---

List any new parts and describe any alteration or modification that was made to restore the instrument to operate satisfactorily and within established limits (use other side if necessary).

Performed Calibration.

INSPECTING OFFICER

SIGNATURE *D.R. Hukill #383*

PRINT NAME
Tpr. D.R. Hukill / #383

TYPE II PERMIT NUMBER/EXPIRATION DATE
290043 / 02/26/2021

TELEPHONE NUMBER
(417) 895-6868

Return completed report to the: Breath Alcohol Program, MO Department of Health and Senior Services, Southeast District Office
2875 James Boulevard
Poplar Bluff, MO 63901

AS 10 Serial no: 108269
Version no: 532B

TEST RECORD 00527

Temp Date Time 210L

Air Blank:
03/04/20 14:04 .000
Calibration Check:
25 03/04/20 14:04 .101

Subject Name
Monthly Maint

Subject I.D.
Test # 1

Operator Name, I.D.
Tr. D. Hukill #383

Location
Zone D-14 Office

Cassville, MD

AS 10 Serial no: 108269
Version no: 532B

TEST RECORD 00528

Temp Date Time 210L

Air Blank:
03/04/20 14:07 .000
Calibration Check:
25 03/04/20 14:07 .101

Subject Name
Monthly Maint

Subject I.D.
Test # 2

Operator Name, I.D.
Tr. D. Hukill #383

Location
Zone D-14 Office

Cassville, MD

AS 10 Serial no: 108269
Version no: 532B

TEST RECORD 00529

Temp Date Time 210L

Air Blank:
03/04/20 14:10 .000
Calibration Check:
25 03/04/20 14:10 .101

Subject Name
Monthly Maint

Subject I.D.
Test # 3

Operator Name, I.D.
Tr. D. Hukill #383

Location
Zone D-14 Office

Cassville, MD

AS 10 Serial no: 108269
Version no: 532B

TEST RECORD 00530

Temp Date Time 210L

VOID: RFI
12 03/04/20 14:12

Subject Name
Monthly Maint

Subject I.D.
RFI Test

Operator Name, I.D.
Tr. D. Hukill #383

Location
Zone D-14 Office

Cassville, MD

CERTIFICATE OF ANALYSIS

MANUFACTURER AND SUPPLIER: RepCo Marketing Co.
LOT NUMBER: 19001
EXPIRATION DATE: March 12, 2021 at 11:59 p.m.

RepCo Marketing Co. certifies the following:


RepCo Marketing Co. prepared, tested and supplied Lot Number 19001 of Alcohol Certified Solution for simulators. This solution was manufactured and tested by RepCo Marketing Co., with confirmation by accredited institution, Data Resources Inc., using NIST standards. Random samples were analyzed by Data Resources Inc. utilizing a gas chromatograph and found to contain .1210 gms/dl +/- .003 gms/dl wt/vol ethanol (95% Confidence).

The alcohol and distilled water used in the solution were found to be free of any interfering substance.

This solution will produce a vapor alcohol value of .100 +/- 3% gms/210L breath when heated to 34 Degrees Celsius +/- 0.2 Degrees Celsius in a simulator (95% Confidence).

The date of manufacture for this lot number is March 13, 2019
The expiration date for this lot number is March 12, 2021 at 11:59 p.m.

This document is a true representation of the original Certificate of Analysis.


Alma Palmer, Operations Manager
RepCo Marketing Co.