



MISSOURI DEPARTMENT OF HEALTH AND SENIOR SERVICES
STATE PUBLIC HEALTH LABORATORY

ALCO-SENSOR IV WITH PRINTER MAINTENANCE REPORT

REPORT #7

Complete this report in duplicate at the time of the regular monthly preventative maintenance check, and whenever instrument is repaired. Send copy to Department of Health and Senior Services; retain original in department file.

ALCO SENSOR IV SN
108392

PRINTER SN
099.3586.814

DATE OF INSPECTION
01/01/2020

LOCATION OF INSTRUMENT (STREET AND CITY)
5490 5th St. Cottleville MO 63304

TIME OF INSPECTION
4:21 am

CHECKLIST: Place a mark in the box by each item if found to be satisfactory or if operating within established limits. (Write in observed values where determined.) Unmarked items must be corrected before using instrument.

☒ DIGITAL READOUT (ALL ELEMENTS OPERATIONAL)

☒ TEMPERATURE OF ALCO SENSOR (10°C - 40°C)

☒ PRINTER WORKING PROPERLY

☒ TIME AND DATE DISPLAYING PROPERLY

BREATH ALCOHOL ACCURACY STANDARDS

☒ SIMULATOR SOLUTION

☐ COMPRESSED ETHANOL-GAS MIXTURE

☒ STANDARD SUPPLIER Guth Laboratories LOT # 18200 EXP. DATE 07/03/2020

☒ SIMULATOR TEMPERATURE (34°C ± 0.2°C) 34C SIMULATOR SN MP4952 SIMULATOR EXP DATE 07/09/2020

☒ CALIBRATION CHECK - (ONLY ONE STANDARD IS TO BE USED PER MAINTENANCE REPORT)

Run three tests using a standard solution. All three tests must be within ±5% of the standard value and must have a spread of .005 or less. Check the box corresponding to the standard solution being used. (PRINTOUT ATTACHED)

☒ 0.100% STANDARD - MUST READ BETWEEN 0.095% and 0.105% INCLUSIVE

☐ 0.080% STANDARD - MUST READ BETWEEN 0.076% and 0.084% INCLUSIVE

☐ 0.040% STANDARD - MUST READ BETWEEN 0.038% and 0.042% INCLUSIVE

TEST 1 .104

TEST 2 .103

TEST 3 .102

☒ RFI DETECTOR OPERATING

INDICATE THE NUMBER OF BREATH TESTS IN THE FOLLOWING RANGES SINCE THE LAST MAINTENANCE REPORT:
(DO NOT INCLUDE SELF-ADMINISTERED TESTS)

REFUSALS 1 (0-.04) 0 (.05-.09) 0 (.10-.14) 0 (.15-.19) 0 (OVER .19)

List any new parts and describe any alteration or modification that was made to restore the instrument to operate satisfactorily and within established limits (use other side if necessary).

INSPECTING OFFICER

SIGNATURE

TYPE II PERMIT NUMBER/EXPIRATION DATE
2680089 02-16-2020

PRINT NAME

Richard Phillips

TELEPHONE NUMBER

(636) 498-6464

Return completed report to the: Breath Alcohol Program, MO Department of Health and Senior Services, Southeast District Office
2875 James Boulevard
Poplar Bluff, MO 63901

AS IV Serial no: 108392
Version no: 532B

TEST RECORD 00599

Temp Date Time 210L

Air Blank:
01/01/20 04:21 .000
Calibration Check:
21 01/01/20 04:21 .104

Subject Name

TEST 1
Subject I.D.

280089
Operator Name, I.D.
R PHILLIPS 468
Location

BOOKING ROOM

AS IV Serial no: 108392
Version no: 532B

TEST RECORD 00600

Temp Date Time 210L

Air Blank:
01/01/20 04:22 .000
Calibration Check:
23 01/01/20 04:22 .103

Subject Name

TEST 2
Subject I.D.

280089
Operator Name, I.D.
R PHILLIPS 468
Location

BOOKING ROOM

AS IV Serial no: 108392
Version no: 532B

TEST RECORD 00601

Temp Date Time 210L

Air Blank:
01/01/20 04:24 .000
Calibration Check:
24 01/01/20 04:24 .102

Subject Name

TEST 3
Subject I.D.

280089
Operator Name, I.D.
R PHILLIPS 468
Location

BOOKING ROOM

AS IV Serial no: 108392
Version no: 532B

TEST RECORD 00602

Temp Date Time 210L

VOID: RFI
12 01/01/20 04:25

Subject Name

R FI TEST
Subject I.D.

280089
Operator Name, I.D.
R PHILLIPS 468
Location
BOOKING ROOM

AS IV Serial no: 108392
Version no: 532B

TEST RECORD 00603

Temp Date Time 210L

Air Blank:
01/01/20 04:26 .000
Subject Test: Auto
25 01/01/20 04:26 .000

Subject Name

ZERO TEST
Subject I.D.

280089
Operator Name, I.D.
R PHILLIPS 468
Location
BOOKING ROOM

**GUTH LABORATORIES, INC.**

890 NORTH 6TH STREET • HARRISBURG, PA 17111-4511 • TELEPHONE: 717-564-5470

CERTIFICATE OF ANALYSIS**Certified Alcohol Reference Solution for Simulator**

Random Samples of Lot Number 18200 of Alcohol Reference Solution for Simulator were analyzed by gas chromatography on July 6, 2018, using a Perkin Elmer Gas Chromatograph Autosystem XL S/N: 610N9030209, and found to contain 0.1215% (w/vol) ethyl alcohol. The expiration date for this lot number is July 3, 2020 at 11:59 PM.

When used in a calibrated Simulator, operating at 34°C +/- .2°C, this solution will give a breath alcohol analysis instrument reading of 0.100 g/210L +/- 3%.

The alcohol and water used in this solution were free of test interfering substances.

Ted L. Pauley, President
GUTH LABORATORIES, INC.

NIST Traceability:

Testing was conducted using Cerilliant Reference Standard lot number FN04271602 whose values are traceable to NIST.

All balances are calibrated annually by an outside agency using NIST traceable weights. Calibration verification is done prior to each use utilizing NIST traceable weights.

**Missouri Department of Health and Senior Services**P.O. Box 570, Jefferson City, MO 65102-0570 Phone: 573-751-6400 FAX: 573-751-6010
RELAY MISSOURI for Hearing and Speech Impaired 1-800-735-2966 VOICE 1-800-735-2466**Randall W. Williams, MD, FACOG**
Director**Michael L. Parson**
Governor

SIMULATOR CERTIFICATION REPORT

SIMULATOR INFORMATION

Simulator Serial Number: MP4952 **Manufacturer:** Guth
Model Number: 12V500
Agency: COTTLEVILLE PD
Agency Address: 5490 FIFTH STREET, COTTLEVILLE, MO 63304

NIST THERMOMETER INFORMATION

Serial Number: 17KMM00689 **Bias:** 0.00
Uncertainty: 0.02
Date of Certification: 12/11/2018 **Date of Expiration:** 12/11/2019

ENVIRONMENTAL CONDITIONS

The environmental conditions during testing are within the tolerances of DHSS BAP method 3.

VERIFICATION RESULTS

<u>Simulator Average</u>	<u>NIST Average</u>	<u>Combined Uncertainty</u>
33.99	34.01	.04

The combined uncertainty is calculated with a k=2 value.

ADJUSTMENT RESULTS

No adjustment was needed.

Date of testing: 7/9/2019
Certification Expiration: 7/9/2020
Simulator testing technician: M. BOND

Notes on Condition: none

Deviation(s) from method: none

DHSS BAP Scientist Approving: B. LUTMER
Certification No: MP4952_792019

X *Brian Lutmer*

DHSS BAP Scientist Approving



STATE OF MISSOURI
DEPARTMENT OF HEALTH AND SENIOR SERVICES
BREATH ALCOHOL PROGRAM



PERMIT
TYPE II
RICHARD PHILLIPS

is hereby authorized to instruct and supervise operators, train instructors, inspect, calibrate, perform field service and repairs, and operate the following breath analyzer(s):

ALCO-SENSOR IV WITH PRINTER

for the determination of the alcoholic content of blood from a sample of expired air. Permit issued under the provisions of sections 577.020 through 577.041, RSMo and 306.111 through 306.119 RSMo.


DATE 2/16/2018

NUMBER 280089

EXPIRES 2/15/2020

MD 680-0771 (6-10)


DIRECTOR OF STATE PUBLIC HEALTH LABORATORY


DIRECTOR OF DEPARTMENT OF HEALTH AND SENIOR SERVICES

LAB-4 (96-10)

	STATE OF MISSOURI DEPARTMENT OF HEALTH AND SENIOR SERVICES BREATH ALCOHOL PROGRAM
	INSTRUMENT OPERATOR CARD
<small>The named cardholder is authorized to operate an evidential breath alcohol instrument for the determination of the alcoholic content in breath form of expired air in Missouri.</small>	
	
Operator PHILLIPS, RICHARD	
Permit No 280089	
Date Issued 2/16/2018 Date Expires 2/16/2020	