



MISSOURI DEPARTMENT OF HEALTH AND SENIOR SERVICES
STATE PUBLIC HEALTH LABORATORY
ALCO-SENSOR IV WITH PRINTER MAINTENANCE REPORT

REPORT #7

Complete this report in duplicate at the time of the regular monthly preventative maintenance check, and whenever instrument is repaired. Send copy to Department of Health and Senior Services; retain original in department file.

ALCO SENSOR IV SN 108392	PRINTER SN 099.3586.814	DATE OF INSPECTION 03-05-2020
LOCATION OF INSTRUMENT (STREET AND CITY) 5490 5th St. Cottleville MO 63304		TIME OF INSPECTION 4:38 am

CHECKLIST: Place a mark in the box by each item if found to be satisfactory or if operating within established limits. (Write in observed values where determined.) Unmarked items must be corrected before using instrument.

- DIGITAL READOUT (ALL ELEMENTS OPERATIONAL)
- TEMPERATURE OF ALCO SENSOR (10°C - 40°C)
- PRINTER WORKING PROPERLY
- TIME AND DATE DISPLAYING PROPERLY

BREATH ALCOHOL ACCURACY STANDARDS

- SIMULATOR SOLUTION COMPRESSED ETHANOL-GAS MIXTURE
- STANDARD SUPPLIER Guth Laboratories LOT # 18200 EXP. DATE 07/03/2020
- SIMULATOR TEMPERATURE (34°C ± 0.2°C) 34C SIMULATOR SN MP4952 SIMULATOR EXP DATE 07/09/2020

- CALIBRATION CHECK - (ONLY ONE STANDARD IS TO BE USED PER MAINTENANCE REPORT)
Run three tests using a standard solution. All three tests must be within ±5% of the standard value and must have a spread of .005 or less. Check the box corresponding to the standard solution being used. (PRINTOUT ATTACHED)
- 0.100% STANDARD - MUST READ BETWEEN 0.095% and 0.105% INCLUSIVE
- 0.080% STANDARD - MUST READ BETWEEN 0.076% and 0.084% INCLUSIVE
- 0.040% STANDARD - MUST READ BETWEEN 0.038% and 0.042% INCLUSIVE

TEST 1 ➡ .103	TEST 2 ➡ .102	TEST 3 ➡ .102
---------------	---------------	---------------

- RFI DETECTOR OPERATING

INDICATE THE NUMBER OF BREATH TESTS IN THE FOLLOWING RANGES SINCE THE LAST MAINTENANCE REPORT; (DO NOT INCLUDE SELF-ADMINISTERED TESTS)

REFUSALS	2	(0-.04)	0	(.05-.09)	0	(.10-.14)	0	(.15-.19)	0	(OVER .19)	0
----------	---	---------	---	-----------	---	-----------	---	-----------	---	------------	---

List any new parts and describe any alteration or modification that was made to restore the instrument to operate satisfactorily and within established limits (use other side if necessary).

INSPECTING OFFICER	
SIGNATURE 	PRINT NAME Richard Phillips
TYPE II PERMIT NUMBER/EXPIRATION DATE 2680089 02-18-2020	TELEPHONE NUMBER (636) 498-6464

Return completed report to the: Breath Alcohol Program, MO Department of Health and Senior Services, Southeast District Office
2876 James Boulevard
Poplar Bluff, MO 63901

AS IU Serial no: 108392
Version no: 532B

TEST RECORD 00615

Temp Date Time ^{9/} 210L

Air Blank:
03/06/20 04:38 .000
Calibration Check:
21 03/06/20 04:38 .103

Subject Name

TEST 1

Subject I.D.

200079

Operator Name, I.D.

RICHARD PHILLIPS 408

Location

BOOKING ROOM

AS IU Serial no: 108392
Version no: 532B

TEST RECORD 00616

Temp Date Time ^{9/} 210L

Air Blank:
03/06/20 04:39 .000
Calibration Check:
23 03/06/20 04:39 .102

Subject Name

TEST 2

Subject I.D.

200079

Operator Name, I.D.

RICHARD PHILLIPS 408

Location

BOOKING ROOM

AS IU Serial no: 108392
Version no: 532B

TEST RECORD 00617

Temp Date Time ^{9/} 210L

Air Blank:
03/06/20 04:43 .000
Calibration Check:
24 03/06/20 04:43 .102

Subject Name

TEST 3

Subject I.D.

200079

Operator Name, I.D.

RICHARD PHILLIPS 408

Location

BOOKING ROOM

AS IU Serial no: 108392
Version no: 532B

TEST RECORD - REPRINT

TEST RECORD 00618

Temp Date Time ^{9/} 210L

VOID: RFI
12 03/06/20 04:49

Subject Name

RFI TEST

Subject I.D.

200079

Operator Name, I.D.

RICHARD PHILLIPS 408

Location

BOOKING ROOM

AS IU Serial no: 108392
Version no: 532B

TEST RECORD 00619

Temp Date Time ^{9/} 210L

Air Blank:
03/06/20 04:51 .000
Subject Test: Auto
25 03/06/20 04:51 .000

Subject Name

ZEZO TEST

Subject I.D.

200079

Operator Name, I.D.

RICHARD PHILLIPS 408

Location

BOOKING ROOM

**GUTH LABORATORIES, INC.**

830 NORTH 67th STREET • HARRISBURG, PA 17111-4811 • TELEPHONE: 717-664-6470

CERTIFICATE OF ANALYSIS**Certified Alcohol Reference Solution for Simulator**

Random Samples of Lot Number 18200 of Alcohol Reference Solution for Simulator were analyzed by gas chromatography on July 6, 2018, using a Perkin Elmer Gas Chromatograph Autosystem XL S/N: 610N9030209, and found to contain 0.1215% (w/vol) ethyl alcohol. The expiration date for this lot number is July 3, 2020 at 11:59 PM.

When used in a calibrated Simulator, operating at 34°C +/- .2°C, this solution will give a breath alcohol analysis instrument reading of 0.100 g/210L +/- 3%.

The alcohol and water used in this solution were free of test interfering substances.

Ted L. Pauley, President
GUTH LABORATORIES, INC.

NIST Traceability:

Testing was conducted using Cerilliant Reference Standard lot number FN04271602 whose values are traceable to NIST.

All balances are calibrated annually by an outside agency using NIST traceable weights. Calibration verification is done prior to each use utilizing NIST traceable weights.



Missouri Department of Health and Senior Services
P.O. Box 570, Jefferson City, MO 65102-0570 Phone: 573-751-6400 FAX: 573-751-6010
RELAY MISSOURI for Hearing and Speech Impaired 1-800-735-2966 VOICE 1-800-735-2466
Randall W. Williams, MD, FACOG
Director



Michael L. Parson
Governor

SIMULATOR CERTIFICATION REPORT

SIMULATOR INFORMATION

Simulator Serial Number: MP4952 Manufacturer: Guth
Model Number: 12V500
Agency: COTTLEVILLE PD
Agency Address: 5490 FIFTH STREET, COTTLEVILLE, MO 63304

NIST THERMOMETER INFORMATION

Serial Number: 17KMM00689 Bias: 0.00
Uncertainty: 0.02
Date of Certification: 12/11/2018 Date of Expiration: 12/11/2019

ENVIRONMENTAL CONDITIONS

The environmental conditions during testing are within the tolerances of DHSS BAP method 3.

VERIFICATION RESULTS

Table with 3 columns: Simulator Average (33.99), NIST Average (34.01), Combined Uncertainty (.04)

The combined uncertainty is calculated with a k=2 value.

ADJUSTMENT RESULTS

No adjustment was needed.

Date of testing: 7/9/2019
Certification Expiration: 7/9/2020
Simulator testing technician: M. BOND

Notes on Condition: none

Deviation(s) from method: none

DHSS BAP Scientist Approving: B. LUTMER
Certification No: MP4952_792019

Handwritten signature of Brian Lutmer

DHSS BAP Scientist Approving



STATE OF MISSOURI
DEPARTMENT OF HEALTH AND SENIOR SERVICES
BREATH ALCOHOL PROGRAM



PERMIT
TYPE II

RICHARD PHILLIPS

Is hereby authorized to instruct and supervise operators, train instructors, inspect, calibrate, perform field service and repairs, and operate the following breath analyzer(s):

ALCO-SENSOR IV WITH PRINTER

for the determination of the alcoholic content of blood from a sample of expired air. Permit issued under the provisions of sections 577.020 through 577.041, RSMo and 306.111 through 306.119 RSMo.

DATE 2/6/2020

NUMBER 200079

EXPIRES 2/6/2022

MO 580-0771 (6-10)

DIRECTOR OF STATE PUBLIC HEALTH LABORATORY

DIRECTOR OF DEPARTMENT OF HEALTH AND SENIOR SERVICES

LAB-4 (18-10)

STATE OF MISSOURI
DEPARTMENT OF HEALTH AND SENIOR SERVICES
BREATH ALCOHOL PROGRAM

INSTRUMENT OPERATOR CARD

The named cardholder is authorized to operate an evidential breath alcohol instrument for the determination of the alcoholic content in breath form of expired air in Missouri.

Operator PHILLIPS, RICHARD
Permit No 200079
Date Issued 2/6/2020 Date Expires 2/6/2022