



**RECEIVED**

By Stephen Wilson at 10:34 am, May 05, 2020

**ALCO-SENSOR IV WITH PRINTER MAINTENANCE REPORT**

REPORT #7

Complete this report in duplicate at the time of the regular monthly preventative maintenance check, and whenever instrument is repaired. Send copy to Department of Health and Senior Services; retain original in department file.

ALCO SENSOR IV SN 108393	PRINTER SN 099.3586.801	DATE OF INSPECTION 05/05/2020
LOCATION OF INSTRUMENT (STREET AND CITY) 14301 SOUTH OUTER 40, CHESTERFIELD		TIME OF INSPECTION 9:25 am

**CHECKLIST:** Place a mark in the box by each item if found to be satisfactory or if operating within established limits. (Write in observed values where determined.) Unmarked items must be corrected before using instrument.

DIGITAL READOUT (ALL ELEMENTS OPERATIONAL)

TEMPERATURE OF ALCO SENSOR (10°C - 40°C)

PRINTER WORKING PROPERLY

TIME AND DATE DISPLAYING PROPERLY

**BREATH ALCOHOL ACCURACY STANDARDS**

SIMULATOR SOLUTION  COMPRESSED ETHANOL-GAS MIXTURE

STANDARD SUPPLIER INTOXIMETERS, INC. LOT # AG829708 EXP. DATE 10/24/2020

SIMULATOR TEMPERATURE (34°C ± 0.2°C) \_\_\_\_\_ SIMULATOR SN \_\_\_\_\_ SIMULATOR EXP DATE \_\_\_\_\_

CALIBRATION CHECK – (ONLY ONE STANDARD IS TO BE USED PER MAINTENANCE REPORT)  
Run three tests using a standard solution. All three tests must be within ±5% of the standard value and must have a spread of .005 or less. Check the box corresponding to the standard solution being used. (PRINTOUT ATTACHED)

- 0.100% STANDARD - MUST READ BETWEEN 0.095% and 0.105% INCLUSIVE
- 0.080% STANDARD - MUST READ BETWEEN 0.076% and 0.084% INCLUSIVE
- 0.040% STANDARD - MUST READ BETWEEN 0.038% and 0.042% INCLUSIVE

TEST 1  .098	TEST 2  .097	TEST 3  .096
--------------	--------------	--------------

RFI DETECTOR OPERATING

**INDICATE THE NUMBER OF BREATH TESTS IN THE FOLLOWING RANGES SINCE THE LAST MAINTENANCE REPORT: (DO NOT INCLUDE SELF-ADMINISTERED TESTS)**

REFUSALS	1	(0-.04)	0	(.05-.09)	0	(.10-.14)	0	(.15-.19)	0	(OVER .19)	0
----------	---	---------	---	-----------	---	-----------	---	-----------	---	------------	---

List any new parts and describe any alteration or modification that was made to restore the instrument to operate satisfactorily and within established limits (use other side if necessary).

Calibration Check #2 was reprint due to paper replacement and printer ribbon error.

**INSPECTING OFFICER**

SIGNATURE	PRINT NAME PO JAHNS, DSN 3725
TYPE II PERMIT NUMBER/EXPIRATION DATE 290030 02/20/2021	TELEPHONE NUMBER (636) 529-8210

**Return completed report to the:** Breath Alcohol Program, MO Department of Health and Senior Services, Southeast District Office  
2875 James Boulevard  
Poplar Bluff, MO 63901

AS IV Serial no: 188393  
Version no: 532B

TEST RECORD 00541

Temp Date Time 210L <sup>s/</sup>

Air Blank:  
05/05/20 09:25 .000  
Calibration Check:  
21 05/05/20 09:26 .099

Subject Name  
*Cal Check #1*

Subject I.D.  
*NIA*

Operator Name, I.D.  
*Jahns 325*

Location  
*MoDOT TME*

AS IV Serial no: 188393  
Version no: 532B

TEST RECORD - REPRINT

TEST RECORD 00542

Temp Date Time 210L <sup>s/</sup>

Air Blank:  
05/05/20 09:30 .003  
Calibration Check:  
22 05/05/20 09:30 .097

Subject Name  
*Cal Check #2*

Subject I.D.  
*NIA*

Operator Name, I.D.  
*Jahns 325*

Location  
*MoDOT TME*

AS IV Serial no: 188393  
Version no: 532B

TEST RECORD 00543

Temp Date Time 210L <sup>s/</sup>

Air Blank:  
05/05/20 09:34 .000  
Calibration Check:  
24 05/05/20 09:34 .096

Subject Name  
*Cal Check #3*

Subject I.D.  
*NIA*

Operator Name, I.D.  
*Jahns 325*

Location  
*MoDOT TME*

AS IV Serial no: 188393  
Version no: 532B

TEST RECORD 00544

Temp Date Time 210L <sup>s/</sup>

VOID: RFI  
12 05/05/20 09:36

Subject Name  
*RFI Check*

Subject I.D.  
*NIA*

Operator Name, I.D.  
*Jahns 325*

Location  
*MoDOT TME*



Airgas USA LLC (LAB)

3500 Bernard Street

St. Louis, Mo. 63103

Ph: (314) 533-3100

Fax: (314) 533-7328

## Certificate of Analysis

**Customer Name**

*Exclusive Supplier*

Intoximeters, Inc.

2081 Craig Road

St. Louis, Mo 63146

**Test Date:** 25-Oct-2018

**Lot # AG829708 Model 108cacc**

**Exp. Date**

24-Oct-2020

**Cyl. Type**

108

**Component**

Ethanol

Nitrogen

**Certified Concentration**

0.100 ± 2% BrAC (272 ppm)

Balance

**Certification Traceable to N.I.S.T. RGM Ethanol Standards:**

**Serial No.**

EB0010581

EB0010570

EB0010285

EB0010561

EB0010681

**Concentration**

392.1 ppm

259.8 ppm

208.0 ppm

103.6 ppm

52.12 ppm

**Serial No.**

EB0010603

EB0010559

EB0010595

EB0010562

EB0010579

**Concentration**

393.0 ppm

258.2 ppm

208.3 ppm

104.2 ppm

52.81 ppm

**Analytical Method:**

NDIR

Digitally signed by Quality Control  
Date: 2018.10.25 14:13:39 -05:00  
Reason: Dry gas standard certification of analysis  
Location: Airgas USA LLC (Lab)

Approved for Release: \_\_\_\_\_

Rod Marsala

**ISO 17025:2005 A2LA accredited. Certificate Number 3082.06**



STATE OF MISSOURI  
 DEPARTMENT OF HEALTH AND SENIOR SERVICES  
 BREATH ALCOHOL PROGRAM

2

**PERMIT**  
**TYPE II**  
**KYLE JAHNS**

Is hereby authorized to instruct and supervise operators, train instructors, inspect, calibrate, perform field service and repairs, and operate the following breath analyzer(s):

**ALCO-SENSOR IV WITH PRINTER, INTOX DMT**

for the determination of the alcoholic content of blood from a sample of expired air. Permit issued under the provisions of sections 577.020 through 577.041, RSMo and 306.111 through 306.119 RSMo.

DATE 2/20/2019

NUMBER 290039

EXPIRES 2/20/2021

DIRECTOR OF STATE PUBLIC HEALTH LABORATORY

DIRECTOR OF DEPARTMENT OF HEALTH AND SENIOR SERVICES