



MISSOURI DEPARTMENT OF HEALTH AND SENIOR SERVICES
 STATE PUBLIC HEALTH LABORATORY
ALCO-SENSOR IV WITH PRINTER MAINTENANCE REPORT

RECEIVED
 By Tracy Crews at 11:34 am, Dec 21, 2020

REPORT #7

Complete this report in duplicate at the time of the regular monthly preventative maintenance check, and whenever instrument is repaired. Send copy to Department of Health and Senior Services; retain original in department file.

ALCO SENSOR IV SN 109482	PRINTER SN 95.1111.053	DATE OF INSPECTION 12/17/2020
LOCATION OF INSTRUMENT (STREET AND CITY) 501 Faraon Street Saint Joseph MO. 64501		TIME OF INSPECTION 1:21 pm

CHECKLIST: Place a mark in the box by each item if found to be satisfactory or if operating within established limits. (Write in observed values where determined.) Unmarked items must be corrected before using instrument.

DIGITAL READOUT (ALL ELEMENTS OPERATIONAL)

TEMPERATURE OF ALCO SENSOR (10°C - 40°C)

PRINTER WORKING PROPERLY

TIME AND DATE DISPLAYING PROPERLY

BREATH ALCOHOL ACCURACY STANDARDS

SIMULATOR SOLUTION COMPRESSED ETHANOL-GAS MIXTURE

STANDARD SUPPLIER Intoximeters LOT # AG836201 EXP. DATE 12/28/2020

SIMULATOR TEMPERATURE (34°C ± 0.2°C) _____ SIMULATOR SN _____ SIMULATOR EXP DATE _____

CALIBRATION CHECK – (ONLY ONE STANDARD IS TO BE USED PER MAINTENANCE REPORT)
 Run three tests using a standard solution. All three tests must be within ±5% of the standard value and must have a spread of .005 or less. Check the box corresponding to the standard solution being used. (PRINTOUT ATTACHED)

- 0.100% STANDARD - MUST READ BETWEEN 0.095% and 0.105% INCLUSIVE
- 0.080% STANDARD - MUST READ BETWEEN 0.076% and 0.084% INCLUSIVE
- 0.040% STANDARD - MUST READ BETWEEN 0.038% and 0.042% INCLUSIVE

TEST 1 → 0.097

TEST 2 → 0.097

TEST 3 → 0.097

RFI DETECTOR OPERATING

**INDICATE THE NUMBER OF BREATH TESTS IN THE FOLLOWING RANGES SINCE THE LAST MAINTENANCE REPORT:
 (DO NOT INCLUDE SELF-ADMINISTERED TESTS)**

REFUSALS	(0-.04)	(.05-.09)	(.10-.14)	(.15-.19)	(OVER .19)
----------	---------	-----------	-----------	-----------	------------

List any new parts and describe any alteration or modification that was made to restore the instrument to operate satisfactorily and within established limits (use other side if necessary).

INSPECTING OFFICER

SIGNATURE 	PRINT NAME John L. Foster
TYPE II PERMIT NUMBER/EXPIRATION DATE 290227 EXP-10/01/2021	TELEPHONE NUMBER (816) 596-8206

Return completed report to the: Breath Alcohol Program, MO Department of Health and Senior Services, Southeast District Office
 2875 James Boulevard
 Poplar Bluff, MO 63901

SAINT JOSEPH POLICE DEPARTMENT MONTHLY MAINTENANCE REPORT

ASIV-W/PRINTER

AS IV Serial no: 109482
 version no: 532B

TEST RECORD 01092
 s/
 Date Time 210L

Air Blank: 12/17/20 13:21 .000
 Calibration Check: 27 12/17/20 13:21 .097

Subject Name
Monthly Maintenance

Subject I.D.
Foster John

Operator Name, I.D.
601 Farson St.

Location
St. Joseph Mo

64101 LEC

AS IV Serial no: 109482
 version no: 532B

TEST RECORD 01093
 s/
 Date Time 210L

Air Blank: 12/17/20 13:24 .000
 Calibration Check: 27 12/17/20 13:24 .097

Subject Name
Monthly

Subject I.D.
Foster John

Operator Name, I.D.

Location
LEC

AS IV Serial no: 109482
 version no: 532B

TEST RECORD 01094
 s/
 Date Time 210L

Air Blank: 12/17/20 13:25 .000
 Calibration Check: 27 12/17/20 13:25 .097

Subject Name
Monthly

Subject I.D.
Foster

Operator Name, I.D.

Location
LEC

AS IV Serial no: 109482
 version no: 532B

TEST RECORD 01095
 s/
 Date Time 210L

Air Blank: 12/17/20 13:26 .000
 Calibration Check: 27 12/17/20 13:26 .097

Subject Name
Monthly

Subject I.D.
Foster

Operator Name, I.D.

Location
LEC



Airgas USA LLC (LAB)
 3500 Bernard Street
 St. Louis, Mo. 63103
 Ph: (314) 533-3100
 Fax: (314) 533-7328

Certificate of Analysis

Customer Name
 Exclusive Supplier
 Intoximeters, Inc.
 2081 Craig Road
 St. Louis, Mo 63146

Test Date: 2-Jan-2019

Lot # AG836201 Model 108cacd

Exp. Date 28-Dec-2020	Cyl. Type 108	Component Ethanol Nitrogen	Certified Concentration 0.100 ± 2% BrAC (272 ppm) Balance
---------------------------------	-------------------------	---	--

Certification Traceable to N.I.S.T. RGM and to CRM Ethanol Standards:

<u>RGM Serial No.</u> EB0010581	<u>Concentration</u> 392.1 ppm	<u>RGM Serial No.</u> EB0010603	<u>Concentration</u> 393.0 ppm
EB0010570	259.8 ppm	EB0010559	258.2 ppm
EB0010285	208.0 ppm	EB0010595	208.3 ppm
EB0010561	103.6 ppm	EB0010562	104.2 ppm
EB0010681	52.12 ppm	EB0010579	52.81 ppm
<u>CRM Serial No.</u> CC434668	<u>Concentration</u> 800.0 ppm	<u>CRM Serial No.</u> 0056649	<u>Concentration</u> 390.1 ppm
CC234503	253.0 ppm	0056662	150.2 ppm

Analytical Method: NDIR

Digitally signed by Quality Control
 Date: 2016.01.02 18:04:08 -08:00
 Reason: Dry gas standard certification of analysis
 Location: Airgas USA LLC (Lab)

Approved for Release:

 Rod Marsala

ISO 17025:2005 A2LA accredited. Certificate Number 3082.06
ISO 17034:2016 A2LA accredited. Certificate Number 3082.07



STATE OF MISSOURI
 DEPARTMENT OF HEALTH AND SENIOR SERVICES
 BREATH ALCOHOL PROGRAM

2

PERMIT
 TYPE II

JOHN L. FOSTER

is hereby authorized to instruct and supervise operators, train instructors, inspect, calibrate, perform field service and repairs, and operate the following breath analyzer(s):

ALCO-SENSOR IV WITH PRINTER, INTOX EC/IR II

for the determination of the alcoholic content of blood from a sample of expired air. Permit issued under the provisions of sections 577.020 through 577.041, RSMo and 306.111 through 306.119 RSMo.

DATE 10/1/2019

NUMBER 290227

EXPIRES 10/1/2021

MO 580-0771 (6-10)

DIRECTOR OF STATE PUBLIC HEALTH LABORATORY

DIRECTOR OF DEPARTMENT OF HEALTH AND SENIOR SERVICES

LAB-4 (R6-10)



STATE OF MISSOURI
 DEPARTMENT OF HEALTH AND SENIOR SERVICES
 BREATH ALCOHOL PROGRAM

INSTRUMENT OPERATOR CARD

The named cardholder is authorized to operate an evidential breath alcohol instrument for the determination of the alcoholic content in breath form of expired air in Missouri.

Operator FOSTER, JOHN
Permit No 290227
Date Issued 10/1/2019 **Date Expires** 10/1/2021

