



MISSOURI DEPARTMENT OF HEALTH AND SENIOR SERVICES
STATE PUBLIC HEALTH LABORATORY
ALCO-SENSOR IV WITH PRINTER MAINTENANCE REPORT

RECEIVED
By Tracy Crews at 11:00 am, Feb 25, 2020

REPORT #7

Complete this report in duplicate at the time of the regular monthly preventative maintenance check, and whenever instrument is repaired. Send copy to Department of Health and Senior Services; retain original in department file.

ALCO SENSOR IV SN 111326	PRINTER SN 099.3586.615	DATE OF INSPECTION 02/20/2020
-----------------------------	----------------------------	----------------------------------

LOCATION OF INSTRUMENT (STREET AND CITY) Richland PD 201 South Chestnut Street Richland Missouri 65556	TIME OF INSPECTION 11:37 am
---	--------------------------------

CHECKLIST: Place a mark in the box by each item if found to be satisfactory or if operating within established limits. (Write in observed values where determined.) Unmarked items must be corrected before using instrument.

DIGITAL READOUT (ALL ELEMENTS OPERATIONAL)

TEMPERATURE OF ALCO SENSOR (10°C - 40°C)

PRINTER WORKING PROPERLY

TIME AND DATE DISPLAYING PROPERLY

BREATH ALCOHOL ACCURACY STANDARDS

SIMULATOR SOLUTION COMPRESSED ETHANOL-GAS MIXTURE

STANDARD SUPPLIER Intoximeters LOT # AG912204 EXP. DATE 05/02/2021

SIMULATOR TEMPERATURE (34°C ± 0.2°C) SIMULATOR SN SIMULATOR EXP DATE

CALIBRATION CHECK - (ONLY ONE STANDARD IS TO BE USED PER MAINTENANCE REPORT)

Run three tests using a standard solution. All three tests must be within ±5% of the standard value and must have a spread of .005 or less. Check the box corresponding to the standard solution being used. (PRINTOUT ATTACHED)

- 0.100% STANDARD - MUST READ BETWEEN 0.095% and 0.105% INCLUSIVE
- 0.080% STANDARD - MUST READ BETWEEN 0.076% and 0.084% INCLUSIVE
- 0.040% STANDARD - MUST READ BETWEEN 0.038% and 0.042% INCLUSIVE

TEST 1 ➔ .104

TEST 2 ➔ .104

TEST 3 ➔ .103

RFI DETECTOR OPERATING

INDICATE THE NUMBER OF BREATH TESTS IN THE FOLLOWING RANGES SINCE THE LAST MAINTENANCE REPORT:
(DO NOT INCLUDE SELF-ADMINISTERED TESTS)

REFUSALS	0	(0-.04)	0	(.05-.09)	0	(.10-.14)	0	(.15-.19)	0	(OVER .19)	0
----------	---	---------	---	-----------	---	-----------	---	-----------	---	------------	---

List any new parts and describe any alteration or modification that was made to restore the instrument to operate satisfactorily and within established limits (use other side if necessary).

Instrument is operating within established limits.

INSPECTING OFFICER

SIGNATURE
David Moser

PRINT NAME
David Moser

TYPE II PERMIT NUMBER/EXPIRATION DATE
280293 10/4/20

TELEPHONE NUMBER
(573) 480-4144

Return completed report to the: Breath Alcohol Program, MO Department of Health and Senior Services, Southeast District Office
2875 James Boulevard
Poplar Bluff, MO 63901

AS IV Serial no: 111326
Version no: 532B

TEST RECORD 00500

Temp Date Time ^{s/} 210L

Air Blank:
02/20/20 11:37 .000
Calibration Check:
2) 02/20/20 11:37 .000

Subject Name

BLANK TEST
Subject I.D.

Operator Name, I.D.

DMOSER 280293
Location
RICHMOND PD

AS IV Serial no: 111326
Version no: 532B

TEST RECORD 00501

Temp Date Time ^{s/} 210L

Air Blank:
02/20/20 11:39 .000
Subject Test: Man
2) 02/20/20 11:39 .104

Subject Name

TEST 1
Subject I.D.

Operator Name, I.D.

DMOSER 280293
Location
RICHMOND PD

AS IV Serial no: 111326
Version no: 532B

TEST RECORD 00502

Temp Date Time ^{s/} 210L

Air Blank:
02/20/20 11:40 .000
Subject Test: Man
2) 02/20/20 11:40 .104

Subject Name

TEST 2
Subject I.D.

Operator Name, I.D.

DMOSER 280293
Location
RICHMOND PD

AS IV Serial no: 111326
Version no: 532B

TEST RECORD 00503

Temp Date Time ^{s/} 210L

Air Blank:
02/20/20 11:42 .000
Subject Test: Man
2) 02/20/20 11:42 .103

Subject Name

TEST 3
Subject I.D.

Operator Name, I.D.

DMOSER 280293
Location
RICHMOND PD

AS IV Serial no: 111326
Version no: 532B

TEST RECORD 00504

Temp Date Time ^{s/} 210L

VOID: RFI
12 02/20/20 11:44

Subject Name

RFI TEST
Subject I.D.

Operator Name, I.D.

DMOSER 280293
Location
RICHMOND PD



STATE OF MISSOURI
DEPARTMENT OF HEALTH AND SENIOR SERVICES
BREATH ALCOHOL PROGRAM



**PERMIT
TYPE II**

DAVID L. MOSER

is hereby authorized to instruct and supervise operators, train instructors, inspect, calibrate, perform field service and repairs, and operate the following breath analyzer(s):

ALCO-SENSOR IV WITH PRINTER

for the determination of the alcoholic content of blood from a sample of expired air. Permit issued under the provisions of sections 577.020 through 577.041, RSMo and 306.111 through 306.119 RSMo.

DATE 10/4/2018

NUMBER 280293

EXPIRES 10/4/2020

DIRECTOR OF STATE PUBLIC HEALTH LABORATORY

DIRECTOR OF DEPARTMENT OF HEALTH AND SENIOR SERVICES

LAJ 4 (R8-10)

MO 580-0771 (5-10)

STATE OF MISSOURI
DEPARTMENT OF HEALTH AND SENIOR SERVICES
BREATH ALCOHOL PROGRAM

INSTRUMENT OPERATOR CARD

The named cardholder is authorized to operate an evidential breath alcohol instrument for the determination of the alcoholic content in breath form of expired air in Missouri.

Operator MOSER, DAVID
Permit No 280293
Date Issued 10/4/2018 Date Expires 10/4/2020