



ALCO-SENSOR IV WITH PRINTER MAINTENANCE REPORT

REPORT #7

Complete this report in duplicate at the time of the regular monthly preventative maintenance check, and whenever instrument is repaired. Send copy to Department of Health and Senior Services; retain original in department file.

ALCO SENSOR IV SN 111327 Lawson PD	PRINTER SN 099.3586.578	DATE OF INSPECTION 04/20/2020
LOCATION OF INSTRUMENT (STREET AND CITY) 104 W 3rd Street, Lawson, MO		TIME OF INSPECTION 2:46 pm

CHECKLIST: Place a mark in the box by each item if found to be satisfactory or if operating within established limits. (Write in observed values where determined.) Unmarked items must be corrected before using instrument.

DIGITAL READOUT (ALL ELEMENTS OPERATIONAL)

TEMPERATURE OF ALCO SENSOR (10°C - 40°C)

PRINTER WORKING PROPERLY

TIME AND DATE DISPLAYING PROPERLY

BREATH ALCOHOL ACCURACY STANDARDS

SIMULATOR SOLUTION COMPRESSED ETHANOL-GAS MIXTURE

STANDARD SUPPLIER Repco Marketing LOT # 19001 EXP. DATE 03/12/2021

SIMULATOR TEMPERATURE (34°C ± 0.2°C) 34.0 SIMULATOR SN DR5392 SIMULATOR EXP DATE 01/22/2021

CALIBRATION CHECK – (ONLY ONE STANDARD IS TO BE USED PER MAINTENANCE REPORT)

Run three tests using a standard solution. All three tests must be within ±5% of the standard value and must have a spread of .005 or less. Check the box corresponding to the standard solution being used. (PRINTOUT ATTACHED)

- 0.100% STANDARD - MUST READ BETWEEN 0.095% and 0.105% INCLUSIVE
- 0.080% STANDARD - MUST READ BETWEEN 0.076% and 0.084% INCLUSIVE
- 0.040% STANDARD - MUST READ BETWEEN 0.038% and 0.042% INCLUSIVE

TEST 1 ➔ .099

TEST 2 ➔ .098

TEST 3 ➔ .098

RFI DETECTOR OPERATING

INDICATE THE NUMBER OF BREATH TESTS IN THE FOLLOWING RANGES SINCE THE LAST MAINTENANCE REPORT: (DO NOT INCLUDE SELF-ADMINISTERED TESTS)

REFUSALS	0	(0-.04)	0	(.05-.09)	0	(.10-.14)	0	(.15-.19)	0	(OVER .19)	0
----------	---	---------	---	-----------	---	-----------	---	-----------	---	------------	---

List any new parts and describe any alteration or modification that was made to restore the instrument to operate satisfactorily and within established limits (use other side if necessary).

Corrected time-slow by three minutes

INSPECTING OFFICER

SIGNATURE
Chief Bruce Summa

PRINT NAME
Chief Bruce Summa

TYPE II PERMIT NUMBER/EXPIRATION DATE
200106/02-27-2022

TELEPHONE NUMBER
(816) 580-7210

Return completed report to the: Breath Alcohol Program, MO Department of Health and Senior Services, Southeast District Office
2875 James Boulevard
Poplar Bluff, MO 63901



STATE OF MISSOURI
 DEPARTMENT OF HEALTH AND SENIOR SERVICES
 BREATH ALCOHOL PROGRAM



PERMIT
TYPE II
BRUCE SUMMA

is hereby authorized to instruct and supervise operators, train instructors, inspect, calibrate, perform field service and repairs, and operate the following breath analyzer(s):

ALCO-SENSOR IV WITH PRINTER

for the determination of the alcoholic content of blood from a sample of expired air. Permit issued under the provisions of sections 577.020 through 577.041, RSMo and 306.111 through 306.119 RSMo.

DATE 2/27/2020

NUMBER 200106

EXPIRES 2/27/2022

DIRECTOR OF STATE PUBLIC HEALTH LABORATORY

DIRECTOR OF DEPARTMENT OF HEALTH AND SENIOR SERVICES



STATE OF MISSOURI
 DEPARTMENT OF HEALTH AND SENIOR SERVICES
 BREATH ALCOHOL PROGRAM

INSTRUMENT OPERATOR CARD

The named cardholder is authorized to operate an evidential breath alcohol instrument for the determination of the alcoholic content in breath form of expired air in Missouri.

Operator SUMMA, BRUCE
Permit No 200106
Date Issued 2/27/2020 **Date Expires** 2/27/2022



CERTIFICATE OF ANALYSIS**MANUFACTURER AND SUPPLIER: RepCo Marketing Co.****LOT NUMBER: 19001****EXPIRATION DATE: March 12, 2021 at 11:59 p.m.**

RepCo Marketing Co. certifies the following:


RepCo Marketing Co. prepared, tested and supplied Lot Number 19001 of Alcohol Certified Solution for simulators. This solution was manufactured and tested by RepCo Marketing Co., with confirmation by accredited institution, Data Resources Inc., using NIST standards. Random samples were analyzed by Data Resources Inc. utilizing a gas chromatograph and found to contain .1210 gms/dl +/- .003 gms/dl wt/vol ethanol (95% Confidence).

The alcohol and distilled water used in the solution were found to be free of any interfering substance.

This solution will produce a vapor alcohol value of .100 +/- 3% gms/210L breath when heated to 34 Degrees Celsius +/- 0.2 Degrees Celsius in a simulator (95% Confidence).

The date of manufacture for this lot number is March 13, 2019
The expiration date for this lot number is March 12, 2021 at 11:59 p.m.

This document is a true representation of the original Certificate of Analysis.



Alma Palmer, Operations Manager
RepCo Marketing Co.

AS IV Serial no: 111327
Version no: 532B

TEST RECORD 00597

Temp Date Time ^{9/} 210L

Air Blank:
04/20/20 14:46 .000
Calibration Check:
23 04/20/20 14:46 .099

Subject Name

Monthly Maint
Subject I.D.

Chief Bruce Summa
Operator Name, I.D.

200106
Location

Lawson PD

Test 1

AS IV Serial no: 111327
Version no: 532B

TEST RECORD 00598

Temp Date Time ^{9/} 210L

Air Blank:
04/20/20 14:48 .000
Subject Test: Man
24 04/20/20 14:48 .098

Subject Name

Monthly Maint
Subject I.D.

Chief Bruce Summa
Operator Name, I.D.

200106
Location

Lawson PD

Test 2

AS IV Serial no: 111327
Version no: 532B

TEST RECORD 00599

Temp Date Time ^{9/} 210L

Air Blank:
04/20/20 14:49 .000
Subject Test: Man
24 04/20/20 14:49 .098

Subject Name

Monthly Maint
Subject I.D.

Chief Bruce Summa
Operator Name, I.D.

200106
Location

Lawson PD

Test 3

AS IV Serial no: 111327
Version no: 532B

TEST RECORD 00600

Temp Date Time ^{9/} 210L

VOID: RFI
12 04/20/20 14:50

Subject Name

Monthly Maint
Subject I.D.

Chief Bruce Summa
Operator Name, I.D.

200106
Location

Lawson PD

RFI Test

AS IV Serial no: 111327
Version no: 532B

TEST RECORD 00601

Temp Date Time ^{9/} 210L

Air Blank:
04/20/20 14:52 .000
Subject Test: Auto
26 04/20/20 14:52 .000

Subject Name

Monthly Maint
Subject I.D.

Chief Bruce Summa
Operator Name, I.D.

200106
Location

Lawson PD

Self Test