

AS IV Serial no: 111327
Version no: 532B

TEST RECORD 00632

Temp Date Time ^{s/} 210L

Air Blank:
09/28/20 12:59 .000
Calibration Check:
20 09/28/20 12:59 .102

Subject Name

Monthly Maint
Subject I.D.

Chief Bruce Summa
Operator Name, I.D.

200106
Location

Lawson PD

Test 1

AS IV Serial no: 111327
Version no: 532B

TEST RECORD 00633

Temp Date Time ^{s/} 210L

Air Blank:
09/28/20 13:01 .000
Calibration Check:
20 09/28/20 13:01 .103

Subject Name

Monthly Maint.
Subject I.D.

Chief Bruce Summa
Operator Name, I.D.

200106
Location

Lawson PD

Test 2

AS IV Serial no: 111327
Version no: 532B

TEST RECORD 00634

Temp Date Time ^{s/} 210L

Air Blank:
09/28/20 13:02 .000
Calibration Check:
21 09/28/20 13:02 .101

Subject Name

Monthly Maint
Subject I.D.

Chief Bruce Summa
Operator Name, I.D.

200106
Location

Lawson PD

Test 3

AS IV Serial no: 111327
Version no: 532B

TEST RECORD 00635

Temp Date Time ^{s/} 210L

VOID: RFI
12 09/28/20 13:03

Subject Name

Monthly Maint
Subject I.D.

Chief Bruce Summa
Operator Name, I.D.

200106
Location

Lawson PD

RFI test

AS IV Serial no: 111327
Version no: 532B

TEST RECORD 00636

Temp Date Time ^{s/} 210L

Air Blank:
09/28/20 13:04 .000
Subject Test: Auto
23 09/28/20 13:04 .000

Subject Name

Monthly Maint
Subject I.D.

Chief Bruce Summa
Operator Name, I.D.

200106
Location

Lawson PD

Self Test

RepCo

RepCo Marketing Co
3101-188 Stony Brook Drive
Raleigh, NC 27604
888-828-0227

CERTIFICATE OF ANALYSIS

MANUFACTURER AND SUPPLIER: RepCo Marketing Co.

LOT NUMBER: 19002

EXPIRATION DATE: October 16, 2021 at 11:59 p.m.

RepCo Marketing Co. certifies the following:


RepCo Marketing Co. prepared, tested and supplied Lot Number 19002 of Alcohol Certified Solution for simulators. This solution was manufactured and tested by RepCo Marketing Co., with confirmation by ISO 17025 and ISO 17034 accredited institution Alcohol Countermeasure Systems, using NIST standards. Random samples were analyzed by Alcohol Countermeasure Systems utilizing a gas chromatograph and found to contain .1231 gms/dl +/- .003 gms/dl wt/vol ethanol (95% Confidence).

The alcohol and distilled water used in the solution were found to be free of any interfering substance.

This solution will produce a vapor alcohol value of .100 +/- 3% gms/210L breath when heated to 34 Degrees Celsius +/- 0.2 Degrees Celsius in a simulator (95% Confidence).

The date of manufacture for this lot number is October 17, 2019. The expiration date for this lot number is October 16, 2021 at 11:59 p.m.

This document is a true representation of the original Certificate of Analysis.



Alma Palmer, Operations Manager
RepCo Marketing Co.



STATE OF MISSOURI
 DEPARTMENT OF HEALTH AND SENIOR SERVICES
 BREATH ALCOHOL PROGRAM

2

PERMIT
TYPE II
BRUCE SUMMA

is hereby authorized to instruct and supervise operators, train instructors, inspect, calibrate, perform field service and repairs, and operate the following breath analyzer(s):

ALCO-SENSOR IV WITH PRINTER

for the determination of the alcoholic content of blood from a sample of expired air. Permit issued under the provisions of sections 577.020 through 577.041, RSMo and 306.111 through 306.119 RSMo.

DATE 2/27/2020

NUMBER 200106

EXPIRES 2/27/2022

DIRECTOR OF STATE PUBLIC HEALTH LABORATORY

DIRECTOR OF DEPARTMENT OF HEALTH AND SENIOR SERVICES

MO 580-0771 (6-10)

LAB-4 (R6-10)



STATE OF MISSOURI
 DEPARTMENT OF HEALTH AND SENIOR SERVICES
 BREATH ALCOHOL PROGRAM

INSTRUMENT OPERATOR CARD

The named cardholder is authorized to operate an evidential breath alcohol instrument for the determination of the alcoholic content in breath form of expired air in Missouri.

Operator SUMMA, BRUCE
Permit No 200106
Date Issued 2/27/2020 **Date Expires** 2/27/2022

