

RECEIVED

By Tracy Crews at 8:49 am, Aug 11, 2020



MISSOURI DEPARTMENT OF HEALTH AND SENIOR SERVICES  
STATE PUBLIC HEALTH LABORATORY  
**ALCO-SENSOR IV WITH PRINTER MAINTENANCE REPORT**

REPORT #7

Complete this report in duplicate at the time of the regular monthly preventative maintenance check, and whenever instrument is repaired. Send copy to Department of Health and Senior Services; retain original in department file.

ALCO SENSOR IV SN 111331      PRINTER SN 099.3586.573      DATE OF INSPECTION 08/07/2020

LOCATION OF INSTRUMENT (STREET AND CITY) 3131 E Kearney, Springfield, Missouri 65803      TIME OF INSPECTION 4:55 pm

**CHECKLIST:** Place a mark in the box by each item if found to be satisfactory or if operating within established limits. (Write in observed values where determined.) Unmarked items must be corrected before using instrument.

- DIGITAL READOUT (ALL ELEMENTS OPERATIONAL)
- TEMPERATURE OF ALCO SENSOR (10°C - 40°C)
- PRINTER WORKING PROPERLY
- TIME AND DATE DISPLAYING PROPERLY

**BREATH ALCOHOL ACCURACY STANDARDS**

SIMULATOR SOLUTION       COMPRESSED ETHANOL-GAS MIXTURE

STANDARD SUPPLIER REPCO      LOT # 19002      EXP. DATE 10/16/2021

SIMULATOR TEMPERATURE (34°C ± 0.2°C) 33.97      SIMULATOR SN MP2306      SIMULATOR EXP DATE 01/07/2021

CALIBRATION CHECK – (ONLY ONE STANDARD IS TO BE USED PER MAINTENANCE REPORT)  
Run three tests using a standard solution. All three tests must be within ±5% of the standard value and must have a spread of .005 or less. Check the box corresponding to the standard solution being used. (PRINTOUT ATTACHED)

- 0.100% STANDARD - MUST READ BETWEEN 0.095% and 0.105% INCLUSIVE
- 0.080% STANDARD - MUST READ BETWEEN 0.076% and 0.084% INCLUSIVE
- 0.040% STANDARD - MUST READ BETWEEN 0.038% and 0.042% INCLUSIVE

TEST 1 .098      TEST 2 .101      TEST 3 .101

RFI DETECTOR OPERATING

**INDICATE THE NUMBER OF BREATH TESTS IN THE FOLLOWING RANGES SINCE THE LAST MAINTENANCE REPORT: (DO NOT INCLUDE SELF-ADMINISTERED TESTS)**

REFUSALS      (0-.04)      (.05-.09)      (.10-.14)      (.15-.19)      (OVER .19)

List any new parts and describe any alteration or modification that was made to restore the instrument to operate satisfactorily and within established limits (use other side if necessary).

Completed calibration  
Changed time +3 minutes

**INSPECTING OFFICER**

SIGNATURE #727      PRINT NAME D W Henley #727

TYPE II PERMIT NUMBER/EXPIRATION DATE 280247 Exp 08/24/2020      TELEPHONE NUMBER (417) 895-6868

**Return completed report to the:** Breath Alcohol Program, MO Department of Health and Senior Services, Southeast District Office  
2875 James Boulevard  
Poplar Bluff, MO 63901

AS IV Serial no: 111331  
Version no: 532B  
TEST RECORD 00686  
Temp Date Time 210L  
Air Blank:  
08/07/20 17:06 .000  
Calibration:  
23 08/07/20 17:06 .100

Subject Name  
Subject I.D.  
Operator Name, I.D.  
D W HENLEY #727  
Location  
3131 E KEARNEY  
SPRINGFIELD MO  
DW #727

AS IV Serial no: 111331  
Version no: 532B  
TEST RECORD - REPRINT  
TEST RECORD 00687  
Temp Date Time 210L  
Air Blank:  
08/07/20 17:17 .000  
Calibration Check:  
23 08/07/20 17:17 .098

Subject Name  
Subject I.D.  
Operator Name, I.D.  
D W HENLEY #727  
Location  
3131 E KEARNEY  
SPRINGFIELD MO  
DW #727

AS IV Serial no: 111331  
Version no: 532B  
TEST RECORD 00688  
Temp Date Time 210L  
Air Blank:  
08/07/20 17:19 .000  
Calibration Check:  
24 08/07/20 17:19 .101

Subject Name  
Subject I.D.  
Operator Name, I.D.  
D W HENLEY #727  
Location  
3131 E KEARNEY  
SPRINGFIELD MO  
DW #727

AS IV Serial no: 111331  
Version no: 532B  
TEST RECORD 00689  
Temp Date Time 210L  
Air Blank:  
08/07/20 17:20 .000  
Calibration Check:  
25 08/07/20 17:20 .101

Subject Name  
Subject I.D.  
Operator Name, I.D.  
D W HENLEY #727  
Location  
3131 E KEARNEY  
SPRINGFIELD MO  
DW #727

AS IV Serial no: 111331  
Version no: 532B  
TEST RECORD 00690  
Temp Date Time 210L  
VOID: RFI  
12 08/07/20 17:21

Subject Name  
Subject I.D.  
Operator Name, I.D.  
D W HENLEY #727  
Location  
3131 E KEARNEY  
SPRINGFIELD MO  
DW #727

# RepCo

RepCo Marketing Co  
3101-188 Stony Brook Drive  
Raleigh, NC 27604  
888-828-0227

## CERTIFICATE OF ANALYSIS

**MANUFACTURER AND SUPPLIER: RepCo Marketing Co.**  
**LOT NUMBER: 19002**  
**EXPIRATION DATE: October 16, 2021 at 11:59 p.m.**

RepCo Marketing Co. certifies the following:


RepCo Marketing Co. prepared, tested and supplied Lot Number 19002 of Alcohol Certified Solution for simulators. This solution was manufactured and tested by RepCo Marketing Co., with confirmation by ISO 17025 and ISO 17034 accredited institution Alcohol Countermeasure Systems, using NIST standards. Random samples were analyzed by Alcohol Countermeasure Systems utilizing a gas chromatograph and found to contain .1231 gms/dl +/- .003 gms/dl wt/vol ethanol (95% Confidence).

The alcohol and distilled water used in the solution were found to be free of any interfering substance.

This solution will produce a vapor alcohol value of .100 +/- 3% gms/210L breath when heated to 34 Degrees Celsius +/- 0.2 Degrees Celsius in a simulator (95% Confidence).

The date of manufacture for this lot number is October 17, 2019 The expiration date for this lot number is October 16, 2021 at 11:59 p.m.

This document is a true representation of the original Certificate of Analysis.

  
Alma Palmer, Operations Manager  
RepCo Marketing Co.





STATE OF MISSOURI  
 DEPARTMENT OF HEALTH AND SENIOR SERVICES  
 BREATH ALCOHOL PROGRAM

2

**PERMIT  
 TYPE II**

**DAVID W HENLEY, JR.**

is hereby authorized to instruct and supervise operators, train instructors, inspect, calibrate, perform field service and repairs, and operate the following breath analyzer(s):

**ALCO-SENSOR IV WITH PRINTER, INTOX DMT**

for the determination of the alcoholic content of blood from a sample of expired air. Permit issued under the provisions of sections 577.020 through 577.041, RSMo and 306.111 through 306.119 RSMo.

DATE 8/24/2018

NUMBER 280247

EXPIRES 8/24/2020

MO 580-0771 (6-10)

DIRECTOR OF STATE PUBLIC HEALTH LABORATORY

DIRECTOR OF DEPARTMENT OF HEALTH AND SENIOR SERVICES

LAB-4 (R6-10)



STATE OF MISSOURI  
 DEPARTMENT OF HEALTH AND SENIOR SERVICES  
 BREATH ALCOHOL PROGRAM

**INSTRUMENT OPERATOR CARD**

*The named cardholder is authorized to operate an evidential breath alcohol instrument for the determination of the alcoholic content in breath form of expired air in Missouri.*

Operator HENLEY, JR., DAVID  
 Permit No 280247  
 Date Issued 8/24/2018 Date Expires 8/24/2020

