



ALCO-SENSOR IV WITH PRINTER MAINTENANCE REPORT

REPORT #7

Complete this report in duplicate at the time of the regular monthly preventative maintenance check, and whenever instrument is repaired. Send copy to Department of Health and Senior Services; retain original in department file.

ALCO SENSOR IV SN 111333	NAME OF AGENCY Mo State Highway Patrol	DATE OF INSPECTION 10/20/2020
LOCATION OF INSTRUMENT (STREET AND CITY) 2920 N. Shamrock Rd, Jefferson City		TIME OF INSPECTION 8:37 AM

CHECKLIST: Place a mark in the box by each item if found to be satisfactory or if operating within established limits. (Write in observed values where determined.) Unmarked items must be corrected before using instrument.

- DIGITAL READOUT (ALL ELEMENTS OPERATIONAL)
- TEMPERATURE OF ALCO SENSOR (10°C - 40°C)
- PRINTER WORKING PROPERLY
- TIME AND DATE DISPLAYING PROPERLY

BREATH ALCOHOL ACCURACY STANDARDS

- SIMULATOR SOLUTION COMPRESSED ETHANOL-GAS MIXTURE
- STANDARD SUPPLIER **Repro Marketing Co** LOT # **19001** EXP. DATE **04-06-2022**
- SIMULATOR TEMPERATURE (34°C ± 0.2°C) **33.99** SIM. SN **MP2316** SIM. NIST EXP DATE **1-14-2021**

- CALIBRATION CHECK – (ONLY ONE STANDARD IS TO BE USED PER MAINTENANCE REPORT)
Run three tests using a standard solution. All three tests must be within ±5% of the standard value and must have a spread of .005 or less. Check the box corresponding to the standard solution being used. (PRINTOUT ATTACHED)
 - 0.100% STANDARD - MUST READ BETWEEN 0.095% and 0.105% INCLUSIVE
 - 0.080% STANDARD - MUST READ BETWEEN 0.076% and 0.084% INCLUSIVE
 - 0.040% STANDARD - MUST READ BETWEEN 0.038% and 0.042% INCLUSIVE

TEST 1 • .100	TEST 2 • .096	TEST 3 • .100
----------------------	----------------------	----------------------

- RFI DETECTOR OPERATING

INDICATE THE NUMBER OF BREATH TESTS IN THE FOLLOWING RANGES SINCE THE LAST MAINTENANCE REPORT: (DO NOT INCLUDE SELF-ADMINISTERED TESTS)

REFUSALS	0	(0-.04)	0	(.05-.09)	0	(.10-.14)	0	(.15-.19)	0	(OVER .19)	0
----------	----------	---------	----------	-----------	----------	-----------	----------	-----------	----------	------------	----------

List any new parts and describe any alteration or modification that was made to restore the instrument to operate satisfactorily and within established limits (use other side if necessary).

Instrument Calibrated

INSPECTING OFFICER	
SIGNATURE Jimmy L. Cleveland	PRINT NAME Jimmy L. Cleveland
TYPE II PERMIT NUMBER/EXPIRATION DATE 290121/06-13-2021	TELEPHONE NUMBER 573-751-4722

Return completed report to the: Breath Alcohol Program, MO Department of Health and Senior Services, Southeast District Office by mail, fax, or email.

AS IV Serial no: 111333
Version no: 532B

TEST RECORD - REPRINT

TEST RECORD 00682

Temp Date Time 210L

Air Blank:
10/20/20 08:37 .000
Calibration Check:
23 10/20/20 08:37 .100

Subject Name
Calibration
Subject I.D.

Operator Name, I.D.
Jimmy L. Cleveland
Location
2920 N. Shamrock
Jefferson City

AS IV Serial no: 111333
Version no: 532B

TEST RECORD 00683

Temp Date Time 210L

Air Blank:
10/20/20 08:47 .000
Calibration Check:
23 10/20/20 08:47 .100

Subject Name
Acc 1
Subject I.D.

Operator Name, I.D.
Jimmy L. Cleveland
Location
2920 N. Shamrock
Jefferson City

AS IV Serial no: 111333
Version no: 532B

TEST RECORD 00684

Temp Date Time 210L

Air Blank:
10/20/20 08:58 .000
Calibration Check:
23 10/20/20 08:58 .096

Subject Name
Acc 2
Subject I.D.

Operator Name, I.D.
Jimmy L. Cleveland
Location
2920 N. Shamrock
Jefferson City

AS IV Serial no: 111333
Version no: 532B

TEST RECORD 00685

Temp Date Time 210L

Air Blank:
10/20/20 09:04 .000
Calibration Check:
23 10/20/20 09:04 .100

Subject Name
Acc 3
RFI Test Sec
Subject I.D.

Operator Name, I.D.
Jimmy L. Cleveland
Location
2920 N. Shamrock
Jefferson City

AS IV Serial no: 111333
Version no: 532B

TEST RECORD 00686

Temp Date Time 210L

VOID: RFI
12 10/20/20 09:05

Subject Name
RFI Test
Subject I.D.

Operator Name, I.D.
Jimmy L. Cleveland
Location
2920 N. Shamrock
Jefferson City



RepCo Marketing Co
3101-188 Stony Brook Drive
Raleigh, NC 27604
888-828-0227

CERTIFICATE OF ANALYSIS

MANUFACTURER AND SUPPLIER: RepCo Marketing Co.

LOT NUMBER: 19001

EXPIRATION DATE: March 12, 2021 at 11:59 p.m.

RepCo Marketing Co. certifies the following:

RepCo Marketing Co. prepared, tested and supplied Lot Number 19001 of Alcohol Certified Solution for simulators. This solution was manufactured and tested by RepCo Marketing Co., with confirmation by accredited institution, Data Resources Inc., using NIST standards. Random samples were analyzed by Data Resources Inc. utilizing a gas chromatograph and found to contain .1210 gms/dl +/- .003 gms/dl wt/vol ethanol (95% Confidence).

The alcohol and distilled water used in the solution were found to be free of any interfering substance.

This solution will produce a vapor alcohol value of .100 +/- 3% gms/210L breath when heated to 34 Degrees Celsius +/- 0.2 Degrees Celsius in a simulator (95% Confidence).

The date of manufacture for this lot number is March 13, 2019
The expiration date for this lot number is March 12, 2021 at 11:59 p.m.

This document is a true representation of the original Certificate of Analysis.

Alma Palmer, Operations Manager
RepCo Marketing Co.



STATE OF MISSOURI
DEPARTMENT OF HEALTH AND SENIOR SERVICES
BREATH ALCOHOL PROGRAM

2

PERMIT
TYPE II

JIMMY L CLEVELAND

is hereby authorized to instruct and supervise operators, train instructors, inspect, calibrate, perform field service and repairs and operate the following breath analyzer(s):

ALCO-SENSOR IV WITH PRINTER, INTOX DMT

for the determination of the alcoholic content of blood from a sample of expired air. Permit issued under the provisions of sections 577.020 through 577.041, RSMo and 306.111 through 306.119 RSMo.

DATE 6/13/2019

NUMBER 290121

EXPIRES 6/13/2021

DIRECTOR OF STATE PUBLIC HEALTH LABORATORY

DIRECTOR OF DEPARTMENT OF HEALTH AND SENIOR SERVICES

MO 580-0771 (6-10)

L.B-4 (R6-0)



STATE OF MISSOURI
DEPARTMENT OF HEALTH AND SENIOR SERVICES
BREATH ALCOHOL PROGRAM

INSTRUMENT OPERATOR CARD

The named cardholder is authorized to operate an evidential breath alcohol instrument for the determination of the alcoholic content in breath form of expired air in Missouri.

Operator **CLEVELAND, JIMMY**
Permit No **290121**
Date issued **6/13/2019** Date Expires **6/13/2021**

