



ALCO-SENSOR IV WITH PRINTER MAINTENANCE REPORT

REPORT #7

Complete this report in duplicate at the time of the regular monthly preventative maintenance check, and whenever instrument is repaired. Send copy to Department of Health and Senior Services; retain original in department file.

ALCO SENSOR IV SN 111629	PRINTER SN 09B.3589.458	DATE OF INSPECTION 07/22/2020
LOCATION OF INSTRUMENT (STREET AND CITY) 901 Virginia Avenue, Columbia		TIME OF INSPECTION 9:28 am

CHECKLIST: Place a mark in the box by each item if found to be satisfactory or if operating within established limits. (Write in observed values where determined.) Unmarked items must be corrected before using instrument.

DIGITAL READOUT (ALL ELEMENTS OPERATIONAL)

TEMPERATURE OF ALCO SENSOR (10°C - 40°C)

PRINTER WORKING PROPERLY

TIME AND DATE DISPLAYING PROPERLY

BREATH ALCOHOL ACCURACY STANDARDS

SIMULATOR SOLUTION COMPRESSED ETHANOL-GAS MIXTURE

STANDARD SUPPLIER Intoximeters LOT # AG824102 EXP. DATE 08/29/2020

SIMULATOR TEMPERATURE (34°C ± 0.2°C) _____ SIMULATOR SN _____ SIMULATOR EXP DATE _____

CALIBRATION CHECK – (ONLY ONE STANDARD IS TO BE USED PER MAINTENANCE REPORT)

Run three tests using a standard solution. All three tests must be within ±5% of the standard value and must have a spread of .005 or less. Check the box corresponding to the standard solution being used. (PRINTOUT ATTACHED)

- 0.100% STANDARD - MUST READ BETWEEN 0.095% and 0.105% INCLUSIVE
 0.080% STANDARD - MUST READ BETWEEN 0.076% and 0.084% INCLUSIVE
 0.040% STANDARD - MUST READ BETWEEN 0.038% and 0.042% INCLUSIVE

TEST 1 ➔ .078

TEST 2 ➔ .077

TEST 3 ➔ .078

RFI DETECTOR OPERATING

**INDICATE THE NUMBER OF BREATH TESTS IN THE FOLLOWING RANGES SINCE THE LAST MAINTENANCE REPORT:
(DO NOT INCLUDE SELF-ADMINISTERED TESTS)**

REFUSALS	1	(0-.04)	1	(.05-.09)	1	(.10-.14)	1	(.15-.19)	3	(OVER .19)	1
----------	---	---------	---	-----------	---	-----------	---	-----------	---	------------	---

List any new parts and describe any alteration or modification that was made to restore the instrument to operate satisfactorily and within established limits (use other side if necessary).

INSPECTING OFFICER

SIGNATURE

PRINT NAME
Eric Moss

TYPE II PERMIT NUMBER/EXPIRATION DATE
280234 08/14/2020

TELEPHONE NUMBER
(573) 882-7201

Return completed report to the: Breath Alcohol Program, MO Department of Health and Senior Services, Southeast District Office
2875 James Boulevard
Poplar Bluff, MO 63901



Airgas USA LLC (LAB)
 3500 Bernard Street
 St. Louis, Mo. 63103
 Ph: (314) 533-3100
 Fax: (314) 533-7328

Certificate of Analysis

Customer Name
 Exclusive Supplier
 Intoximeters, Inc.
 2081 Craig Road
 St. Louis, Mo 63146

Test Date: 29-Aug-2018

Lot # AG824102 Model 108cacd

<u>Exp. Date</u> 29-Aug-2020	<u>Cyl. Type</u> 108	<u>Component</u> Ethanol Nitrogen	<u>Certified Concentration</u> 0.080 ± 0.002 BrAC (218 ppm) Balance
---------------------------------	-------------------------	---	---

Certification Traceable to N.I.S.T. RGM Ethanol Standards:

<u>Serial No.</u>	<u>Concentration</u>	<u>Serial No.</u>	<u>Concentration</u>
EB0010581	392.1 ppm	EB0010603	393.0 ppm
EB0010570	259.8 ppm	EB0010559	258.2 ppm
EB0010285	208.0 ppm	EB0010595	208.3 ppm
EB0010561	103.6 ppm	EB0010562	104.2 ppm
EB0010681	52.12 ppm	EB0010579	52.81 ppm

Analytical Method: NDIR

Digitally signed by Quality Control
 Date: 2018.08.29 18:38:26 -05:00
 Reason: Dry gas standard certification of analysis
 Location: Airgas USA LLC (Lab)

Approved for Release: _____

Rod Marsala

ISO 17025:2005 A2LA accredited. Certificate Number 3082.06



STATE OF MISSOURI
 DEPARTMENT OF HEALTH AND SENIOR SERVICES
 BREATH ALCOHOL PROGRAM

2

PERMIT
TYPE II
ERIC L MOSS

is hereby authorized to instruct and supervise operators, train instructors, inspect, calibrate; perform field service and repairs, and operate the following breath analyzer(s):

ALCO-SENSOR IV WITH PRINTER, INTOX DMT

for the determination of the alcoholic content of blood from a sample of expired air. Permit issued under the provisions of sections 577.020 through 577.041, RSMo and 306.111 through 306.119 RSMo.

DATE 8/14/2018

NUMBER 280234

EXPIRES 8/14/2020

MO 580-0771 (6-10)

DIRECTOR OF STATE PUBLIC HEALTH LABORATORY

DIRECTOR OF DEPARTMENT OF HEALTH AND SENIOR SERVICES

LAB-4 (R6-10)



STATE OF MISSOURI
 DEPARTMENT OF HEALTH AND SENIOR SERVICES
 BREATH ALCOHOL PROGRAM

INSTRUMENT OPERATOR CARD

The named cardholder is authorized to operate an evidential breath alcohol instrument for the determination of the alcoholic content in breath form of expired air in Missouri.



Operator MOSS, ERIC
 Permit No 280234
 Date Issued 8/14/2018 Date Expires 8/14/2020

AS IV Serial no: 111629
Version no: 532B

TEST RECORD 00766

Temp Date Time ^{s/} 210L

Air Blank:
07/22/20 09:28 .000
Calibration Check:
24 07/22/20 09:28 .078

Subject Name

Cal check 1+3 OK

Subject I.D.

E. Moss MUPD #37

Operator Name, I.D.

280234 8/14/20

Location
MUPD

AS IV Serial no: 111629
Version no: 532B

TEST RECORD 00767

Temp Date Time ^{s/} 210L

Air Blank:
07/22/20 09:30 .000
Calibration Check:
24 07/22/20 09:30 .077

Subject Name

Cal check 2+3 OK

Subject I.D.

E. Moss MUPD #37

Operator Name, I.D.

280234 8/14/20

Location
MUPD

AS IV Serial no: 111629
Version no: 532B

TEST RECORD 00768

Temp Date Time ^{s/} 210L

Air Blank:
07/22/20 09:32 .000
Calibration Check:
24 07/22/20 09:32 .078

Subject Name

Cal check 3+3 OK

Subject I.D.

E. Moss MUPD #37

Operator Name, I.D.

280234 8/14/20

Location
MUPD

AS IV Serial no: 111629
Version no: 532B

TEST RECORD 00769

Temp Date Time ^{s/} 210L

Air Blank:
07/22/20 09:33

Subject Name

FI test OK

Subject I.D.

E. Moss MUPD #37

Operator Name, I.D.

280234 8/14/20

Location
MUPD

AS IV Serial no: 111629
Version no: 532B

TEST RECORD 00770

Temp Date Time ^{s/} 210L

Air Blank:
07/22/20 09:34 .000
Subject Test: Auto
24 07/22/20 09:34 .000

Subject Name

Blank test OK

Subject I.D.

280234 8/14/20

Operator Name, I.D.

E. Moss MUPD #37

Location
MUPD