



ALCO-SENSOR IV WITH PRINTER MAINTENANCE REPORT

REPORT #7

Complete this report in duplicate at the time of the regular monthly preventative maintenance check, and whenever instrument is repaired. Send copy to Department of Health and Senior Services; retain original in department file.

ALCO SENSOR IV SN 111642	PRINTER SN 09B.3589.463	DATE OF INSPECTION 08/01/2020
LOCATION OF INSTRUMENT (STREET AND CITY) Campbell Police Dept. 204 W. Grand Campbell, MO 63933		TIME OF INSPECTION 7:21 am

CHECKLIST: Place a mark in the box by each item if found to be satisfactory or if operating within established limits. (Write in observed values where determined.) Unmarked items must be corrected before using instrument.

DIGITAL READOUT (ALL ELEMENTS OPERATIONAL)

TEMPERATURE OF ALCO SENSOR (10°C - 40°C)

PRINTER WORKING PROPERLY

TIME AND DATE DISPLAYING PROPERLY

BREATH ALCOHOL ACCURACY STANDARDS

SIMULATOR SOLUTION COMPRESSED ETHANOL-GAS MIXTURE

STANDARD SUPPLIER RepCo LOT # 19002 EXP. DATE 10/16/2021

SIMULATOR TEMPERATURE (34°C ± 0.2°C) 34 SIMULATOR SN SD174301 SIMULATOR EXP DATE 01/07/2021

CALIBRATION CHECK – (ONLY ONE STANDARD IS TO BE USED PER MAINTENANCE REPORT)

Run three tests using a standard solution. All three tests must be within ±5% of the standard value and must have a spread of .005 or less. Check the box corresponding to the standard solution being used. (PRINTOUT ATTACHED)

0.100% STANDARD - MUST READ BETWEEN 0.095% and 0.105% INCLUSIVE

0.080% STANDARD - MUST READ BETWEEN 0.076% and 0.084% INCLUSIVE

0.040% STANDARD - MUST READ BETWEEN 0.038% and 0.042% INCLUSIVE

TEST 1 ← .099%

TEST 2 ← .099%

TEST 3 ← .099%

RFI DETECTOR OPERATING

INDICATE THE NUMBER OF BREATH TESTS IN THE FOLLOWING RANGES SINCE THE LAST MAINTENANCE REPORT: (DO NOT INCLUDE SELF-ADMINISTERED TESTS)

REFUSALS	0	(0-.04)	1	(.05-.09)	0	(.10-.14)	0	(.15-.19)	0	(OVER .19)	0
----------	---	---------	---	-----------	---	-----------	---	-----------	---	------------	---

List any new parts and describe any alteration or modification that was made to restore the instrument to operate satisfactorily and within established limits (use other side if necessary).

Instrument is within D.O.H. specifications. Simulator solution RepCo lot # 19002 exp. date 10/16/2021. Vapor alcohol value .100% +/-3%

INSPECTING OFFICER

SIGNATURE

PRINT NAME
Weldon Wallace

TYPE II PERMIT NUMBER/EXPIRATION DATE
200183 5/26/2022

TELEPHONE NUMBER
573-246-2511

Return completed report to the: Breath Alcohol Program, MO Department of Health and Senior Services, Southeast District Office
2875 James Boulevard
Poplar Bluff, MO 63901

AS IV Serial no: 111642
Version no: 532R

TEST RECORD 88483

Temp	Date	Time	216L
Air Blank			
	02/01/78	07:05	.000
Calibration Check			
	02/01/78	07:06	.000

Sim.

Subject I.D.

Operator Name: I.D.

Welder Waller #23
Location
Campbell P.D.

AS IV Serial no: 111642
Version no: 532R

TEST RECORD 88484

Temp	Date	Time	216L
Air Blank			
	02/01/78	07:05	.000
Calibration Check			
	02/01/78	07:05	.000

Subject Name

Sim.

Subject I.D.

Operator Name: I.D.

Welder Waller #23
Location
Campbell P.D.

AS IV Serial no: 111642
Version no: 532R

TEST RECORD 88484

Temp	Date	Time	216L
VOID: RFI			
	02/01/78	07:09	

Subject Name

Sim.

Subject I.D.

Operator Name: I.D.

Welder Waller #23
Location
Campbell P.D.

AS IV Serial no: 111642
Version no: 532R

TEST RECORD 88482

Temp	Date	Time	216L
Air Blank			
	02/01/78	07:06	.000
Calibration Check			
	02/01/78	07:06	.000

Subject Name

Sim.

Subject I.D.

Operator Name: I.D.

Welder Waller #23
Location
Campbell P.D.

RepCo Marketing Co
3101-188 Stony Brook Drive
Raleigh, NC 27604
888-828-0227

CERTIFICATE OF ANALYSIS

MANUFACTURER AND SUPPLIER: RepCo Marketing Co.
LOT NUMBER: 19002
EXPIRATION DATE: October 16, 2021 at 11:59 p.m.

RepCo Marketing Co. certifies the following:


RepCo Marketing Co. prepared, tested and supplied Lot Number 19002 of Alcohol Certified Solution for simulators. This solution was manufactured and tested by RepCo Marketing Co., with confirmation by ISO 17025 and ISO 17034 accredited institution Alcohol Countermeasure Systems, using NIST standards. Random samples were analyzed by Alcohol Countermeasure Systems utilizing a gas chromatograph and found to contain .1231 gms/dl +/- .003 gms/dl wt/vol ethanol (95% Confidence).

The alcohol and distilled water used in the solution were found to be free of any interfering substance.

This solution will produce a vapor alcohol value of .100 +/- 3% gms/210L breath when heated to 34 Degrees Celsius +/- 0.2 Degrees Celsius in a simulator (95% Confidence).

The date of manufacture for this lot number is October 17, 2019 The expiration date for this lot number is October 16, 2021 at 11:59 p.m.

This document is a true representation of the original Certificate of Analysis.


Alma Palmer, Operations Manager
RepCo Marketing Co.



STATE OF MISSOURI
DEPARTMENT OF HEALTH AND SENIOR SERVICES
BREATH ALCOHOL PROGRAM

2

**PERMIT
TYPE II
WELDON WALLACE**

is hereby authorized to instruct and supervise operators, train instructors, inspect, calibrate, perform field service and repairs, and operate the following breath analyzer(s):

ALCO-SENSOR IV WITH PRINTER

for the determination of the alcoholic content of blood from a sample of expired air. Permit issued under the provisions of sections 577.020 through 577.041, RSMo and 308.111 through 308.119 RSMo.

DATE 5/26/2020

DIRECTOR OF STATE PUBLIC HEALTH LABORATORY

NUMBER 200183

DIRECTOR OF DEPARTMENT OF HEALTH AND SENIOR SERVICES

EXPIRES 5/26/2022

LAB-4 (R6-10)

MO 690-0771 (6-10)

STATE OF MISSOURI
DEPARTMENT OF HEALTH AND SENIOR SERVICES
BREATH ALCOHOL PROGRAM

INSTRUMENT OPERATOR CARD

The named cardholder is authorized to operate an evidential breath alcohol instrument for the determination of the alcoholic content in breath form of expired air in Missouri.

Operator **WALLACE, WELDON**
Permit No **200183**
Date Issued **5/26/2020** Date Expires **5/26/2022**