



**ALCO-SENSOR IV WITH PRINTER MAINTENANCE REPORT**

REPORT #7

Complete this report in duplicate at the time of the regular monthly preventative maintenance check, and whenever instrument is repaired. Send copy to Department of Health and Senior Services; retain original in department file.

ALCO SENSOR IV SN <b>111646</b>	PRINTER SN <b>09B, 3589, 473</b>	DATE OF INSPECTION <b>7/18/2020</b>
LOCATION OF INSTRUMENT (STREET AND CITY) <b>211 4th St, Vienna MO, Maries Co. Sheriff's Office</b>		TIME OF INSPECTION <b>1313</b>

**CHECKLIST:** Place a mark in the box by each item if found to be satisfactory or if operating within established limits. (Write in observed values where determined.) Unmarked items must be corrected before using instrument.

DIGITAL READOUT (ALL ELEMENTS OPERATIONAL)

TEMPERATURE OF ALCO SENSOR (10°C - 40°C)

PRINTER WORKING PROPERLY

TIME AND DATE DISPLAYING PROPERLY

**BREATH ALCOHOL ACCURACY STANDARDS**

SIMULATOR SOLUTION

COMPRESSED ETHANOL-GAS MIXTURE

STANDARD SUPPLIER **Intoximeters** LOT # **AG9K006** EXP. DATE **10 Apr 2021**

SIMULATOR TEMPERATURE (34°C ± 0.2°C) \_\_\_\_\_ SIMULATOR SN \_\_\_\_\_ SIMULATOR EXP DATE \_\_\_\_\_

CALIBRATION CHECK – (ONLY ONE STANDARD IS TO BE USED PER MAINTENANCE REPORT)

Run three tests using a standard solution. All three tests must be within ±5% of the standard value and must have a spread of .005 or less. Check the box corresponding to the standard solution being used. (PRINTOUT ATTACHED)

0.100% STANDARD - MUST READ BETWEEN 0.095% and 0.105% INCLUSIVE

0.080% STANDARD - MUST READ BETWEEN 0.076% and 0.084% INCLUSIVE

0.040% STANDARD - MUST READ BETWEEN 0.038% and 0.042% INCLUSIVE

TEST 1 • **.105**

TEST 2 • **.105**

TEST 3 • **.104**

RFI DETECTOR OPERATING

**INDICATE THE NUMBER OF BREATH TESTS IN THE FOLLOWING RANGES SINCE THE LAST MAINTENANCE REPORT: (DO NOT INCLUDE SELF-ADMINISTERED TESTS)**

REFUSALS **0** (0-.04) **0** (.05-.09) **0** (.10-.14) **0** (.15-.19) **0** (OVER .19) **0**

List any new parts and describe any alteration or modification that was made to restore the instrument to operate satisfactorily and within established limits (use other side if necessary).

**Monthly Maintenance**

**INSPECTING OFFICER**

SIGNATURE

PRINT NAME **Dale Harp**

TYPE II PERMIT NUMBER/EXPIRATION DATE **200197 7/8/2022**

TELEPHONE NUMBER **573-422-3381**

**Return completed report to the:** Breath Alcohol Program, MO Department of Health and Senior Services, Southeast District Office  
2875 James Boulevard  
Poplar Bluff, MO 63901

AS IV Serial no: 111646  
Version no: 532B

TEST RECORD 00359

Temp	Date	Time	s/ 210L
Air Blank:			
	07/18/20	13:19	.000
Calibration Check:			
	20 07/18/20	13:19	.105

Air Blank:  
07/18/20 13:19 .000  
Calibration Check:  
20 07/18/20 13:19 .105

Subject Name

Test # 1

Subject I.D.

Operator Name, I.D.

Dale Harp 917

Location

MCSO

AS IV Serial no: 111646  
Version no: 532B

TEST RECORD 00360

Temp	Date	Time	s/ 210L
Air Blank:			
	07/18/20	13:21	.000
Calibration Check:			
	21 07/18/20	13:21	.105

Air Blank:  
07/18/20 13:21 .000  
Calibration Check:  
21 07/18/20 13:21 .105

Subject Name

Test # 2

Subject I.D.

Operator Name, I.D.

Dale Harp 917

Location

MCSO

AS IV Serial no: 111646  
Version no: 532B

TEST RECORD 00361

Temp	Date	Time	s/ 210L
Air Blank:			
	07/18/20	13:23	.000
Calibration Check:			
	22 07/18/20	13:23	.104

Air Blank:  
07/18/20 13:23 .000  
Calibration Check:  
22 07/18/20 13:23 .104

Subject Name

Test # 3

Subject I.D.

Operator Name, I.D.

Dale Harp 917

Location

MCSO

AS IV Serial no: 111646  
Version no: 532B

TEST RECORD 00362

Temp	Date	Time	s/ 210L
VOID: RFI			
	12 07/18/20	13:25	

VOID: RFI

12 07/18/20 13:25

Subject Name

RFI Test

Subject I.D.

Operator Name, I.D.

Dale Harp 917

Location

MCSO



Airgas USA LLC (LAB)

3500 Bernard Street

St. Louis, Mo. 63103

Ph: (314) 533-3100

Fax: (314) 533-7328

## Certificate of Analysis

**Customer Name**

*Exclusive Supplier*  
Intoximeters, Inc.  
2081 Craig Road  
St. Louis, Mo 63146

**Test Date:** 12-Apr-2019

**Lot # AG910006 Model 55cadd**

**Exp. Date**

10-Apr-2021

**Cyl. Type**

55

**Component**

Ethanol  
Nitrogen

**Certified Concentration**

0.100 ± 2% BrAC (272 ppm)  
Balance

**Certification Traceable to N.I.S.T. RGM and to CRM Ethanol Standards:**

**RGM Serial No.**

EB0010581

EB0010570

EB0010285

EB0010561

EB0010681

**Concentration**

392.1 ppm

259.8 ppm

208.0 ppm

103.6 ppm

52.12 ppm

**RGM Serial No.**

EB0010603

EB0010559

EB0010595

EB0010562

EB0010579

**Concentration**

393.0 ppm

258.2 ppm

208.3 ppm

104.2 ppm

52.81 ppm

**CRM Serial No.**

CC434668

CC234503

**Concentration**

800.0 ppm

253.0 ppm

**CRM Serial No.**

0056649

0056662

**Concentration**

390.1 ppm

150.2 ppm

**Analytical Method:**

NDIR

Digitally signed by Quality Control  
Date: 2019.04.16 14:42:56 -05:00  
Reason: Dry gas standard certification of analysis  
Location: Airgas USA LLC (Lab)

Approved for Release: \_\_\_\_\_

Rod Marsala

**ISO 17025:2005 A2LA accredited. Certificate Number 3082.06**

**ISO 17034:2016 A2LA accredited. Certificate Number 3082.07**



STATE OF MISSOURI  
 DEPARTMENT OF HEALTH AND SENIOR SERVICES  
 BREATH ALCOHOL PROGRAM

2

**PERMIT  
 TYPE II**

**DALE M. HARP**

is hereby authorized to instruct and supervise operators, train instructors, inspect, calibrate, perform field service and repairs, and operate the following breath analyzer(s):

**ALCO-SENSOR IV WITH PRINTER**

for the determination of the alcoholic content of blood from a sample of expired air. Permit issued under the provisions of sections 577.020 through 577.041, RSMo and 306.111 through 306.119 RSMo.

DATE 7/8/2020

NUMBER 200197

EXPIRES 7/8/2022

DIRECTOR OF STATE PUBLIC HEALTH LABORATORY

DIRECTOR OF DEPARTMENT OF HEALTH AND SENIOR SERVICES

MO 580-0771 (6-10)

LAB-4 (R6-10)



**STATE OF MISSOURI**  
 DEPARTMENT OF HEALTH AND SENIOR SERVICES  
 BREATH ALCOHOL PROGRAM

**INSTRUMENT OPERATOR CARD**

*The named cardholder is authorized to operate an evidential breath alcohol instrument for the determination of the alcoholic content in breath form of expired air in Missouri.*

**Operator** HARP, DALE  
**Permit No** 200197  
**Date Issued** 7/8/2020 **Date Expires** 7/8/2022

