

AS IV Serial no: 111667
Version no: 532B

TEST RECORD 00297

Temp Date Time ^{s/} 210L

Air Blank:
04/23/20 19:45 .000
Calibration Check:
22 04/23/20 19:45 .104

Subject Name

Test # 1

Subject I.D.

Operator Name, I.D.

William Beeler ²⁸⁰¹⁹⁰ 5/22/2020

Location

415 Third St

Platte City MO

AS IV Serial no: 111667
Version no: 532B

TEST RECORD 00298

Temp Date Time ^{s/} 210L

Air Blank:
04/23/20 19:46 .000
Calibration Check:
23 04/23/20 19:46 .102

Subject Name

Test # 2

Subject I.D.

Operator Name, I.D.

William Beeler ²⁸⁰¹⁹⁰ 5/22/2020

Location

415 Third St

Platte City MO

AS IV Serial no: 111667
Version no: 532B

TEST RECORD 00299

Temp Date Time ^{s/} 210L

Air Blank:
04/23/20 19:48 .000
Calibration Check:
24 04/23/20 19:48 .100

Subject Name

Test # 3

Subject I.D.

Operator Name, I.D.

William Beeler ²⁸⁰¹⁹⁰ 5/22/2020

Location

415 Third St

Platte City MO

AS IV Serial no: 111667
Version no: 532B

TEST RECORD 00300

Temp Date Time ^{s/} 210L

Air Blank:
04/23/20 19:50 .000
Subject Test: Auto
24 04/23/20 19:50 .000

Subject Name

Sobe-Sample

Subject I.D.

Operator Name, I.D.

William Beeler ²⁸⁰¹⁹⁰ 5/22/2020

Location

415 Third St

Platte City MO

AS IV Serial no: 111667
Version no: 532B

TEST RECORD 00301

Temp Date Time ^{s/} 210L

VOID: RFI
12 04/23/20 19:52

Subject Name

R=1

Subject I.D.

Operator Name, I.D.

William Beeler ²⁸⁰¹⁹⁰ 5/22/2020

Location

415 Third St

Platte City MO



Airgas USA LLC (LAB)

3500 Bernard Street

St. Louis, Mo. 63103

Ph: (314) 533-3100

Fax: (314) 533-7328

Certificate of Analysis

Customer Name

Exclusive Supplier
Intoximeters, Inc.
2081 Craig Road
St. Louis, Mo 63146

Test Date: 11-Oct-2019

Lot # AG928303 **Model** 108cacd

<u>Exp. Date</u>	<u>Cyl. Type</u>	<u>Component</u>	<u>Certified Concentration</u>
10-Oct-2021	108	Ethanol Nitrogen	0.100 ± 2% BrAC (272 ppm) Balance

Certification Traceable to N.I.S.T. RGM and to CRM Ethanol Standards:

<u>RGM Serial No.</u>	<u>Concentration</u>	<u>RGM Serial No.</u>	<u>Concentration</u>
EB0010581	392.1 ppm	EB0010603	393.0 ppm
EB0010570	259.8 ppm	EB0010559	258.2 ppm
EB0010285	208.0 ppm	EB0010595	208.3 ppm
EB0010561	103.6 ppm	EB0010562	104.2 ppm
EB0010681	52.12 ppm	EB0010579	52.81 ppm

<u>CRM Serial No.</u>	<u>Concentration</u>	<u>CRM Serial No.</u>	<u>Concentration</u>
CC434668	800.0 ppm	0056649	390.1 ppm
CC234503	253.0 ppm	0056662	150.2 ppm

Analytical Method: NDIR

Digitally signed by Quality Control
Date: 2019.10.11 17:27:37 -05:00
Reason: Dry gas standard certification of analysis
Location: Airgas USA LLC (Lab)

Approved for Release:

Rod Marsala

ISO 17025:2005 A2LA accredited. Certificate Number 3082.06
ISO 17034:2016 A2LA accredited. Certificate Number 3082.07



STATE OF MISSOURI
 DEPARTMENT OF HEALTH AND SENIOR SERVICES
 BREATH ALCOHOL PROGRAM

2

PERMIT
TYPE II
WILLIAM BEELER

is hereby authorized to instruct and supervise operators, train instructors, inspect, calibrate, perform field service and repairs, and operate the following breath analyzer(s):

ALCO-SENSOR IV WITH PRINTER, INTOX DMT

for the determination of the alcoholic content of blood from a sample of expired air. Permit issued under the provisions of sections 577.020 through 577.041, RSMo and 306.111 through 306.119 RSMo.

DATE 5/22/2018

NUMBER 280190

EXPIRES 5/22/2020

DIRECTOR OF STATE PUBLIC HEALTH LABORATORY

DIRECTOR OF DEPARTMENT OF HEALTH AND SENIOR SERVICES



STATE OF MISSOURI
 DEPARTMENT OF HEALTH AND SENIOR SERVICES
 BREATH ALCOHOL PROGRAM

INSTRUMENT OPERATOR CARD

The named cardholder is authorized to operate an evidential breath alcohol instrument for the determination of the alcoholic content in breath form of expired air in Missouri.



Operator BEELER, WILLIAM
Permit No 280190
Date Issued 5/22/2018 **Date Expires** 5/22/2020