



AS IV Serial no: 111732  
Version no: 532B

TEST RECORD 00351

Temp Date Time <sup>s/</sup> 210L

Air Blank:  
06/30/20 13:44 .000  
Calibration Check:  
24 06/30/20 13:44 .101

Subject Name

Subject I.D.

DONALD DEBOARD

Operator Name, I.D.

290122 6/13/2021

Location

Warrensburg

AS IV Serial no: 111732  
Version no: 532B

TEST RECORD 00352

Temp Date Time <sup>s/</sup> 210L

Air Blank:  
06/30/20 13:45 .000  
Calibration Check:  
25 06/30/20 13:45 .099

Subject Name

Subject I.D.

Donald DeBoard

Operator Name, I.D.

290122 6/13/2021

Location

Warrensburg

AS IV Serial no: 111732  
Version no: 532B

TEST RECORD 00353

Temp Date Time <sup>s/</sup> 210L

Air Blank:  
06/30/20 13:47 .000  
Calibration Check:  
26 06/30/20 13:47 .100

Subject Name

Subject I.D.

DONALD DEBOARD

Operator Name, I.D.

290122 6/13/2021

Location

Warrensburg

AS IV Serial no: 111732  
Version no: 532B

TEST RECORD 00354

Temp Date Time <sup>s/</sup> 210L

VOID: RFI  
12 06/30/20 13:48

Subject Name

Subject I.D.

DONALD DEBOARD

Operator Name, I.D.

290122 6/13/2021

Location

Warrensburg



## GUTH LABORATORIES, INC.

590 NORTH 67th STREET • HARRISBURG, PA 17111-4511 • TELEPHONE: 717-564-5470

### CERTIFICATE OF ANALYSIS

Certified Alcohol Reference Solution for Simulator

Random Samples of Lot Number 18200 of Alcohol Reference Solution for Simulator were analyzed by gas chromatography on **July 6, 2018**, using a Perkin Elmer Gas Chromatograph Autosystem XL S/N: 610N9030209, and found to contain **0.1215%** (w/vol) ethyl alcohol. The expiration date for this lot number is **July 3, 2020** at 11:59 PM.

When used in a calibrated Simulator, operating at  $34^{\circ}\text{C} \pm .2^{\circ}\text{C}$ , this solution will give a breath alcohol analysis instrument reading of **0.100 g/210L  $\pm$  3%**.

The alcohol and water used in this solution were free of test interfering substances.

Ted L. Pauley, President  
GUTH LABORATORIES, INC.

***NIST Traceability:***

*Testing was conducted using Cerilliant Reference Standard lot number FN04271602 whose values are traceable to NIST.*

*All balances are calibrated annually by an outside agency using NIST traceable weights. Calibration verification is done prior to each use utilizing NIST traceable weights.*



STATE OF MISSOURI  
 DEPARTMENT OF HEALTH AND SENIOR SERVICES  
 BREATH ALCOHOL PROGRAM



**PERMIT  
 TYPE II**

**DONALD D DEBOARD**

is hereby authorized to instruct and supervise operators, train instructors, inspect, calibrate, perform field service and repairs, and operate the following breath analyzer(s):

**INTOX DMT, INTOXILYZER 8000, INTOX EC/IR II, ASIV W/PRINTER**

for the determination of the alcoholic content of blood from a sample of expired air. Permit issued under the provisions of sections 577.020 through 577.041, RSMo and 306.111 through 306.119 RSMo.

DATE 6/13/2019

NUMBER 290122

EXPIRES 6/13/2021

DIRECTOR OF STATE PUBLIC HEALTH LABORATORY

DIRECTOR OF DEPARTMENT OF HEALTH AND SENIOR SERVICES