



MISSOURI DEPARTMENT OF HEALTH AND SENIOR SERVICES  
 STATE PUBLIC HEALTH LABORATORY  
**ALCO-SENSOR IV WITH PRINTER MAINTENANCE REPORT**

**RECEIVED**  
 By Tracy Crews at 10:16 am, Sep 24, 2020

REPORT #7

Complete this report in duplicate at the time of the regular monthly preventative maintenance check, and whenever instrument is repaired. Send copy to Department of Health and Senior Services; retain original in department file.

ALCO SENSOR IV SN 111732	NAME OF AGENCY Ozark Police Department	DATE OF INSPECTION 09/23/2020
LOCATION OF INSTRUMENT (STREET AND CITY) Ozark Police Department. 201 E. Brick Street, Ozark, MO 65721		TIME OF INSPECTION 14:21

**CHECKLIST:** Place a mark in the box by each item if found to be satisfactory or if operating within established limits. (Write in observed values where determined.) Unmarked items must be corrected before using instrument.

- DIGITAL READOUT (ALL ELEMENTS OPERATIONAL)
- TEMPERATURE OF ALCO SENSOR (10°C - 40°C)
- PRINTER WORKING PROPERLY
- TIME AND DATE DISPLAYING PROPERLY

**BREATH ALCOHOL ACCURACY STANDARDS**

<input checked="" type="checkbox"/> SIMULATOR SOLUTION	<input type="checkbox"/> COMPRESSED ETHANOL-GAS MIXTURE
<input checked="" type="checkbox"/> STANDARD SUPPLIER <u>Guth</u>	LOT # <u>19370</u> EXP. DATE <u>12/09/2021</u>
<input checked="" type="checkbox"/> SIMULATOR TEMPERATURE (34°C ± 0.2°C) <u>34.0</u>	SIM. SN <u>DR6932</u> SIM. NIST EXP DATE <u>12/04/2020</u>

**CALIBRATION CHECK – (ONLY ONE STANDARD IS TO BE USED PER MAINTENANCE REPORT)**  
 Run three tests using a standard solution. All three tests must be within ±5% of the standard value and must have a spread of .005 or less. Check the box corresponding to the standard solution being used. (PRINTOUT ATTACHED)

- 0.100% STANDARD - MUST READ BETWEEN 0.095% and 0.105% INCLUSIVE
- 0.080% STANDARD - MUST READ BETWEEN 0.076% and 0.084% INCLUSIVE
- 0.040% STANDARD - MUST READ BETWEEN 0.038% and 0.042% INCLUSIVE

TEST 1 $\blacktriangleleft$ .100	TEST 2 $\blacktriangleleft$ .101	TEST 3 $\blacktriangleleft$ .101
----------------------------------	----------------------------------	----------------------------------

RFI DETECTOR OPERATING

**INDICATE THE NUMBER OF BREATH TESTS IN THE FOLLOWING RANGES SINCE THE LAST MAINTENANCE REPORT: (DO NOT INCLUDE SELF-ADMINISTERED TESTS)**

REFUSALS	(0-.04)	(.05-.09)	(.10-.14)	(.15-.19)	(OVER .19)
----------	---------	-----------	-----------	-----------	------------

List any new parts and describe any alteration or modification that was made to restore the instrument to operate satisfactorily and within established limits (use other side if necessary).

---



---



---

<b>INSPECTING OFFICER</b>	
SIGNATURE ▶	PRINT NAME <b>Stephen Conner</b>
TYPE II PERMIT NUMBER/EXPIRATION DATE <b>200194/07-08-2022</b>	TELEPHONE NUMBER <b>(417 )581-6600</b>

**Return completed report to the:** Breath Alcohol Program, MO Department of Health and Senior Services, Southeast District Office by mail, fax, or email.

AS IV Serial no: 111732  
Version no: 532B

TEST RECORD 00581

Temp Date Time 210L  
s/

Air Blank:  
09/23/20 14:21 .000  
Calibration Check:  
23 09/23/20 14:21 .100

Subject Name

Maintenance  
Subject I.D.  
1

Operator Name, I.D.

S. Conner / 200194  
Location  
Ozark PD

201 E. Brack St

AS IV Serial no: 111732  
Version no: 532B

TEST RECORD 00583

Temp Date Time 210L  
s/

Air Blank:  
09/23/20 14:25 .000  
Calibration Check:  
23 09/23/20 14:25 .101

Subject Name

Maintenance  
Subject I.D.  
3

Operator Name, I.D.

S. Conner / 200194  
Location  
Ozark PD

201 E. Brack St

AS IV Serial no: 111732  
Version no: 532B

TEST RECORD 00582

Temp Date Time 210L  
s/

Air Blank:  
09/23/20 14:23 .000  
Calibration Check:  
23 09/23/20 14:23 .101

Subject Name

Maintenance  
Subject I.D.  
2

Operator Name, I.D.

S. Conner / 200194  
Location  
Ozark PD

201 E. Brack St

AS IV Serial no: 111732  
Version no: 532B

TEST RECORD 00586

Temp Date Time 210L  
s/

VOID: RFI  
12 09/23/20 14:32

Subject Name

Maintenance  
Subject I.D.  
RFI

Operator Name, I.D.

S. Conner / 200194  
Location  
Ozark PD

201 E. Brack St



**GUTH LABORATORIES, INC.**

590 NORTH 67th STREET • HARRISBURG, PA 17111-4511 • TELEPHONE: 717-564-5470

## CERTIFICATE OF ANALYSIS

Certified Alcohol Reference Solution for Simulator

Random Samples of Lot Number **19370** of Alcohol Reference Solution for Simulator were analyzed by gas chromatography on **December 10, 2019**, using a Perkin Elmer Gas Chromatograph Autosystem XL S/N: 610N9030209, and found to contain **0.1199%** (w/vol) ethyl alcohol. The expiration date for this lot number is **December 9, 2021** at 11:59 PM.

When used in a calibrated Simulator, operating at  $34^{\circ}\text{C} \pm 0.2^{\circ}\text{C}$ , this solution will give a breath alcohol analysis instrument reading of **0.100 g/210L  $\pm$  3%**.


The alcohol and water used in this solution were free of test interfering substances.

Ted L. Pauley, President  
GUTH LABORATORIES, INC.

*NIST Traceability:*

*Testing was conducted using Cerilliant Reference Standard lot number FN04271602 whose values are traceable to NIST.*

*All balances are calibrated annually by an outside agency using NIST traceable weights. Calibration verification is done prior to each use utilizing NIST traceable weights.*



STATE OF MISSOURI  
 DEPARTMENT OF HEALTH AND SENIOR SERVICES  
 BREATH ALCOHOL PROGRAM

2

**PERMIT  
 TYPE II**

**STEPHEN L. CONNER**

is hereby authorized to instruct and supervise operators, train instructors, inspect, calibrate, perform field service and repairs, and operate the following breath analyzer(s):

**ALCO-SENSOR IV WITH PRINTER**

for the determination of the alcoholic content of blood from a sample of expired air. Permit issued under the provisions of sections 577.020 through 577.041, RSMo and 306.111 through 306.119 RSMo.

DATE 7/8/2020

NUMBER 200194

EXPIRES 7/8/2022

MO 580-0771 (6-10)

DIRECTOR OF STATE PUBLIC HEALTH LABORATORY

DIRECTOR OF DEPARTMENT OF HEALTH AND SENIOR SERVICES

LAB-4 (R6-10)

STATE OF MISSOURI  
 DEPARTMENT OF HEALTH AND SENIOR SERVICES  
 BREATH ALCOHOL PROGRAM

**INSTRUMENT OPERATOR CARD**

*The named cardholder is authorized to operate an evidential breath alcohol instrument for the determination of the alcoholic content in breath form of expired air in Missouri.*

Operator CONNER, STEPHEN  
 Permit No 200194  
 Date Issued 7/8/2020 Date Expires 7/8/2022



MISSOURI DEPARTMENT OF HEALTH AND SENIOR SERVICES  
STATE PUBLIC HEALTH LABORATORY  
BREATH ALCOHOL PROGRAM

APPLICATION FOR TYPE II PERMIT FOR OPERATION OF BREATH ALCOHOL ANALYZERS

RECEIVED

By Tracy Crews at 8:10 am, Jul 02, 2020

APPROVED

By Stephen Wilson at 8:56 am, Jul 08, 2020

THIS APPLICATION IS FOR  
 NEW PERMIT  RENEWAL  
 CURRENT PERMIT NUMBER AND EXPIRATION DATE: 280267 / 09.06.2020

PRINT FULL NAME: Stephen Louis Corner TITLE: Detective AGE: 31

DEPARTMENT OR GROUP: Ozark Police Department

BUSINESS ADDRESS (STREET, CITY, STATE, ZIP CODE): 201 E. Birch Street, Ozark, MO 65721 TELEPHONE: (417) 531-6600

EMAIL ADDRESS: S. Corner @ ozarkpol.org

A disclosure concerning your SSN number is available at:  
<http://www.health.mo.gov/lab/breathalcohol/>

LIST ALL ORIGINAL TRAINING COURSES FOR OPERATION OF BREATH ANALYZERS  
 (Also, please place a checkmark beside ALL breath analyzer(s) for which you are requesting a permit.)

DATES OF COURSE	LOCATION OF COURSE	COURSE LENGTH (HRS.)	NAME & MODEL OF BREATH ANALYZER	PLACE A CHECKMARK BESIDE INSTRUMENTS FOR WHICH YOU REQUEST	NAME OF INSTRUCTOR
6/30/20	MSC	8	AS-EV w/Printer Lab	<input checked="" type="checkbox"/>	Bond
09/2016	MSC	40	Inter DMT Supervisor	<input type="checkbox"/>	Welsh
09/2016	MSC	14	Inter DMT	<input type="checkbox"/>	Welsh

List the manufacturer and name of instruments for which you are currently performing maintenance reports on and the number of maintenance reports performed on EACH type in the last year.

MANUFACTURER AND NAME OF INSTRUMENT	NUMBER OF MAINTENANCE REPORTS	NUMBER OF SUBJECT TESTS
1. <u>AS-EV w./Printer</u>	<u>10</u> <input type="checkbox"/> OK SGW	<u>5</u> <input type="checkbox"/> SELF-TESTS OK SGW
2.		
3.		

When adding a new instrument, you receive a new two (2) year permit. Therefore, normal renewal procedures apply for the instrument(s) on your current permit that you wish to transfer to the new permit. Disregarding these renewal procedures will result in a new permit for the new instrument only.

To renew a Type II Permit, the applicant shall have completed two (2) Maintenance Reports and shall have performed at least ten (10) tests on drinking subjects in the past year on each instrument for which renewal is requested. If these conditions are not met, or the permit has expired for more than thirty (30) days, the applicant shall perform two (2) Maintenance Reports and five (5) self-administered tests for each breath analyzer for which renewal is requested. Copies of the Maintenance Reports along with the Operational checklists and printouts for the five (5) self-administered tests shall accompany the application for renewal.

SIGNATURE OF APPLICANT: [Signature] DATE: 6/30/2020

URN COMPLETED APPLICATION TO THE: Breath Alcohol Program, Missouri Department of Health and Senior Services  
 Southeast District Office  
 2875 James Blvd.  
 Poplar Bluff, MO 63901