



ALCO-SENSOR IV WITH PRINTER MAINTENANCE REPORT

REPORT #7

Complete this report in duplicate at the time of the regular monthly preventative maintenance check, and whenever instrument is repaired. Send copy to Department of Health and Senior Services; retain original in department file.

ALCO SENSOR IV SN 111733	PRINTER SN 09B.3589.519	DATE OF INSPECTION 02/21/2020
-----------------------------	----------------------------	----------------------------------

LOCATION OF INSTRUMENT (STREET AND CITY) 14301 SOUTH OUTER FORTY RD, CHESTERFIELD	TIME OF INSPECTION 11:32 am
--	--------------------------------

CHECKLIST: Place a mark in the box by each item if found to be satisfactory or if operating within established limits. (Write in observed values where determined.) Unmarked items must be corrected before using instrument.

DIGITAL READOUT (ALL ELEMENTS OPERATIONAL)

TEMPERATURE OF ALCO SENSOR (10°C - 40°C)

PRINTER WORKING PROPERLY

TIME AND DATE DISPLAYING PROPERLY

BREATH ALCOHOL ACCURACY STANDARDS

SIMULATOR SOLUTION COMPRESSED ETHANOL-GAS MIXTURE

STANDARD SUPPLIER INTOXIMETERS, INC. LOT # AG829708 EXP. DATE 10/24/2020

SIMULATOR TEMPERATURE (34°C ± 0.2°C) _____ SIMULATOR SN _____ SIMULATOR EXP DATE _____

CALIBRATION CHECK – (ONLY ONE STANDARD IS TO BE USED PER MAINTENANCE REPORT)
Run three tests using a standard solution. All three tests must be within ±5% of the standard value and must have a spread of .005 or less. Check the box corresponding to the standard solution being used. (PRINTOUT ATTACHED)

- 0.100% STANDARD - MUST READ BETWEEN 0.095% and 0.105% INCLUSIVE
- 0.080% STANDARD - MUST READ BETWEEN 0.076% and 0.084% INCLUSIVE
- 0.040% STANDARD - MUST READ BETWEEN 0.038% and 0.042% INCLUSIVE

TEST 1 .102

TEST 2 .102

TEST 3 .101

RFI DETECTOR OPERATING

**INDICATE THE NUMBER OF BREATH TESTS IN THE FOLLOWING RANGES SINCE THE LAST MAINTENANCE REPORT:
(DO NOT INCLUDE SELF-ADMINISTERED TESTS)**

REFUSALS	0	(0-.04)	0	(.05-.09)	1	(.10-.14)	0	(.15-.19)	1	(OVER .19)	1
----------	---	---------	---	-----------	---	-----------	---	-----------	---	------------	---

List any new parts and describe any alteration or modification that was made to restore the instrument to operate satisfactorily and within established limits (use other side if necessary).

TIME UPDATED

INSPECTING OFFICER

SIGNATURE 	PRINT NAME PO JAHNS, DSN 3725
---------------	----------------------------------

TYPE II PERM NUMBER/EXPIRATION DATE 290039 02/20/2021	TELEPHONE NUMBER (636) 529-8210
--	------------------------------------

Return completed report to the: Breath Alcohol Program, MO Department of Health and Senior Services, Southeast District Office
2875 James Boulevard
Poplar Bluff, MO 63901

AS IV Serial no: 111733
Version no: 532B

TEST RECORD 00165

Temp Date Time ^{s/} 210L

Air Blank:
02/21/20 11:30 .000
Calibration Check:
21 02/21/20 11:30 .102

Subject Name

Cal Check #1

Subject I.D.

N/A

Operator Name, I.D.

Jahns 322S

Location

MoPot TMC

AS IV Serial no: 111733
Version no: 532B

TEST RECORD 00166

Temp Date Time ^{s/} 210L

Air Blank:
02/21/20 11:32 .000
Calibration Check:
21 02/21/20 11:32 .102

Subject Name

Cal Check #2

Subject I.D.

N/A

Operator Name, I.D.

Jahns 322S

Location

MoPot TMC

AS IV Serial no: 111733
Version no: 532B

TEST RECORD 00167

Temp Date Time ^{s/} 210L

Air Blank:
02/21/20 11:33 .000
Calibration Check:
22 02/21/20 11:33 .101

Subject Name

Cal Check #3

Subject I.D.

N/A

Operator Name, I.D.

Jahns 322S

Location

MoPot TMC

AS IV Serial no: 111733
Version no: 532B

TEST RECORD 00168

Temp Date Time ^{s/} 210L

VOID: RFI
12 02/21/20 11:34

Subject Name

RFI Check

Subject I.D.

N/A

Operator Name, I.D.

Jahns 322S

Location

MoPot TMC



Airgas USA LLC (LAB)
 3500 Bernard Street
 St. Louis, Mo. 63103
 Ph: (314) 533-3100
 Fax: (314) 533-7328

Certificate of Analysis

Customer Name
 Exclusive Supplier
 Intoximeters, Inc.
 2081 Craig Road
 St. Louis, Mo 63146

Test Date: 25-Oct-2018

Lot # AG829708 Model 108cacd

<u>Exp. Date</u>	<u>Cyl. Type</u>	<u>Component</u>	<u>Certified Concentration</u>
24-Oct-2020	108	Ethanol Nitrogen	0.100 ± 2% BrAC (272 ppm) Balance

Certification Traceable to N.I.S.T. RGM Ethanol Standards:

<u>Serial No.</u>	<u>Concentration</u>	<u>Serial No.</u>	<u>Concentration</u>
EB0010581	392.1 ppm	EB0010603	393.0 ppm
EB0010570	259.8 ppm	EB0010559	258.2 ppm
EB0010285	208.0 ppm	EB0010595	208.3 ppm
EB0010561	103.6 ppm	EB0010562	104.2 ppm
EB0010681	52.12 ppm	EB0010579	52.81 ppm

Analytical Method: NDIR

Digitally signed by Quality Control
 Date: 2018.10.25 14:13:39 -05:00
 Reason: Dry gas standard certification of analysis
 Location: Airgas USA LLC (Lab)

Approved for Release: _____

Rod Marsala
 Rod Marsala

ISO 17025:2005 A2LA accredited. Certificate Number 3082.06



STATE OF MISSOURI
 DEPARTMENT OF HEALTH AND SENIOR SERVICES
 BREATH ALCOHOL PROGRAM

2

**PERMIT
 TYPE II**

KYLE JAHNS

Is hereby authorized to instruct and supervise operators, train instructors, inspect, calibrate, perform field service and repairs, and operate the following breath analyzer(s):

ALCO-SENSOR IV WITH PRINTER, INTOX DMT

for the determination of the alcoholic content of blood from a sample of expired air. Permit issued under the provisions of sections 577.020 through 577.041, RSMo and 306.111 through 306.119 RSMo.

DATE 2/20/2019

NUMBER 290039

EXPIRES 2/20/2021

DIRECTOR OF STATE PUBLIC HEALTH LABORATORY

DIRECTOR OF DEPARTMENT OF HEALTH AND SENIOR SERVICES