

AS IU Serial no: 111749
Version no: 532B

TEST RECORD 00631

Temp Date Time 210L
s/
Air Blank:
03/06/20 21:43 .000
Calibration Check:
19 03/06/20 21:43 .102

Subject Name

TEST # 1

Subject I.D.

Operator Name, I.D.

M.S. HAWKHAHAW

Location A 290089

A-3 WEST

MAUIEAW, MO

AS IU Serial no: 111749
Version no: 532B

TEST RECORD 00632

Temp Date Time 210L
s/
Air Blank:
03/06/20 21:45 .000
Calibration Check:
20 03/06/20 21:45 .101

Subject Name

TEST # 2

Subject I.D.

Operator Name, I.D.

M.S. HAWKHAHAW

Location A 290089

A-3 WEST

MAUIEAW, MO

AS IU Serial no: 111749
Version no: 532B

TEST RECORD 00633

Temp Date Time 210L
s/
Air Blank:
03/06/20 21:46 .000
Calibration Check:
21 03/06/20 21:46 .099

Subject Name

TEST # 3

Subject I.D.

Operator Name, I.D.

M.S. HAWKHAHAW

Location A 290089

A-3 WEST

MAUIEAW, MO

AS IU Serial no: 111749
Version no: 532B

TEST RECORD 00634

Temp Date Time 210L
s/
UOID: RTI
12 03/06/20 21:48

Subject Name

R F I TEST

Subject I.D.

Operator Name, I.D.

M.S. HAWKHAHAW

Location A 290089

A-3 WEST

MAUIEAW, MO

CERTIFICATE OF ANALYSIS

MANUFACTURER AND SUPPLIER: RepCo Marketing Co.
LOT NUMBER: 19002
EXPIRATION DATE: October 16, 2021 at 11:59 p.m.

RepCo Marketing Co. certifies the following:


RepCo Marketing Co. prepared, tested and supplied Lot Number 19002 of Alcohol Certified Solution for simulators. This solution was manufactured and tested by RepCo Marketing Co., with confirmation by ISO 17025 and ISO 17034 accredited institution Alcohol Countermeasure Systems, using NIST standards. Random samples were analyzed by Alcohol Countermeasure Systems utilizing a gas chromatograph and found to contain .1231 gms/dl +/- .003 gms/dl wt/vol ethanol (95% Confidence).

The alcohol and distilled water used in the solution were found to be free of any interfering substance.

This solution will produce a vapor alcohol value of .100 +/- 3% gms/210L breath when heated to 34 Degrees Celsius +/- 0.2 Degrees Celsius in a simulator (95% Confidence).

The date of manufacture for this lot number is October 17, 2019 The expiration date for this lot number is October 16, 2021 at 11:59 p.m.

This document is a true representation of the original Certificate of Analysis.


Alma Palmer, Operations Manager
RepCo Marketing Co.



STATE OF MISSOURI
DEPARTMENT OF HEALTH AND SENIOR SERVICES
BREATH ALCOHOL PROGRAM

2

**PERMIT
TYPE II**

MATTHEW HANRAHAN

is hereby authorized to instruct and supervise operators, train instructors, inspect, calibrate, perform field service and repairs, and operate the following breath analyzer(s):

ALCO-SENSOR IV WITH PRINTER, INTOX DMT

for the determination of the alcoholic content of blood from a sample of expired air. Permit issued under the provisions of sections 577.020 through 577.041, RSMo and 306.111 through 306.119 RSMo.

DATE 4/22/2019

NUMBER 290089

EXPIRES 4/22/2021

MO 580-0771 (6-10)

DIRECTOR OF STATE PUBLIC HEALTH LABORATORY

DIRECTOR OF DEPARTMENT OF HEALTH AND SENIOR SERVICES

LAB-4 (R6-10)

STATE OF MISSOURI
DEPARTMENT OF HEALTH AND SENIOR SERVICES
BREATH ALCOHOL PROGRAM

INSTRUMENT OPERATOR CARD

The named cardholder is authorized to operate an evidential breath alcohol instrument for the determination of the alcoholic content in breath form of expired air in Missouri.

Operator HANRAHAN, MATTHEW
Permit No 290089
Date Issued 4/22/2019 Date Expires 4/22/2021