



MISSOURI DEPARTMENT OF HEALTH AND SENIOR SERVICES
 STATE PUBLIC HEALTH LABORATORY
ALCO-SENSOR IV WITH PRINTER MAINTENANCE REPORT

REPORT #7

Complete this report in duplicate at the time of the regular monthly preventative maintenance check, and whenever instrument is repaired. Send copy to Department of Health and Senior Services; retain original in department file.

ALCO SENSOR IV SN 111753	PRINTER SN 09B.3589.512	DATE OF INSPECTION 04/04/2020
LOCATION OF INSTRUMENT (STREET AND CITY) 709 W. Main Street, Greenwood, Missouri 64034		TIME OF INSPECTION 6:52 pm

CHECKLIST: Place a mark in the box by each item if found to be satisfactory or if operating within established limits. (Write in observed values where determined.) Unmarked items must be corrected before using instrument.

- DIGITAL READOUT (ALL ELEMENTS OPERATIONAL)
- TEMPERATURE OF ALCO SENSOR (10°C - 40°C)
- PRINTER WORKING PROPERLY
- TIME AND DATE DISPLAYING PROPERLY

BREATH ALCOHOL ACCURACY STANDARDS

- SIMULATOR SOLUTION COMPRESSED ETHANOL-GAS MIXTURE

STANDARD SUPPLIER Guth Laboratories LOT # 19370 EXP. DATE 12/09/2021

SIMULATOR TEMPERATURE (34°C ± 0.2°C) 34.0 SIMULATOR SN SD2285 SIMULATOR EXP DATE 08/13/2020

- CALIBRATION CHECK – (ONLY ONE STANDARD IS TO BE USED PER MAINTENANCE REPORT)
 Run three tests using a standard solution. All three tests must be within ±5% of the standard value and must have a spread of .005 or less. Check the box corresponding to the standard solution being used. (PRINTOUT ATTACHED)
 - 0.100% STANDARD - MUST READ BETWEEN 0.095% and 0.105% INCLUSIVE
 - 0.080% STANDARD - MUST READ BETWEEN 0.076% and 0.084% INCLUSIVE
 - 0.040% STANDARD - MUST READ BETWEEN 0.038% and 0.042% INCLUSIVE

TEST 1 .100 TEST 2 .103 TEST 3 .103

- RFI DETECTOR OPERATING

INDICATE THE NUMBER OF BREATH TESTS IN THE FOLLOWING RANGES SINCE THE LAST MAINTENANCE REPORT: (DO NOT INCLUDE SELF-ADMINISTERED TESTS)

REFUSALS	0	(0-.04)	2	(.05-.09)	0	(.10-.14)	0	(.15-.19)	0	(OVER .19)	0
----------	---	---------	---	-----------	---	-----------	---	-----------	---	------------	---

List any new parts and describe any alteration or modification that was made to restore the instrument to operate satisfactorily and within established limits (use other side if necessary).

INSPECTING OFFICER	
SIGNATURE <i>J. Hutchinson - #171</i>	PRINT NAME Hutchinson #171
TYPE II PERMIT NUMBER/EXPIRATION DATE 200128 / 04/04/2020	TELEPHONE NUMBER (816) 537-5020

Return completed report to the: Breath Alcohol Program, MO Department of Health and Senior Services, Southeast District Office
 2875 James Boulevard
 Poplar Bluff, MO 63901

AS IV Serial no: 111753
Version no: 532B

TEST RECORD 00191
Temp Date Time 210L
s/

Air Blank:
04/04/20 18:53 .000
Calibration Check:
20 04/04/20 18:53 .100

Subject Name
Test 1 00191
Subject I.D.
Hutchinson #171
Operator Name, I.D.
200128
Location
Greenwood PD

AS IV Serial no: 111753
Version no: 532B

TEST RECORD 00192
Temp Date Time 210L
s/

Air Blank:
04/04/20 18:55 .000
Calibration Check:
21 04/04/20 18:55 .103

Subject Name
Test 2 00192
Subject I.D.
Hutchinson #171
Operator Name, I.D.
200128
Location
Greenwood PD

AS IV Serial no: 111753
Version no: 532B

TEST RECORD 00193
Temp Date Time 210L
s/

Air Blank:
04/04/20 18:57 .000
Calibration Check:
21 04/04/20 18:57 .103

Subject Name
Test 3 00193
Subject I.D.
Hutchinson #171
Operator Name, I.D.
200128
Location
Greenwood PD

AS IV Serial no: 111753
Version no: 532B

TEST RECORD 00194
Temp Date Time 210L
s/

VOID: RFI
12 04/04/20 19:01

Subject Name
Test - RFI
Subject I.D.
Hutchinson #171
Operator Name, I.D.
200128
Location
Greenwood PD

Monthly AS-IV Maintenance Check

Greenwood PD 4-4-2020



GUTH LABORATORIES, INC.

590 NORTH 67th STREET • HARRISBURG, PA 17111-4511 • TELEPHONE: 717-564-5470

CERTIFICATE OF ANALYSIS

Certified Alcohol Reference Solution for Simulator

Random Samples of Lot Number **19370** of Alcohol Reference Solution for Simulator were analyzed by gas chromatography on **December 10, 2019**, using a Perkin Elmer Gas Chromatograph Autosystem XL S/N: 610N9030209, and found to contain **0.1199%** (w/vol) ethyl alcohol. The expiration date for this lot number is **December 9, 2021** at 11:59 PM.

When used in a calibrated Simulator, operating at **34°C +/- .2°C**, this solution will give a breath alcohol analysis instrument reading of **0.100 g/210L +/- 3%**.

The alcohol and water used in this solution were free of test interfering substances.

Ted L. Pauley, President
GUTH LABORATORIES, INC.

NIST Traceability:

Testing was conducted using Cerilliant Reference Standard lot number FN04271602 whose values are traceable to NIST.

All balances are calibrated annually by an outside agency using NIST traceable weights. Calibration verification is done prior to each use utilizing NIST traceable weights.



STATE OF MISSOURI
DEPARTMENT OF HEALTH AND SENIOR SERVICES
BREATH ALCOHOL PROGRAM

2

PERMIT
TYPE II

ANDREW HUTCHINSON

is hereby authorized to instruct and supervise operators, train instructors, inspect, calibrate, perform field service and repairs, and operate the following breath analyzer(s):

ALCO-SENSOR IV WITH PRINTER

for the determination of the alcoholic content of blood from a sample of expired air. Permit issued under the provisions of sections 577.020 through 577.041, RSMo and 306.111 through 306.119 RSMo.

DATE 3/9/2020

NUMBER 200128

EXPIRES 3/9/2022

DIRECTOR OF STATE PUBLIC HEALTH LABORATORY

DIRECTOR OF DEPARTMENT OF HEALTH AND SENIOR SERVICES



STATE OF MISSOURI
DEPARTMENT OF HEALTH AND SENIOR SERVICES
BREATH ALCOHOL PROGRAM

INSTRUMENT OPERATOR CARD

The named cardholder is authorized to operate an evidential breath alcohol instrument for the determination of the alcoholic content in breath form of expired air in Missouri.

Operator HUTCHINSON, ANDREW
Permit No 200128
Date Issued 3/9/2020 **Date Expires** 3/9/2022

