

RECEIVED

By Tracy Crews at 10:38 am, Jul 27, 2020

MISSOURI DEPARTMENT OF HEALTH AND SENIOR SERVICES
STATE PUBLIC HEALTH LABORATORY**ALCO-SENSOR IV WITH PRINTER MAINTENANCE REPORT**

REPORT #7

Complete this report in duplicate at the time of the regular monthly preventative maintenance check, and whenever instrument is repaired. Send copy to Department of Health and Senior Services; retain original in department file.

ALCO SENSOR IV SN 111754 PRINTER SN 09B.3589.480 DATE OF INSPECTION 07/25/2020

LOCATION OF INSTRUMENT (STREET AND CITY) Ray County Zone Office, 102 W N Main St. Richmond, Missouri 64085 TIME OF INSPECTION 11:33 pm

CHECKLIST: Place a mark in the box by each item if found to be satisfactory or if operating within established limits. (Write in observed values where determined.) Unmarked items must be corrected before using instrument. DIGITAL READOUT (ALL ELEMENTS OPERATIONAL) TEMPERATURE OF ALCO SENSOR (10°C - 40°C) PRINTER WORKING PROPERLY TIME AND DATE DISPLAYING PROPERLY**BREATH ALCOHOL ACCURACY STANDARDS** SIMULATOR SOLUTION COMPRESSED ETHANOL-GAS MIXTURE STANDARD SUPPLIER RepCo Marketing Co LOT # 19001 EXP. DATE 03/12/2021 SIMULATOR TEMPERATURE (34°C ± 0.2°C) 34.00 SIMULATOR SN MP2125 SIMULATOR EXP DATE 02/13/2021 CALIBRATION CHECK - (ONLY ONE STANDARD IS TO BE USED PER MAINTENANCE REPORT)

Run three tests using a standard solution. All three tests must be within ±5% of the standard value and must have a spread of .005 or less. Check the box corresponding to the standard solution being used. (PRINTOUT ATTACHED)

 0.100% STANDARD - MUST READ BETWEEN 0.095% and 0.105% INCLUSIVE 0.080% STANDARD - MUST READ BETWEEN 0.076% and 0.084% INCLUSIVE 0.040% STANDARD - MUST READ BETWEEN 0.038% and 0.042% INCLUSIVE

TEST 1 • .099

TEST 2 • .098

TEST 3 • .098

 RFI DETECTOR OPERATING**INDICATE THE NUMBER OF BREATH TESTS IN THE FOLLOWING RANGES SINCE THE LAST MAINTENANCE REPORT:
(DO NOT INCLUDE SELF-ADMINISTERED TESTS)**

REFUSALS 0 (0-.04) 0 (.05-.09) 0 (.10-.14) 0 (.15-.19) 0 (OVER .19) 0

List any new parts and describe any alteration or modification that was made to restore the instrument to operate satisfactorily and within established limits (use other side if necessary).

INSPECTING OFFICERSIGNATURE
*Tpr. A. M. Beckwith*PRINT NAME
Tpr. A. M. BeckwithTYPE II PERMIT NUMBER/EXPIRATION DATE
290049 / 3/1/2021TELEPHONE NUMBER
(816) 622-0800**Return completed report to the:** Breath Alcohol Program, MO Department of Health and Senior Services, Southeast District Office
2875 James Boulevard
Poplar Bluff, MO 63901



STATE OF MISSOURI
 DEPARTMENT OF HEALTH AND SENIOR SERVICES
 BREATH ALCOHOL PROGRAM

2

PERMIT
TYPE II
ANDREW BECKWITH

is hereby authorized to instruct and supervise operators, train instructors, inspect, calibrate, perform field service and repairs, and operate the following breath analyzer(s):

ALCO-SENSOR IV WITH PRINTER, INTOX DMT

for the determination of the alcoholic content of blood from a sample of expired air. Permit issued under the provisions of sections 577.020 through 577.041, RSMo and 306.111 through 306.119 RSMo.

DATE **3/1/2019**

NUMBER **290049**

EXPIRES **3/1/2021**

DIRECTOR OF STATE PUBLIC HEALTH LABORATORY

DIRECTOR OF DEPARTMENT OF HEALTH AND SENIOR SERVICES

MO 160.0771 (5-10)

LAB 4 (REV 10)

 STATE OF MISSOURI
 DEPARTMENT OF HEALTH AND SENIOR SERVICES
 BREATH ALCOHOL PROGRAM

INSTRUMENT OPERATOR CARD

The named cardholder is authorized to operate an evidential breath alcohol instrument for the determination of the alcoholic content in breath form of expired air in Missouri.

Operator BECKWITH, ANDREW
 Permit No 290049
 Date Issued 3/1/2019 Date Expires 3/1/2021



RepCo

RepCo Marketing Co
3101-188 Stony Brook Drive
Raleigh, NC 27604
888-828-0227

CERTIFICATE OF ANALYSIS

MANUFACTURER AND SUPPLIER: RepCo Marketing Co.
LOT NUMBER: 19001
EXPIRATION DATE: March 12, 2021 at 11:59 p.m.

RepCo Marketing Co. certifies the following:


RepCo Marketing Co. prepared, tested and supplied Lot Number 19001 of Alcohol Certified Solution for simulators. This solution was manufactured and tested by RepCo Marketing Co., with confirmation by accredited institution, Data Resources Inc., using NIST standards. Random samples were analyzed by Data Resources Inc. utilizing a gas chromatograph and found to contain .1210 gms/dl +/- .003 gms/dl wt/vol ethanol (95% Confidence).

The alcohol and distilled water used in the solution were found to be free of any interfering substance.

This solution will produce a vapor alcohol value of .100 +/- 3% gms/210L breath when heated to 34 Degrees Celsius +/- 0.2 Degrees Celsius in a simulator (95% Confidence).

The date of manufacture for this lot number is March 13, 2019
The expiration date for this lot number is March 12, 2021 at 11:59 p.m.

This document is a true representation of the original Certificate of Analysis.


Alma Palmer, Operations Manager
RepCo Marketing Co.

AS IV Serial no: 111754
Version no: 532B

TEST RECORD 00165

Temp	Date	Time	s/ 210L
------	------	------	------------

Air Blank:
07/25/20 23:38 .000
Calibration Check:
22 07/25/20 23:38 .099

Subject Name

Maintenance

Subject I.D.

290049 3/1/21

Operator Name, I.D.

Beckwith 314

Location

Ray County Zone Office

AS IV Serial no: 111754
Version no: 532B

TEST RECORD 00166

Temp	Date	Time	s/ 210L
------	------	------	------------

Air Blank:
07/25/20 23:41 .000
Calibration Check:
23 07/25/20 23:41 .098

Subject Name

Maintenance

Subject I.D.

290049 3/1/21

Operator Name, I.D.

Beckwith 314

Location

Ray County Zone Office

AS IV Serial no: 111754
Version no: 532B

TEST RECORD 00167

Temp	Date	Time	s/ 210L
------	------	------	------------

Air Blank:
07/25/20 23:43 .000
Calibration Check:
23 07/25/20 23:43 .098

Subject Name

Maintenance

Subject I.D.

290049 3/1/21

Operator Name, I.D.

Beckwith 314

Location

Ray County Zone Office

AS IV Serial no: 111754
Version no: 532B

TEST RECORD 00168

Temp	Date	Time	s/ 210L
------	------	------	------------

VOID: RFI
12 07/25/20 23:45

Subject Name

Maintenance RFI

Subject I.D.

290049 3/1/21

Operator Name, I.D.

Beckwith 314

Location

Ray County Zone Office