



MISSOURI DEPARTMENT OF HEALTH AND SENIOR SERVICES  
 STATE PUBLIC HEALTH LABORATORY  
**ALCO-SENSOR IV WITH PRINTER MAINTENANCE REPORT**

REPORT #7

Complete this report in duplicate at the time of the regular monthly preventative maintenance check, and whenever instrument is repaired. Send copy to Department of Health and Senior Services; retain original in department file.

ALCO SENSOR IV SN 111755	PRINTER SN 09B.3589.489	DATE OF INSPECTION 01/02/2020
-----------------------------	----------------------------	----------------------------------

LOCATION OF INSTRUMENT (STREET AND CITY) 1287 County Road 8890	TIME OF INSPECTION 12:41 pm
---	--------------------------------

**CHECKLIST:** Place a mark in the box by each item if found to be satisfactory or if operating within established limits. (Write in observed values where determined.) Unmarked items must be corrected before using instrument.

DIGITAL READOUT (ALL ELEMENTS OPERATIONAL)

TEMPERATURE OF ALCO SENSOR (10°C - 40°C) 23°C

PRINTER WORKING PROPERLY

TIME AND DATE DISPLAYING PROPERLY

**BREATH ALCOHOL ACCURACY STANDARDS**

SIMULATOR SOLUTION  COMPRESSED ETHANOL-GAS MIXTURE

STANDARD SUPPLIER RepCo Marketing Inc LOT # 18001 EXP. DATE 07/31/2020

SIMULATOR TEMPERATURE (34°C ± 0.2°C) 34.01 SIMULATOR SN MP2323 SIMULATOR EXP DATE 02/07/2020

CALIBRATION CHECK – (ONLY ONE STANDARD IS TO BE USED PER MAINTENANCE REPORT)

Run three tests using a standard solution. All three tests must be within ±5% of the standard value and must have a spread of .005 or less. Check the box corresponding to the standard solution being used. (PRINTOUT ATTACHED)

- 0.100% STANDARD - MUST READ BETWEEN 0.095% and 0.105% INCLUSIVE
- 0.080% STANDARD - MUST READ BETWEEN 0.076% and 0.084% INCLUSIVE
- 0.040% STANDARD - MUST READ BETWEEN 0.038% and 0.042% INCLUSIVE

TEST 1 .099

TEST 2 .099

TEST 3 .098

RFI DETECTOR OPERATING

**INDICATE THE NUMBER OF BREATH TESTS IN THE FOLLOWING RANGES SINCE THE LAST MAINTENANCE REPORT: (DO NOT INCLUDE SELF-ADMINISTERED TESTS)**

REFUSALS	2	(0-.04)	(.05-.09)	2	(.10-.14)	1	(.15-.19)	(OVER .19)	1
----------	---	---------	-----------	---	-----------	---	-----------	------------	---

List any new parts and describe any alteration or modification that was made to restore the instrument to operate satisfactorily and within established limits (use other side if necessary).

Calibrated instrument due to new simulator.

**INSPECTING OFFICER**

SIGNATURE

PRINT NAME  
 Thomas Meyer

TYPE II PERMIT NUMBER/EXPIRATION DATE  
 280276 09/13/2020

TELEPHONE NUMBER  
 (417) 469-3121

**Return completed report to the:** Breath Alcohol Program, MO Department of Health and Senior Services, Southeast District Office  
 2875 James Boulevard  
 Poplar Bluff, MO 63901

AS IV Serial no: 111755  
Version no: 532B

TEST RECORD 00188

Temp Date Time 210L s/  
Air Blank: 01/02/20 12:44 .000  
Calibration: 23 01/02/20 12:44 .100  
Subject Name

Subject I.D.

Operator Name, I.D.

T. MEYER 080276  
Location

1287 COUNTY ROAD 8898

WEST PLAINS

AS IV Serial no: 111755  
Version no: 532B

TEST RECORD 00189

Temp Date Time 210L s/  
Air Blank: 01/02/20 12:52 .000  
Calibration Check: 23 01/02/20 12:52 .099  
Subject Name

Subject I.D.

Operator Name, I.D.

T. MEYER 080276  
Location

1287 COUNTY ROAD 8898

WEST PLAINS

AS IV Serial no: 111755  
Version no: 532B

TEST RECORD 00190

Temp Date Time 210L s/  
Air Blank: 01/02/20 12:54 .000  
Calibration Check: 23 01/02/20 12:54 .099  
Subject Name

Subject I.D.

Operator Name, I.D.

T. MEYER 080276  
Location

1287 COUNTY ROAD 8898

WEST PLAINS

AS IV Serial no: 111755  
Version no: 532B

TEST RECORD 00191

Temp Date Time 210L s/  
Air Blank: 01/02/20 13:07 .000  
Calibration Check: 23 01/02/20 13:07 .098  
Subject Name

Subject I.D.

Operator Name, I.D.

T. MEYER 080276  
Location

1287 COUNTY ROAD 8898

WEST PLAINS

AS IV Serial no: 111755  
Version no: 5928

TEST RECORD - REPRINT

TEST RECORD 03192

Temp	Date	Time	s/
VOIDR PPT	12 01/02/20	13:59	

Subject Name

Subject I.D.

Operator Name, I.D.

T. MEYER 280276  
Location

1287 COUNTY ROAD 8896

WEST PLAINS



STATE OF MISSOURI  
DEPARTMENT OF HEALTH AND SENIOR SERVICES  
BREATH ALCOHOL PROGRAM



**PERMIT**  
**TYPE II**  
**THOMAS W. MEYER**

is hereby authorized to instruct and supervise operators, train instructors, inspect, calibrate, perform field service and repairs, and operate the following breath analyzer(s):


**ALCO-SENSOR IV WITH PRINTER, INTOX DMT**

for the determination of the alcoholic content of blood from a sample of expired air. Permit issued under the provisions of sections 577.020 through 577.041, RSMo and 306.111 through 306.119 RSMo.

DATE 9/13/2018

NUMBER 280276

EXPIRES 9/13/2020

  
DIRECTOR OF STATE PUBLIC HEALTH LABORATORY

  
DIRECTOR OF DEPARTMENT OF HEALTH AND SENIOR SERVICES



**STATE OF MISSOURI**  
DEPARTMENT OF HEALTH AND SENIOR SERVICES  
BREATH ALCOHOL PROGRAM

**INSTRUMENT OPERATOR CARD**

The named cardholder is authorized to operate an evidential breath alcohol instrument for the determination of the alcoholic content in breath form of expired air in Missouri.

Operator MEYER, THOMAS  
Permit No 280276  
Date Issued 9/13/2018 Date Expires 9/13/2020





RepCo

RepCo Marketing Co  
3101-188 Stony Brook Drive  
Raleigh, NC 27604  
888-828-0227

CERTIFICATE OF ANALYSIS

MANUFACTURER AND SUPPLIER: RepCo Marketing Co.  
LOT NUMBER: 18001  
EXPIRATION DATE: July 31, 2020 at 11:59 p.m.

RepCo Marketing Co. certifies the following:

RepCo Marketing Co. prepared, tested and supplied Lot Number 18001 of Alcohol Certified Solution for simulators. This solution was manufactured and tested by RepCo Marketing Co., with confirmation by accredited institution, Data Resources Inc., using NIST standards. Random samples were analyzed by Data Resources Inc. utilizing a gas chromatograph and found to contain .1206 gms/dl  $\pm$  .003 gms/dl wt/vol ethanol (95% Confidence).

The alcohol and distilled water used in the solution were found to be free of any interfering substance.

This solution will produce a vapor alcohol value of .100  $\pm$  3% gms/210L breath when heated to 34 Degrees Celsius  $\pm$  0.2 Degrees Celsius in a simulator (95% Confidence).

The date of manufacture for this lot number is August 1, 2018  
The expiration date for this lot number is July 31, 2020 at 11:59 p.m.

This document is a true representation of the original Certificate of Analysis.

Alma Palmer  
Alma Palmer, Operations Manager  
RepCo Marketing Co.