



MISSOURI DEPARTMENT OF HEALTH AND SENIOR SERVICES
STATE PUBLIC HEALTH LABORATORY
ALCO-SENSOR IV WITH PRINTER MAINTENANCE REPORT

REPORT #7

Complete this report in duplicate at the time of the regular monthly preventative maintenance check, and whenever instrument is repaired. Send copy to Department of Health and Senior Services; retain original in department file.

ALCO SENSOR IV SN 111766	PRINTER SN 09B.3589.48	DATE OF INSPECTION 6-18-2020
LOCATION OF INSTRUMENT (STREET AND CITY) 1915 W. Arrow St Marshall, Mo 65340		TIME OF INSPECTION 16.25

CHECKLIST: Place a mark in the box by each item if found to be satisfactory or if operating within established limits. (Write in observed values where determined.) Unmarked items must be corrected before using instrument.

DIGITAL READOUT (ALL ELEMENTS OPERATIONAL)

TEMPERATURE OF ALCO SENSOR (10°C - 40°C)

PRINTER WORKING PROPERLY

TIME AND DATE DISPLAYING PROPERLY

BREATH ALCOHOL ACCURACY STANDARDS

SIMULATOR SOLUTION

COMPRESSED ETHANOL-GAS MIXTURE

STANDARD SUPPLIER Intoximeters

LOT # AG912204 EXP. DATE 05-02-2021

SIMULATOR TEMPERATURE (34°C ± 0.2°C) _____

SIMULATOR SN _____ SIMULATOR EXP DATE _____

CALIBRATION CHECK – (ONLY ONE STANDARD IS TO BE USED PER MAINTENANCE REPORT)

Run three tests using a standard solution. All three tests must be within ±5% of the standard value and must have a spread of .005 or less. Check the box corresponding to the standard solution being used. (PRINTOUT ATTACHED)

0.100% STANDARD - MUST READ BETWEEN 0.095% and 0.105% INCLUSIVE

0.080% STANDARD - MUST READ BETWEEN 0.076% and 0.084% INCLUSIVE

0.040% STANDARD - MUST READ BETWEEN 0.038% and 0.042% INCLUSIVE

TEST 1 ← 0.097

TEST 2 ← 0.099

TEST 3 ← 0.099

RFI DETECTOR OPERATING

**INDICATE THE NUMBER OF BREATH TESTS IN THE FOLLOWING RANGES SINCE THE LAST MAINTENANCE REPORT:
(DO NOT INCLUDE SELF-ADMINISTERED TESTS)**

REFUSALS 0 | (0-.04) 0 | (.05-.09) 0 | (.10-.14) 0 | (.15-.19) 0 | (OVER .19) 0

List any new parts and describe any alteration or modification that was made to restore the instrument to operate satisfactorily and within established limits (use other side if necessary).

INSPECTING OFFICER

SIGNATURE

▶

PRINT NAME

Richard Kendall

TYPE II PERMIT NUMBER/EXPIRATION DATE

280261/ 08-29-2020

TELEPHONE NUMBER

660-886-5511

Return completed report to the: Breath Alcohol Program, MO Department of Health and Senior Services, Southeast District Office
2875 James Boulevard
Poplar Bluff, MO 63901

AS IV Serial no: 111766
Version no: 532B

TEST RECORD 00186

Temp Date Time ^{s/} 210L

Air Blank:
06/18/20 20:49 .000
Calibration Check:
33 06/18/20 20:49 .097

Subject Name

TEST 1

Subject I.D.

TEST 1

Operator Name, I.D.

RICHARD KENDAN 280241

Location

1915 W ARROW

SCSD

AS IV Serial no: 111766
Version no: 532B

TEST RECORD 00187

Temp Date Time ^{s/} 210L

Air Blank:
06/18/20 20:51 .000
Calibration Check:
33 06/18/20 20:51 .099

Subject Name

TEST 2

Subject I.D.

TEST 2

Operator Name, I.D.

RICHARD KENDAN 280241

Location

1915 W ARROW

SCSD

AS IV Serial no: 111766
Version no: 532B

TEST RECORD 00188

Temp Date Time ^{s/} 210L

Air Blank:
06/18/20 20:52 .000
Calibration Check:
33 06/18/20 20:52 .099

Subject Name

TEST - 3

Subject I.D.

TEST 3

Operator Name, I.D.

RICHARD KENDAN 280241

Location

1915 W ARROW

SCSD

AS IV Serial no: 111766
Version no: 532B

TEST RECORD 00189

Temp Date Time ^{s/} 210L

VOID: RFI
12 06/18/20 20:54

Subject Name

RFI

Subject I.D.

RFI

Operator Name, I.D.
RICHARD KENDAN 280241

RICHARD KENDAN

Location

1915 W ARROW

SCSD

AS IV Serial no: 111766
Version no: 532B

TEST RECORD 00190

Temp Date Time ^{s/} 210L

Air Blank:
06/18/20 20:55 .000
Subject Test: Auto
33 06/18/20 20:55 .000

Subject Name

RICHARD KENDAN

Subject I.D. 7417

RICHARD KENDAN

Operator Name, I.D.

RICHARD KENDAN

Location

1915 W ARROW

SCSD



Airgas USA LLC (LAB)

3500 Bernard Street

St. Louis, Mo. 63103

Ph: (314) 533-3100

Fax: (314) 533-7328

Certificate of Analysis

Customer Name

Exclusive Supplier

Intoximeters, Inc.

2081 Craig Road

St. Louis, Mo 63146

Test Date: 6-May-2019

Lot # AG912204 Model 108cadd

Exp. Date

2-May-2021

Cyl. Type

108

Component

Ethanol

Nitrogen

Certified Concentration

0.100 ± 2% BrAC (272 ppm)

Balance

Certification Traceable to N.I.S.T. RGM and to CRM Ethanol Standards:

RGM Serial No.

EB0010581

EB0010570

EB0010285

EB0010561

EB0010681

Concentration

392.1 ppm

259.8 ppm

208.0 ppm

103.6 ppm

52.12 ppm

RGM Serial No.

EB0010603

EB0010559

EB0010595

EB0010562

EB0010579

Concentration

393.0 ppm

258.2 ppm

208.3 ppm

104.2 ppm

52.81 ppm

CRM Serial No.

CC434668

CC234503

Concentration

800.0 ppm

253.0 ppm

CRM Serial No.

0056649

0056662

Concentration

390.1 ppm

150.2 ppm

Analytical Method:

NDIR

Digitally signed by Quality Control
Date: 2019.05.17 12:35:24 -05:00
Reason: Dry gas standard certification of analysis
Location: Airgas USA LLC (Lab)

Approved for Release:

Rod Marsala

ISO 17025:2005 A2LA accredited. Certificate Number 3082.06

ISO 17034:2016 A2LA accredited. Certificate Number 3082.07



STATE OF MISSOURI
DEPARTMENT OF HEALTH AND SENIOR SERVICES
BREATH ALCOHOL PROGRAM

2

**PERMIT
TYPE II**

RICHARD L. KENDALL

is hereby authorized to instruct and supervise operators, train instructors, inspect, calibrate, perform field service and repairs, and operate the following breath analyzer(s):

ALCO-SENSOR IV WITH PRINTER

for the determination of the alcoholic content of blood from a sample of expired air. Permit issued under the provisions of sections 577.020 through 577.041, RSMo and 306.111 through 306.119 RSMo.

DATE 8/29/2018

NUMBER 280261

EXPIRES 8/29/2020

MO 580-0771 (6-10)

DIRECTOR OF STATE PUBLIC HEALTH LABORATORY

DIRECTOR OF DEPARTMENT OF HEALTH AND SENIOR SERVICES

LAB-4 (R6-10)



STATE OF MISSOURI
DEPARTMENT OF HEALTH AND SENIOR SERVICES
BREATH ALCOHOL PROGRAM

INSTRUMENT OPERATOR CARD

The named cardholder is authorized to operate an evidential breath alcohol instrument for the determination of the alcoholic content in breath form of expired air in Missouri.

Operator KENDALL, RICHARD
Permit No 280261
Date Issued 8/29/2018 Date Expires 8/29/2020

