



**ALCO-SENSOR IV WITH PRINTER MAINTENANCE REPORT**

REPORT #7

Complete this report in duplicate at the time of the regular monthly preventative maintenance check, and whenever instrument is repaired. Send copy to Department of Health and Senior Services; retain original in department file.

ALCO SENSOR IV SN 111767	PRINTER SN 09B.3589.455	DATE OF INSPECTION 08/17/2020
LOCATION OF INSTRUMENT (STREET AND CITY) 501 Faraon Street Saint Joseph MO. 64501		TIME OF INSPECTION 1:01 pm

**CHECKLIST:** Place a mark in the box by each item if found to be satisfactory or if operating within established limits. (Write in observed values where determined.) Unmarked items must be corrected before using instrument.

- DIGITAL READOUT (ALL ELEMENTS OPERATIONAL)
- TEMPERATURE OF ALCO SENSOR (10°C - 40°C)
- PRINTER WORKING PROPERLY
- TIME AND DATE DISPLAYING PROPERLY

**BREATH ALCOHOL ACCURACY STANDARDS**

- SIMULATOR SOLUTION  COMPRESSED ETHANOL-GAS MIXTURE
- STANDARD SUPPLIER Intoximeters LOT # AG836201 EXP. DATE 12/28/2020
- SIMULATOR TEMPERATURE (34°C ± 0.2°C) \_\_\_\_\_ SIMULATOR SN \_\_\_\_\_ SIMULATOR EXP DATE \_\_\_\_\_

- CALIBRATION CHECK – (ONLY ONE STANDARD IS TO BE USED PER MAINTENANCE REPORT)  
Run three tests using a standard solution. All three tests must be within ±5% of the standard value and must have a spread of .005 or less. Check the box corresponding to the standard solution being used. (PRINTOUT ATTACHED)
- 0.100% STANDARD - MUST READ BETWEEN 0.095% and 0.105% INCLUSIVE
- 0.080% STANDARD - MUST READ BETWEEN 0.076% and 0.084% INCLUSIVE
- 0.040% STANDARD - MUST READ BETWEEN 0.038% and 0.042% INCLUSIVE

TEST 1 ➡ 0.097	TEST 2 ➡ 0.096	TEST 3 ➡ 0.096
----------------	----------------	----------------

- RFI DETECTOR OPERATING

**INDICATE THE NUMBER OF BREATH TESTS IN THE FOLLOWING RANGES SINCE THE LAST MAINTENANCE REPORT: (DO NOT INCLUDE SELF-ADMINISTERED TESTS)**

REFUSALS	(0-.04)	(.05-.09)	(.10-.14)	(.15-.19)	(OVER .19)
----------	---------	-----------	-----------	-----------	------------

List any new parts and describe any alteration or modification that was made to restore the instrument to operate satisfactorily and within established limits (use other side if necessary).

<b>INSPECTING OFFICER</b>	
SIGNATURE 	PRINT NAME John L. Foster
TYPE I PERMIT NUMBER/EXPIRATION DATE 290227 EXP-10/01/2021	TELEPHONE NUMBER (816) 596-8206

**Return completed report to the:** Breath Alcohol Program, MO Department of Health and Senior Services, Southeast District Office  
2875 James Boulevard  
Poplar Bluff, MO 63901

# SAINT JOSEPH POLICE DEPARTMENT MONTHLY MAINTENANCE

AS IU Serial no: 111767  
Version no: 532B

TEST RECORD 00163

Temp Date Time 210L  
s/  
Air Blank: 08/17/20 13:01 .000  
Calibration Check: 21 08/17/20 13:01 .097

Subject Name

Subject I.D.

Operator Name, I.D.

LEC

Location

AS IU Serial no: 111767  
Version no: 532B

TEST RECORD 00164

Temp Date Time 210L  
s/  
VOID: RFI 12 08/17/20 13:04

Subject Name

Subject I.D.

Operator Name, I.D.

LEC

Location

AS IU Serial no: 111767  
Version no: 532B

TEST RECORD 00165

Temp Date Time 210L  
s/  
Air Blank: 08/17/20 13:06 .000  
Calibration Check: 22 08/17/20 13:06 .096

Subject Name

Subject I.D.

Operator Name, I.D.

LEC

Location

AS IU Serial no: 111767  
Version no: 532B

TEST RECORD 00166

Temp Date Time 210L  
s/  
Air Blank: 08/17/20 13:08 .000  
Calibration Check: 22 08/17/20 13:08 .096

Subject Name

Subject I.D.

Operator Name, I.D.

LEC

Location



Airgas USA LLC (LAB)  
 3500 Bernard Street  
 St. Louis, Mo. 63103  
 Ph: (314) 533-3100  
 Fax: (314) 533-7328

## Certificate of Analysis

**Customer Name**  
 Exclusive Supplier  
 Intoximeters, Inc.  
 2081 Craig Road  
 St. Louis, Mo 63146

**Test Date:** 2-Jan-2019

**Lot # AG836201 Model 108cacc**

<b>Exp. Date</b> 28-Dec-2020	<b>Cyl. Type</b> 108	<b>Component</b> Ethanol Nitrogen	<b>Certified Concentration</b> 0.100 ± 2% BrAC (272 ppm) Balance
---------------------------------	-------------------------	---	--

**Certification Traceable to N.I.S.T. RGM and to CRM Ethanol Standards:**

<b><u>RGM Serial No.</u></b> EB0010581 EB0010570 EB0010285 EB0010561 EB0010681	<b><u>Concentration</u></b> 392.1 ppm 259.8 ppm 208.0 ppm 103.6 ppm 52.12 ppm	<b><u>RGM Serial No.</u></b> EB0010603 EB0010559 EB0010595 EB0010562 EB0010579	<b><u>Concentration</u></b> 393.0 ppm 258.2 ppm 208.3 ppm 104.2 ppm 52.81 ppm
<b><u>CRM Serial No.</u></b> CC434668 CC234503	<b><u>Concentration</u></b> 800.0 ppm 253.0 ppm	<b><u>CRM Serial No.</u></b> 0056649 0056662	<b><u>Concentration</u></b> 390.1 ppm 150.2 ppm

**Analytical Method:** NDIR

Digitally signed by Quality Control  
 Date: 2019.01.02 18:04:08 -06:00  
 Reason: Dry gas standard certification of analysis  
 Location: Airgas USA LLC (Lab)

Approved for Release:

Rod Marsala

**ISO 17025:2005 A2LA accredited. Certificate Number 3082.06**  
**ISO 17034:2016 A2LA accredited. Certificate Number 3082.07**



MISSOURI DEPARTMENT OF HEALTH AND SENIOR SERVICES  
 STATE PUBLIC HEALTH LABORATORY  
**ALCO-SENSOR IV WITH PRINTER MAINTENANCE REPORT**

REPORT #7

Complete this report in duplicate at the time of the regular monthly preventative maintenance check, and whenever instrument is repaired. Send copy to Department of Health and Senior Services; retain original in department file.

ALCO SENSOR IV SN 111767	PRINTER SN 09B.3589.455	DATE OF INSPECTION 07/21/2020
LOCATION OF INSTRUMENT (STREET AND CITY) 501 Faraon Street Saint Joseph MO. 64501		TIME OF INSPECTION 10:02 am

**CHECKLIST:** Place a mark in the box by each item if found to be satisfactory or if operating within established limits. (Write in observed values where determined.) Unmarked items must be corrected before using instrument.

DIGITAL READOUT (ALL ELEMENTS OPERATIONAL)

TEMPERATURE OF ALCO SENSOR (10°C - 40°C)

PRINTER WORKING PROPERLY

TIME AND DATE DISPLAYING PROPERLY

**BREATH ALCOHOL ACCURACY STANDARDS**

SIMULATOR SOLUTION

COMPRESSED ETHANOL-GAS MIXTURE

STANDARD SUPPLIER Intoximeters LOT # AG836201 EXP. DATE 12/28/2020

SIMULATOR TEMPERATURE (34°C ± 0.2°C) SIMULATOR SN SIMULATOR EXP DATE

CALIBRATION CHECK – (ONLY ONE STANDARD IS TO BE USED PER MAINTENANCE REPORT)

Run three tests using a standard solution. All three tests must be within ±5% of the standard value and must have a spread of .005 or less. Check the box corresponding to the standard solution being used. (PRINTOUT ATTACHED)

0.100% STANDARD - MUST READ BETWEEN 0.095% and 0.105% INCLUSIVE

0.080% STANDARD - MUST READ BETWEEN 0.076% and 0.084% INCLUSIVE

0.040% STANDARD - MUST READ BETWEEN 0.038% and 0.042% INCLUSIVE

TEST 1 ← 0.098

TEST 2 ← 0.097

TEST 3 ← 0.097

RFI DETECTOR OPERATING

**INDICATE THE NUMBER OF BREATH TESTS IN THE FOLLOWING RANGES SINCE THE LAST MAINTENANCE REPORT: (DO NOT INCLUDE SELF-ADMINISTERED TESTS)**

REFUSALS	(0-.04)	(.05-.09)	(.10-.14)	(.15-.19)	(OVER .19)
----------	---------	-----------	-----------	-----------	------------

List any new parts and describe any alteration or modification that was made to restore the instrument to operate satisfactorily and within established limits (use other side if necessary).

**INSPECTING OFFICER**

SIGNATURE

PRINT NAME

John L. Foster

TYPE II PERMIT NUMBER/EXPIRATION DATE

29022 EXP-10/01/2021

TELEPHONE NUMBER

(816) 596-8206

**Return completed report to the:** Breath Alcohol Program, MO Department of Health and Senior Services, Southeast District Office  
 2875 James Boulevard  
 Poplar Bluff, MO 63901



STATE OF MISSOURI  
DEPARTMENT OF HEALTH AND SENIOR SERVICES  
BREATH ALCOHOL PROGRAM

2

**PERMIT  
TYPE II**

**JOHN L. FOSTER**

is hereby authorized to instruct and supervise operators, train instructors, inspect, calibrate, perform field service and repairs, and operate the following breath analyzer(s):

**ALCO-SENSOR IV WITH PRINTER, INTOX EC/IR II**

for the determination of the alcoholic content of blood from a sample of expired air. Permit issued under the provisions of sections 577.020 through 577.041, RSMo and 306.111 through 306.119 RSMo.

DATE 10/1/2019

NUMBER 290227

EXPIRES 10/1/2021

DIRECTOR OF STATE PUBLIC HEALTH LABORATORY

DIRECTOR OF DEPARTMENT OF HEALTH AND SENIOR SERVICES

MO 580-0771 (6-10)

LAB-4 (R6-10)



**STATE OF MISSOURI**  
DEPARTMENT OF HEALTH AND SENIOR SERVICES  
BREATH ALCOHOL PROGRAM

**INSTRUMENT OPERATOR CARD**

*The named cardholder is authorized to operate an evidential breath alcohol instrument for the determination of the alcoholic content in breath form of expired air in Missouri.*

**Operator** FOSTER, JOHN  
**Permit No** 290227  
**Date Issued** 10/1/2019    **Date Expires** 10/1/2021

