



ALCO-SENSOR IV WITH PRINTER MAINTENANCE REPORT

REPORT #7

Complete this report in duplicate at the time of the regular monthly preventative maintenance check, and whenever instrument is repaired. Send copy to Department of Health and Senior Services; retain original in department file.

ALCO SENSOR IV SN 111774	PRINTER SN 09B.3589.450	DATE OF INSPECTION 02/11/2020
LOCATION OF INSTRUMENT (STREET AND CITY) 106 S. Folger St, Carrollton		TIME OF INSPECTION 9:15 am

CHECKLIST: Place a mark in the box by each item if found to be satisfactory or if operating within established limits. (Write in observed values where determined.) Unmarked items must be corrected before using instrument.

DIGITAL READOUT (ALL ELEMENTS OPERATIONAL)

TEMPERATURE OF ALCO SENSOR (10°C - 40°C)

PRINTER WORKING PROPERLY

TIME AND DATE DISPLAYING PROPERLY

BREATH ALCOHOL ACCURACY STANDARDS

SIMULATOR SOLUTION COMPRESSED ETHANOL-GAS MIXTURE

STANDARD SUPPLIER Guth Labs LOT # 19370 EXP. DATE 12/09/2021

SIMULATOR TEMPERATURE (34°C ± 0.2°C) 34.01 SIMULATOR SN MP3879 SIMULATOR EXP DATE 08/05/2020

CALIBRATION CHECK – (ONLY ONE STANDARD IS TO BE USED PER MAINTENANCE REPORT)

Run three tests using a standard solution. All three tests must be within ±5% of the standard value and must have a spread of .005 or less. Check the box corresponding to the standard solution being used. (PRINTOUT ATTACHED)

- 0.100% STANDARD - MUST READ BETWEEN 0.095% and 0.105% INCLUSIVE
- 0.080% STANDARD - MUST READ BETWEEN 0.076% and 0.084% INCLUSIVE
- 0.040% STANDARD - MUST READ BETWEEN 0.038% and 0.042% INCLUSIVE

TEST 1 → .100

TEST 2 → .099

TEST 3 → .098

RFI DETECTOR OPERATING

**INDICATE THE NUMBER OF BREATH TESTS IN THE FOLLOWING RANGES SINCE THE LAST MAINTENANCE REPORT:
(DO NOT INCLUDE SELF-ADMINISTERED TESTS)**

REFUSALS	0	(0-.04)	0	(.05-.09)	0	(.10-.14)	0	(.15-.19)	0	(OVER .19)	0
----------	---	---------	---	-----------	---	-----------	---	-----------	---	------------	---

List any new parts and describe any alteration or modification that was made to restore the instrument to operate satisfactorily and within established limits (use other side if necessary).

INSPECTING OFFICER

SIGNATURE

PRINT NAME
James H. Plaggemeyer II

TYPE II PERMIT NUMBER/EXPIRATION DATE
280236 08/14/2020

TELEPHONE NUMBER
(660) 542-2200

Return completed report to the: Breath Alcohol Program, MO Department of Health and Senior Services, Southeast District Office
2875 James Boulevard
Poplar Bluff, MO 63901



STATE OF MISSOURI
 DEPARTMENT OF HEALTH AND SENIOR SERVICES
 BREATH ALCOHOL PROGRAM



PERMIT
TYPE II

JAMES H PLAGGEMEYER II

is hereby authorized to instruct and supervise operators, train instructors, inspect, calibrate, perform field service and repairs, and operate the following breath analyzer(s):

ALCO-SENSOR IV WITH PRINTER

for the determination of the alcoholic content of blood from a sample of expired air. Permit issued under the provisions of sections 577.020 through 577.041, RSMo and 306.111 through 306.119 RSMo.

DATE 8/14/2018

DIRECTOR OF STATE PUBLIC HEALTH LABORATORY

NUMBER 280236

DIRECTOR OF DEPARTMENT OF HEALTH AND SENIOR SERVICES

EXPIRES 8/14/2020

MO-580-0771 (6-10)

LAB-4 (R6-10)

STATE OF MISSOURI
 DEPARTMENT OF HEALTH AND SENIOR SERVICES
 BREATH ALCOHOL PROGRAM

INSTRUMENT OPERATOR CARD

The named cardholder is authorized to operate an evidential breath alcohol instrument for the determination of the alcoholic content in breath form of expired air in Missouri.

Operator PLAGGEMEYER II, JAMES
 Permit No 280236
 Date issued 8/14/2018 Date Expires 8/14/2020



GUTH LABORATORIES, INC.

590 NORTH 67th STREET • HARRISBURG, PA 17111-4511 • TELEPHONE: 717-564-5470

CERTIFICATE OF ANALYSIS

Certified Alcohol Reference Solution for Simulator

Random Samples of Lot Number **19370** of Alcohol Reference Solution for Simulator were analyzed by gas chromatography on **December 10, 2019**, using a Perkin Elmer Gas Chromatograph Autosystem XL S/N: 610N9030209, and found to contain **0.1199%** (w/vol) ethyl alcohol. The expiration date for this lot number is **December 9, 2021** at 11:59 PM.

When used in a calibrated Simulator, operating at $34^{\circ}\text{C} \pm .2^{\circ}\text{C}$, this solution will give a breath alcohol analysis instrument reading of **0.100 g/210L** \pm 3%.

The alcohol and water used in this solution were free of test interfering substances.

Ted L. Pauley, President
GUTH LABORATORIES, INC.

NIST Traceability:

Testing was conducted using Cerilliant Reference Standard lot number FN04271602 whose values are traceable to NIST.

All balances are calibrated annually by an outside agency using NIST traceable weights. Calibration verification is done prior to each use utilizing NIST traceable weights.

CK 1

AS IV Serial no: 111774
Version no: 532B

TEST RECORD 00143

Temp Date Time ^{s/} 210L

Air Blank:
02/11/20 09:15 .000
Calibration Check:
20 02/11/20 09:15 .100

Subject Name

Subject I.D. #785
PLAGGEMEYER

Operator Name, I.D.

Location

CK 2

AS IV Serial no: 111774
Version no: 532B

TEST RECORD 00144

Temp Date Time ^{s/} 210L

Air Blank:
02/11/20 09:17 .000
Calibration Check:
21 02/11/20 09:17 .099

Subject Name

Subject I.D. #785
PLAGGEMEYER

Operator Name, I.D.

Location

CK 3

AS IV Serial no: 111774
Version no: 532B

TEST RECORD 00145

Temp Date Time ^{s/} 210L

Air Blank:
02/11/20 09:18 .000
Calibration Check:
22 02/11/20 09:18 .098

Subject Name

Subject I.D. #785
PLAGGEMEYER

Operator Name, I.D.

Location

RFI

AS IV Serial no: 111774
Version no: 532B

TEST RECORD 00146

Temp Date Time ^{s/} 210L

VOID: RFI
12 02/11/20 09:20

Subject Name

Subject I.D. #785
PLAGGEMEYER

Operator Name, I.D.

Location

TEST

AS IV Serial no: 111774
Version no: 532B

TEST RECORD 00147

Temp Date Time ^{s/} 210L

Air Blank:
02/11/20 09:21 .000
Subject Test: Auto
23 02/11/20 09:21 .000

Subject Name

Subject I.D. #785
PLAGGEMEYER

Operator Name, I.D.

Location