



**ALCO-SENSOR IV WITH PRINTER MAINTENANCE REPORT**

REPORT #7

Complete this report in duplicate at the time of the regular monthly preventative maintenance check, and whenever instrument is repaired. Send copy to Department of Health and Senior Services; retain original in department file.

|  |                            |                                  |
|--|----------------------------|----------------------------------|
| ALCO SENSOR IV SN<br>111774  | PRINTER SN<br>09B.3589.450 | DATE OF INSPECTION<br>09/05/2020 |
| LOCATION OF INSTRUMENT (STREET AND CITY)<br>106 S. Folger St, Carrollton |                            | TIME OF INSPECTION<br>12:37 pm   |

**CHECKLIST:** Place a mark in the box by each item if found to be satisfactory or if operating within established limits. (Write in observed values where determined.) Unmarked items must be corrected before using instrument.

DIGITAL READOUT (ALL ELEMENTS OPERATIONAL)

TEMPERATURE OF ALCO SENSOR (10°C - 40°C)

PRINTER WORKING PROPERLY

TIME AND DATE DISPLAYING PROPERLY

**BREATH ALCOHOL ACCURACY STANDARDS**

SIMULATOR SOLUTION  COMPRESSED ETHANOL-GAS MIXTURE

STANDARD SUPPLIER Guth Labs LOT # 20190 EXP. DATE 04/06/2022

SIMULATOR TEMPERATURE (34°C ± 0.2°C) 34.01 SIMULATOR SN MP3879 SIMULATOR EXP DATE 07/20/2021

CALIBRATION CHECK – (ONLY ONE STANDARD IS TO BE USED PER MAINTENANCE REPORT)

Run three tests using a standard solution. All three tests must be within ±5% of the standard value and must have a spread of .005 or less. Check the box corresponding to the standard solution being used. (PRINTOUT ATTACHED)

- 0.100% STANDARD - MUST READ BETWEEN 0.095% and 0.105% INCLUSIVE
- 0.080% STANDARD - MUST READ BETWEEN 0.076% and 0.084% INCLUSIVE
- 0.040% STANDARD - MUST READ BETWEEN 0.038% and 0.042% INCLUSIVE

TEST 1 .099

TEST 2 .100

TEST 3 .099

RFI DETECTOR OPERATING

**INDICATE THE NUMBER OF BREATH TESTS IN THE FOLLOWING RANGES SINCE THE LAST MAINTENANCE REPORT:  
(DO NOT INCLUDE SELF-ADMINISTERED TESTS)**

|          |   |         |   |           |   |           |   |           |   |            |   |
|----------|---|---------|---|-----------|---|-----------|---|-----------|---|------------|---|
| REFUSALS | 0 | (0-.04) | 0 | (.05-.09) | 0 | (.10-.14) | 0 | (.15-.19) | 0 | (OVER .19) | 0 |
|----------|---|---------|---|-----------|---|-----------|---|-----------|---|------------|---|

List any new parts and describe any alteration or modification that was made to restore the instrument to operate satisfactorily and within established limits (use other side if necessary).

Change paper in printer (operational)

**INSPECTING OFFICER**

SIGNATURE

PRINT NAME  
Brian C. Woods

TYPE II PERMIT NUMBER/EXPIRATION DATE  
200226 08/13/2022

TELEPHONE NUMBER  
(660) 542-2200

**Return completed report to the:** Breath Alcohol Program, MO Department of Health and Senior Services, Southeast District Office  
2875 James Boulevard  
Poplar Bluff, MO 63901



STATE OF MISSOURI  
DEPARTMENT OF HEALTH AND SENIOR SERVICES  
BREATH ALCOHOL PROGRAM

2

**PERMIT  
TYPE II  
BRIAN C. WOODS**

is hereby authorized to instruct and supervise operators, train instructors, inspect, calibrate, perform field service and repairs, and operate the following breath analyzer(s):

**ALCO-SENSOR IV WITH PRINTER**

for the determination of the alcoholic content of blood from a sample of expired air. Permit issued under the provisions of sections 577.020 through 577.041, RSMo and 306.111 through 306.119 RSMo.

DATE 8/13/2020

\_\_\_\_\_  
DIRECTOR OF STATE PUBLIC HEALTH LABORATORY

NUMBER 200226

EXPIRES 8/13/2022

\_\_\_\_\_  
DIRECTOR OF DEPARTMENT OF HEALTH AND SENIOR SERVICES

MO 580-0771 (6-10)

LAB-4 (F6-10)

STATE OF MISSOURI  
DEPARTMENT OF HEALTH AND SENIOR SERVICES  
BREATH ALCOHOL PROGRAM

**INSTRUMENT OPERATOR CARD**

*The named cardholder is authorized to operate an evidential breath alcohol instrument for the determination of the alcoholic content in breath form of expired air in Missouri.*

Operator WOODS, BRIAN  
Permit No 200226  
Date Issued 8/13/2020 Date Expires 8/13/2022

#####  
\* 78 87 96 05 10 57 99 55 9 2 7 1 3 3 \* ( 5 6 8 2 0 2 1 )  
#####



**GUTH LABORATORIES, INC.**

590 NORTH 67th STREET • HARRISBURG, PA 17111-4511 • TELEPHONE: 717-564-5470

## CERTIFICATE OF ANALYSIS

Certified Alcohol Reference Solution for Simulator

Random Samples of Lot Number **20190** of Alcohol Reference Solution for Simulator were analyzed by gas chromatography on **April 8, 2020**, using a Perkin Elmer Gas Chromatograph Autosystem XL S/N: 610N9030209, and found to contain **0.1206%** (w/vol) ethyl alcohol. The expiration date for this lot number is **April 6, 2022** at 11:59 PM.

When used in a calibrated Simulator, operating at  $34^{\circ}\text{C} \pm .2^{\circ}\text{C}$ , this solution will give a breath alcohol analysis instrument reading of **0.100 g/210L  $\pm$  3%**.

The alcohol and water used in this solution were free of test interfering substances.

Ted L. Pauley, President  
GUTH LABORATORIES, INC.

*NIST Traceability:*

*Testing was conducted using Cerilliant Reference Standard lot number FN02271802 whose values are traceable to NIST.*

*All balances are calibrated annually by an outside agency using NIST traceable weights. Calibration verification is done prior to each use utilizing NIST traceable weights.*

AS IV Serial no: 111774  
Version no: 532B

TEST RECORD 00185

Temp Date Time <sup>s/</sup> 210L

Air Blank:  
09/05/20 12:37 .000  
Calibration Check:  
21 09/05/20 12:37 .099

Subject Name  
Test #1

Subject I.D.

Operator Name, I.D.  
B.C. Woods, 782  
Location

AS IV Serial no: 111774  
Version no: 532B

TEST RECORD 00186

Temp Date Time <sup>s/</sup> 210L

Air Blank:  
09/05/20 12:39 .000  
Calibration Check:  
22 09/05/20 12:39 .100

Subject Name  
Test #2

Subject I.D.

Operator Name, I.D.  
B.C. Woods, 782  
Location

AS IV Serial no: 111774  
Version no: 532B

TEST RECORD 00187

Temp Date Time <sup>s/</sup> 210L

Air Blank:  
09/05/20 12:41 .000  
Calibration Check:  
23 09/05/20 12:41 .099

Subject Name  
Test #3

Subject I.D.

Operator Name, I.D.  
B.C. Woods, 782  
Location

AS IV Serial no: 111774  
Version no: 532B

TEST RECORD 00188

Temp Date Time <sup>s/</sup> 210L

VOID: RFI  
12 09/05/20 12:43

Subject Name  
RFI Test

Subject I.D.

Operator Name, I.D.  
B.C. Woods, 782  
Location

AS IV Serial no: 111774  
Version no: 532B

TEST RECORD 00189

Temp Date Time <sup>s/</sup> 210L

Air Blank:  
09/05/20 12:44 .000  
Subject Test: Auto  
23 09/05/20 12:44 .000

Subject Name  
Blank Test

Subject I.D.

Operator Name, I.D.  
B.C. Woods, 782  
Location