



RECEIVED

By Tracy Crews at 12:08 pm, Aug 14, 2020

REPORT #7

Complete this report in duplicate at the time of the regular monthly preventative maintenance check, and whenever instrument is repaired. Send copy to Department of Health and Senior Services; retain original in department file.

ALCO SENSOR IV SN 111776	NAME OF AGENCY KCPD	DATE OF INSPECTION 08/07/2020
LOCATION OF INSTRUMENT (STREET AND CITY) 9701 MARION PARK DR KANSAS CITY, MO 64137		TIME OF INSPECTION 1947

CHECKLIST: Place a mark in the box by each item if found to be satisfactory or if operating within established limits. (Write in observed values where determined.) Unmarked items must be corrected before using instrument.

- DIGITAL READOUT (ALL ELEMENTS OPERATIONAL)
- TEMPERATURE OF ALCO SENSOR (10°C - 40°C)
- PRINTER WORKING PROPERLY
- TIME AND DATE DISPLAYING PROPERLY

BREATH ALCOHOL ACCURACY STANDARDS

- SIMULATOR SOLUTION COMPRESSED ETHANOL-GAS MIXTURE
- STANDARD SUPPLIER INTOXIMETERS LOT # AG905605 EXP. DATE 02/25/2021
- SIMULATOR TEMPERATURE (34°C ± 0.2°C) _____ SIM. SN _____ SIM. NIST EXP DATE _____

- CALIBRATION CHECK - (ONLY ONE STANDARD IS TO BE USED PER MAINTENANCE REPORT)
 Run three tests using a standard solution. All three tests must be within ±5% of the standard value and must have a spread of .005 or less. Check the box corresponding to the standard solution being used. (PRINTOUT ATTACHED)
 - 0.100% STANDARD - MUST READ BETWEEN 0.095% and 0.105% INCLUSIVE
 - 0.080% STANDARD - MUST READ BETWEEN 0.076% and 0.084% INCLUSIVE
 - 0.040% STANDARD - MUST READ BETWEEN 0.038% and 0.042% INCLUSIVE

TEST 1 .078	TEST 2 .080	TEST 3 .080
--------------------	--------------------	--------------------

- RFI DETECTOR OPERATING

INDICATE THE NUMBER OF BREATH TESTS IN THE FOLLOWING RANGES SINCE THE LAST MAINTENANCE REPORT: (DO NOT INCLUDE SELF-ADMINISTERED TESTS)

REFUSALS	(0-.04)	(.05-.09)	(.10-.14)	(.15-.19)	(OVER .19)
----------	---------	-----------	-----------	-----------	------------

List any new parts and describe any alteration or modification that was made to restore the instrument to operate satisfactorily and within established limits (use other side if necessary).

Instrument meets all DOHSS Standards and Guidelines.

INSPECTING OFFICER

SIGNATURE # 5266	PRINT NAME COREY CARLISLE
TYPE II PERMIT NUMBER/EXPIRATION DATE 200123 03/09/2022	TELEPHONE NUMBER ()

Return completed report to the: Breath Alcohol Program, MO Department of Health and Senior Services, Southeast District Office by mail, fax, or email.

AS IV Serial no: 111776
Version no: 532B

TEST RECORD 00134

Temp Date Time 210L s/

Air Blank: 08/07/20 19:41 .000
Calibration Check: 26 08/07/20 19:41 .078

Subject Name

Test 1

Subject I.D.

Carlisle # 200123

Operator Name: I.D.

Location

9701 Warren Park Dr

AS IV Serial no: 111776
Version no: 532B

TEST RECORD 00135

Temp Date Time 210L s/

Air Blank: 08/07/20 19:43 .000
Calibration Check: 26 08/07/20 19:43 .080

Subject Name

Test 2

Subject I.D.

Carlisle # 200123

Operator Name: I.D.

Location

9701 Warren Park Dr

AS IV Serial no: 111776
Version no: 532B

TEST RECORD 00136

Temp Date Time 210L s/

Air Blank: 08/07/20 19:45 .000
Calibration Check: 27 08/07/20 19:45 .080

Subject Name

Test 3

Subject I.D.

Carlisle # 200123

Operator Name: I.D.

Location

9701 Warren Park Dr

AS IV Serial no: 111776
Version no: 532B

TEST RECORD 00137

Temp Date Time 210L s/

VOID: RTI
12 08/07/20 19:47

Subject Name

RET

Subject I.D.

Carlisle # 200123

Operator Name: I.D.

Location

9701 Warren Park Dr



STATE OF MISSOURI
DEPARTMENT OF HEALTH AND SENIOR SERVICES
BREATH ALCOHOL PROGRAM



Airgas

Airgas USA LLC (LA#)
3500 Bernard Street
St. Louis, Mo. 63103
Ph: (314) 533-3100
Fax: (314) 533-7328

**PERMIT
TYPE II**

COREY CARLISLE

Customer Name
Exclusive Supplier
Intoximeters, Inc.
2081 Craig Road
St. Louis, Mo. 63146

Certificate of Analysis

Test Date: 26-Feb-2019

is hereby authorized to instruct and supervise operators, train instructors, inspect, calibrate, perform field service and repairs, and operate the following breath analyzer(s):

ALCO-SENSOR IV WITH PRINTER, INTOXILYZER 8000

for the determination of the alcoholic content of blood from a sample of expired air. Permit issued under the provisions of sections 577.020 through 577.041, RSMo and 306.111 through 306.119 RSMo.

Lot # AG905605 Model 108cacd

Exp. Date 26-Feb-2021
Cyl. Type 108
Component Ethanol
Nitrogen
Certified Concentration 0.082 ± 0.002 BAC (23 ppm)
Balance

Certification Traceable to N.I.S.T. RGM and to CRM Ethanol Standards:

RGM Serial No.	Concentration	RGM Serial No.	Concentration
EB0010581	392.1 ppm	EB0010579	52.81 ppm
EB0010570	259.8 ppm	EB0010562	104.2 ppm
EB0010285	208.0 ppm	EB0010595	208.3 ppm
EB0010561	103.6 ppm	EB0010559	104.2 ppm
EB0010681	52.12 ppm	EB0010579	52.81 ppm
CRM Serial No.	Concentration	CRM Serial No.	Concentration
CC434668	800.0 ppm	0056649	390.1 ppm
CC234503	253.0 ppm	0056662	150.2 ppm

Analytical Method: NDIR



Reliability verified by Quality Control
Division of the Missouri Department of Health
Reason: Dry gas standard concentration of analysis
Certification Agency: USA ILCC (LA#)

Approved for Release:

Rod Marsala
Rod Marsala

ISO 17025:2005 A2LA accredited. Certificate Number 3082.06
ISO 17034:2016 A2LA accredited. Certificate Number 3082.07