



ALCO-SENSOR IV WITH PRINTER MAINTENANCE REPORT

REPORT #7

Complete this report in duplicate at the time of the regular monthly preventative maintenance check, and whenever instrument is repaired. Send copy to Department of Health and Senior Services; retain original in department file.

ALCO SENSOR IV SN 111780	PRINTER SN 09B.3589.546	DATE OF INSPECTION 02/07/2020
-----------------------------	----------------------------	----------------------------------

LOCATION OF INSTRUMENT (STREET AND CITY) 9701 MARION PARK DR, KANSAS CITY	TIME OF INSPECTION 5:59 pm
--	-------------------------------

CHECKLIST: Place a mark in the box by each item if found to be satisfactory or if operating within established limits. (Write in observed values where determined.) Unmarked items must be corrected before using instrument.

DIGITAL READOUT (ALL ELEMENTS OPERATIONAL)

TEMPERATURE OF ALCO SENSOR (10°C - 40°C)

PRINTER WORKING PROPERLY

TIME AND DATE DISPLAYING PROPERLY

BREATH ALCOHOL ACCURACY STANDARDS

SIMULATOR SOLUTION COMPRESSED ETHANOL-GAS MIXTURE

STANDARD SUPPLIER INTOXIMETERS LOT # AG905605 EXP. DATE 02/25/2021

SIMULATOR TEMPERATURE (34°C ± 0.2°C) SIMULATOR SN SIMULATOR EXP DATE

CALIBRATION CHECK – (ONLY ONE STANDARD IS TO BE USED PER MAINTENANCE REPORT)

Run three tests using a standard solution. All three tests must be within ±5% of the standard value and must have a spread of .005 or less. Check the box corresponding to the standard solution being used. (PRINTOUT ATTACHED)

- 0.100% STANDARD - MUST READ BETWEEN 0.095% and 0.105% INCLUSIVE
- 0.080% STANDARD - MUST READ BETWEEN 0.076% and 0.084% INCLUSIVE
- 0.040% STANDARD - MUST READ BETWEEN 0.038% and 0.042% INCLUSIVE

TEST 1 ➔ .078	TEST 2 ➔ .079	TEST 3 ➔ .079
---------------	---------------	---------------

RFI DETECTOR OPERATING

INDICATE THE NUMBER OF BREATH TESTS IN THE FOLLOWING RANGES SINCE THE LAST MAINTENANCE REPORT: (DO NOT INCLUDE SELF-ADMINISTERED TESTS)

REFUSALS	0	(0-.04)	0	(.05-.09)	0	(.10-.14)	1	(.15-.19)	0	(OVER .19)	2
----------	---	---------	---	-----------	---	-----------	---	-----------	---	------------	---

List any new parts and describe any alteration or modification that was made to restore the instrument to operate satisfactorily and within established limits (use other side if necessary).

Instrument meets all DOHSS standards and guidelines.

INSPECTING OFFICER

SIGNATURE P.O. [Signature] R 5396	PRINT NAME Shawn Davis
--------------------------------------	---------------------------

TYPE II PERMIT NUMBER/EXPIRATION DATE 290088 / 04/22/2021	TELEPHONE NUMBER (816) 234-5000
--	------------------------------------

Return completed report to the: Breath Alcohol Program, MO Department of Health and Senior Services, Southeast District Office
2875 James Boulevard
Poplar Bluff, MO 63901

AS IV Serial no: 111780
Version no: 532B

TEST RECORD 00252

Temp Date Time ^{a/} 210L

Air Blank:
02/07/20 17:59 .000
Calibration Check:
20 02/07/20 17:59 .078

Subject Name

Test #1

Subject I.D.

Operator Name, I.D.

Davis 290088

Location

AS IV Serial no: 111780
Version no: 532B

TEST RECORD 00254

Temp Date Time ^{a/} 210L

Air Blank:
02/07/20 18:02 .000
Calibration Check:
22 02/07/20 18:02 .079

Subject Name

Test #3

Subject I.D.

Operator Name, I.D.

Davis 290088

Location

AS IV Serial no: 111780
Version no: 532B

TEST RECORD 00253

Temp Date Time ^{a/} 210L

Air Blank:
02/07/20 18:01 .000
Calibration Check:
21 02/07/20 18:01 .079

Subject Name

Test #2

Subject I.D.

Operator Name, I.D.

Davis 290088

Location

AS IV Serial no: 111780
Version no: 532B

TEST RECORD 00255

Temp Date Time ^{a/} 210L

VOID: RFI
12 02/07/20 18:03

Subject Name

RFI Test

Subject I.D.

Operator Name, I.D.

Davis 290088

Location



STATE OF MISSOURI
DEPARTMENT OF HEALTH AND SENIOR SERVICES
BREATH ALCOHOL PROGRAM



**PERMIT
TYPE II**

SHAWN DAVIS

is hereby authorized to instruct and supervise operators, train instructors, inspect, calibrate, perform field service and repairs, and operate the following breath analyzer(s):

ALCO-SENSOR IV WITH PRINTER, INTOXILYZER 8000

for the determination of the alcoholic content of blood from a sample of expired air. Permit issued under the provisions of sections 577.020 through 577.041, RSMo and 306.111 through 306.118 RSMo.

DATE 4/22/2019
 NUMBER 290088
 EXPIRES 4/22/2021
 MO 280-0771 (6-10)

Shawn Davis
 DIRECTOR OF STATE PUBLIC HEALTH LABORATORY

Shawn Davis
 DIRECTOR OF DEPARTMENT OF HEALTH AND SENIOR SERVICES
 LAB-4 (R6-10)

STATE OF MISSOURI
DEPARTMENT OF HEALTH AND SENIOR SERVICES
BREATH ALCOHOL PROGRAM

INSTRUMENT OPERATOR CARD

The issued operator is authorized to operate an evidential breath alcohol instrument for the determination of the alcoholic content of the breath from a sample of expired air in Missouri.

Operator: **DAVIS, SHAWN**
 Permit No: **290088**
 Date Issued: **4/22/2019** Date Expires: **4/22/2021**




Airgas USA LLC (LAB)
3500 Bernard Street
St. Louis, Mo. 63103
Ph: (314) 533-3100
Fax: (314) 533-7328

Certificate of Analysis

Test Date: 26-Feb-2019

Customer Name
Exclusive Supplier
Intoximeters, Inc.
2081 Craig Road
St. Louis, Mo 63146

Lot # AG905605 Model 108caod

Exp. Date	Cyl. Type	Component	Certified Concentration
25-Feb-2021	108	Ethanol	0.062 ± 0.002 BrAC (223 ppm)
		Nitrogen	Balance

Certification Traceable to N.I.S.T. RGM and to CRM Ethanol Standards:

RGM Serial No.	Concentration	RGM Serial No.	Concentration
EB0010581	392.1 ppm	EB0010603	393.0 ppm
EB0010570	259.8 ppm	EB0010559	258.2 ppm
EB0010285	208.0 ppm	EB0010595	208.3 ppm
EB0010561	103.6 ppm	EB0010562	104.2 ppm
EB0010681	52.12 ppm	EB0010579	52.81 ppm
CRM Serial No.	Concentration	CRM Serial No.	Concentration
CC434668	800.0 ppm	0056649	390.1 ppm
CC234503	253.0 ppm	0056662	150.2 ppm

Analytical Method: NDIR

Digitally signed by Quality Control
 Date: 2019.02.26 4:23:45 -0800
 Reason: Dry gas standard certification of analysis
 Location: Airgas USA, LLC (Lab)

Approved for Release:

Rod Marsala
Rod Marsala

ISO 17025:2005 A2LA accredited. Certificate Number 3082.06
 ISO 17034:2016 A2LA accredited. Certificate Number 3082.07