



MISSOURI DEPARTMENT OF HEALTH AND SENIOR SERVICES
 STATE PUBLIC HEALTH LABORATORY
ALCO-SENSOR IV WITH PRINTER MAINTENANCE REPORT

RECEIVED
 by Missy Green at 8:30am, Aug 17, 2020

REPORT #7

Complete this report in duplicate at the time of the regular monthly preventative maintenance check, and whenever instrument is repaired. Send copy to Department of Health and Senior Services; retain original in department file.

ALCO SENSOR IV SN <u>11752</u>	NAME OF AGENCY <u>STRAFFORD POLICE DEPT</u>	DATE OF INSPECTION <u>8-14-2020</u>
LOCATION OF INSTRUMENT (STREET AND CITY) <u>125 S. JEFFERSON STRAFFORD, MO 65757</u>		TIME OF INSPECTION <u>15:09</u>

CHECKLIST: Place a mark in the box by each item if found to be satisfactory or if operating within established limits. (Write in observed values where determined.) Unmarked items must be corrected before using instrument.

<input checked="" type="checkbox"/> DIGITAL READOUT (ALL ELEMENTS OPERATIONAL)
<input checked="" type="checkbox"/> TEMPERATURE OF ALCO SENSOR (10°C - 40°C)
<input checked="" type="checkbox"/> PRINTER WORKING PROPERLY
<input checked="" type="checkbox"/> TIME AND DATE DISPLAYING PROPERLY

BREATH ALCOHOL ACCURACY STANDARDS

<input checked="" type="checkbox"/> SIMULATOR SOLUTION	<input type="checkbox"/> COMPRESSED ETHANOL-GAS MIXTURE
<input checked="" type="checkbox"/> STANDARD SUPPLIER <u>GUTH LABS</u> LOT # <u>19341</u> EXP. DATE <u>11-18-2021</u>	
<input checked="" type="checkbox"/> SIMULATOR TEMPERATURE (34°C ± 0.2°C) <u>34.0</u> SIM. SN <u>D25379</u> SIM. NIST EXP DATE <u>5-27-2021</u>	
<input checked="" type="checkbox"/> CALIBRATION CHECK – (ONLY ONE STANDARD IS TO BE USED PER MAINTENANCE REPORT) Run three tests using a standard solution. All three tests must be within ±5% of the standard value and must have a spread of .005 or less. Check the box corresponding to the standard solution being used. (PRINTOUT ATTACHED)	
<input checked="" type="checkbox"/> 0.100% STANDARD - MUST READ BETWEEN 0.095% and 0.105% INCLUSIVE	
<input type="checkbox"/> 0.080% STANDARD - MUST READ BETWEEN 0.076% and 0.084% INCLUSIVE	
<input type="checkbox"/> 0.040% STANDARD - MUST READ BETWEEN 0.038% and 0.042% INCLUSIVE	

TEST 1 • <u>.096</u>	TEST 2 • <u>.098</u>	TEST 3 • <u>.098</u>
----------------------	----------------------	----------------------

RFI DETECTOR OPERATING

INDICATE THE NUMBER OF BREATH TESTS IN THE FOLLOWING RANGES SINCE THE LAST MAINTENANCE REPORT: (DO NOT INCLUDE SELF-ADMINISTERED TESTS)

REFUSALS	(0-.04)	(.05-.09)	(.10-.14)	(.15-.19)	(OVER .19)
----------	---------	-----------	-----------	-----------	------------

List any new parts and describe any alteration or modification that was made to restore the instrument to operate satisfactorily and within established limits (use other side if necessary).

AT TIME OF MAINTENANCE, INSTRUMENT WAS OPERATING
IN ACCORDANCE WITH DASS REGULATIONS

INSPECTING OFFICER

SIGNATURE <u>[Signature]</u>	PRINT NAME <u>KRISTEN MALLORY</u>
TYPE II PERMIT NUMBER/EXPIRATION DATE <u>290135 exp 6-21-2021</u>	TELEPHONE NUMBER <u>(417) 736-4000</u>

Return completed report to the: Breath Alcohol Program, MO Department of Health and Senior Services, Southeast District Office by mail, fax, or email.

AS IV Serial no: 111752
Version no: 532B

TEST RECORD 00609

Temp	Date	Time	g/ 210L
Air Blank:	08/14/20	15:09	.000
Calibration Check:	22 08/14/20	15:09	.096

Subject Name
TEST # 1
Subject I.D.
KRISTEN MALLORY
Operator Name, I.D.
290135 exp 6-21-21
Location
125 S. JEFFERSON
STRAFFORD, MO

AS IV Serial no: 111752
Version no: 532B

TEST RECORD 00610

Temp	Date	Time	g/ 210L
Air Blank:	08/14/20	15:11	.000
Calibration Check:	22 08/14/20	15:11	.098

Subject Name
TEST # 2
Subject I.D.
KRISTEN MALLORY
Operator Name, I.D.
290135 exp 6-21-21
Location
125 S. JEFFERSON
STRAFFORD, MO

AS IV Serial no: 111752
Version no: 532B

TEST RECORD 00611

Temp	Date	Time	g/ 210L
Air Blank:	08/14/20	15:12	.000
Calibration Check:	23 08/14/20	15:12	.098

Subject Name
TEST # 3
Subject I.D.
KRISTEN MALLORY
Operator Name, I.D.
290135 exp 6-21-21
Location
125 S. JEFFERSON
STRAFFORD, MO

AS IV Serial no: 111752
Version no: 532B

TEST RECORD 00612

Temp	Date	Time	g/ 210L
VOID: RFI	12 08/14/20	15:13	

Subject Name
RFI TEST
Subject I.D.
KRISTEN MALLORY
Operator Name, I.D.
290135 exp 6-21-21
Location
125 S. JEFFERSON
STRAFFORD, MO



GUTH LABORATORIES, INC.

590 NORTH 67th STREET • HARRISBURG, PA 17111-4511 • TELEPHONE: 717-564-5470

CERTIFICATE OF ANALYSIS

Certified Alcohol Reference Solution for Simulator

Random Samples of Lot Number **19341** of Alcohol Reference Solution for Simulator were analyzed by gas chromatography on **November 20, 2019**, using a Perkin Elmer Gas Chromatograph Autosystem XL S/N: 610N9030209, and found to contain **0.1211%** (w/vol) ethyl alcohol. The expiration date for this lot number is **November 18, 2021** at 11:59 PM.

When used in a calibrated Simulator, operating at $34^{\circ}\text{C} \pm .2^{\circ}\text{C}$, this solution will give a breath alcohol analysis instrument reading of **0.100 g/210L \pm 3%**.

The alcohol and water used in this solution were free of test interfering substances.

Ted L. Pauley, President
GUTH LABORATORIES, INC.

NIST Traceability:

Testing was conducted using Cerilliant Reference Standard lot number FN04271602 whose values are traceable to NIST.

All balances are calibrated annually by an outside agency using NIST traceable weights. Calibration verification is done prior to each use utilizing NIST traceable weights.



STATE OF MISSOURI
DEPARTMENT OF HEALTH AND SENIOR SERVICES
BREATH ALCOHOL PROGRAM



**PERMIT
TYPE II**

KRISTEN J. MALLORY

is hereby authorized to instruct and supervise operators, train instructors, inspect, calibrate, perform field service and repairs, and operate the following breath analyzer(s):

ALCO-SENSOR IV WITH PRINTER

for the determination of the alcoholic content of blood from a sample of expired air. Permit issued under the provisions of sections 577.020 through 577.041, RSMo and 306.111 through 306.119 RSMo.

DATE 6/21/2019

DIRECTOR OF STATE PUBLIC HEALTH LABORATORY

NUMBER 290135

EXPIRES 6/21/2021

DIRECTOR OF DEPARTMENT OF HEALTH AND SENIOR SERVICES

MO 580-0771 (6-10)

LAB-4 (R6-10)



STATE OF MISSOURI
DEPARTMENT OF HEALTH AND SENIOR SERVICES
BREATH ALCOHOL PROGRAM

INSTRUMENT OPERATOR CARD

The named cardholder is authorized to operate an evidential breath alcohol instrument for the determination of the alcoholic content in breath form of expired air in Missouri.

Operator MALLORY, KRISTEN
Permit No 290135
Date Issued 6/21/2019 Date Expires 6/21/2021

