

RECEIVED

By Tracy Crews at 3:44 pm, Dec 02, 2020



MISSOURI DEPARTMENT OF HEALTH AND SENIOR SERVICES
STATE PUBLIC HEALTH LABORATORY
BREATH ALCOHOL PROGRAM

INTOX EC/IR II MAINTENANCE REPORT

REPORT #3

Complete this report at the time of the regular monthly preventive maintenance check (not to exceed 35 days). Complete this report whenever the instrument is serviced or repaired and whenever it is placed into service. Retain the original and send a copy within 15 days to the Breath Alcohol Program, DHSS.

INTOX EC/IR II SN
12680

NAME OF AGENCY
SLMPD BAT VAN

DATE OF INSPECTION
12/02/2020

LOCATION OF INSTRUMENT (STREET AND CITY)
2150 S 59th St ST LOUIS

TIME OF INSPECTION
08:50 CST

CHECKLIST: Place a mark in the box by each item if found to be satisfactory or is operating within established limits. (Write in observed values where determined). Unmarked items must be corrected before using instrument.

 DIAGNOSTIC RECORD BLANK CHECK CO2 CHECK FC 1 TEMP FLOW CHECK SRC TEMP FCB CHECK DET TEMP CRC COMP CHECK BT TEMP CRC CAL CHECK STD 2 TEMP PRINT TEST ETH CHECK**BREATH ANALYZER ACCURACY STANDARDS** SIMULATOR SOLUTION COMPRESSED ETHANOL-GAS MIXTURE STANDARD SUPPLIER

INTOXIMETERS

LOT# AG911501

EXP. DATE 04/25/2021

 SIMULATOR TEMP (34°C ±0.2°C)

SIM. SN

SIM. NIST EXP DATE

 CALIBRATION CHECK - (ONLY ONE STANDARD IS TO BE USED PER MAINTENANCE REPORT)

Run three tests using a standard solution. All three tests must be within +5% of the standard value and must have a spread of .005 or less. Mark the box corresponding to the standard solution being used.

 0.10% STANDARD - MUST READ BETWEEN 0.095% AND 0.105% INCLUSIVE 0.08% STANDARD - MUST READ BETWEEN 0.076% AND 0.084% INCLUSIVE 0.04% STANDARD - MUST READ BETWEEN 0.038% AND 0.042% INCLUSIVE

TEST 1 0.078 g/210L

TEST 2 0.077 g/210L

TEST 3 0.077 g/210L

INDICATE THE NUMBER OF BREATH TESTS IN THE FOLLOWING RANGES SINCE THE LAST MAINTENANCE REPORT:

REFUSALS 0

0-.04 0

.05-.09 0

.10-.14 0

.15-.19 0

OVER .19 0

LIST ANY NEW PARTS AND DESCRIBE ANY ALTERATION OR MODIFICATION THAT WAS MADE TO RESTORE THE INSTRUMENT TO OPERATE SATISFACTORILY AND WITHIN ESTABLISHED LIMITS (USE OTHER SIDE IF NECESSARY).

INSPECTING OFFICER

SIGNATURE

PRINT FULL NAME

TYPE II PERMIT NUMBER

EXPIRATION DATE

TELEPHONE NUMBER

290177

08/09/2021

(314) 444-5345

RETURN COMPLETED REPORT TO THE:

Breath Alcohol Program, Missouri Department of Health and Senior Services,
by mail, fax, or e-mail