



MISSOURI DEPARTMENT OF HEALTH AND SENIOR SERVICES  
STATE PUBLIC HEALTH LABORATORY  
BREATH ALCOHOL PROGRAM

## INTOX EC/IR II MAINTENANCE REPORT

REPORT #3

Complete this report at the time of the regular monthly preventive maintenance check (not to exceed 35 days). Complete this report whenever the instrument is serviced or repaired and whenever it is placed into service. Retain the original and send a copy within 15 days to the Breath Alcohol Program, DHSS.											
INTOX EC/IR II SN 12823		NAME OF AGENCY Pevely Police Department		DATE OF INSPECTION 11/21/2020							
LOCATION OF INSTRUMENT (STREET AND CITY) 401 Main Street Pevely				TIME OF INSPECTION 14:50 CST							
CHECKLIST: Place a mark in the box by each item if found to be satisfactory or is operating within established limits. (write in observed values where determined). Unmarked items must be corrected before using instrument.											
<input checked="" type="checkbox"/> DIAGNOSTIC RECORD											
<input checked="" type="checkbox"/> BLANK CHECK		<input checked="" type="checkbox"/> CO2 CHECK									
<input checked="" type="checkbox"/> FC 1 TEMP		<input checked="" type="checkbox"/> FLOW CHECK									
<input checked="" type="checkbox"/> SRC TEMP		<input checked="" type="checkbox"/> FCB CHECK									
<input checked="" type="checkbox"/> DET TEMP		<input checked="" type="checkbox"/> CRC COMP CHECK									
<input checked="" type="checkbox"/> BT TEMP		<input checked="" type="checkbox"/> CRC CAL CHECK									
<input checked="" type="checkbox"/> STD 2 TEMP		<input checked="" type="checkbox"/> PRINT TEST									
<input checked="" type="checkbox"/> ETH CHECK											
BREATH ANALYZER ACCURACY STANDARDS											
<input type="checkbox"/> SIMULATOR SOLUTION		<input checked="" type="checkbox"/> COMPRESSED ETHANOL-GAS MIXTURE									
<input checked="" type="checkbox"/> STANDARD SUPPLIER Intoximeters		LOT# AG933901		EXP DATE 12/05/2021							
<input type="checkbox"/> SIMULATOR TEMP (34°C ±0.2°C)		SIM. SN		SIM. NIST EXP DATE							
<input checked="" type="checkbox"/> CALIBRATION CHECK - (ONLY ONE STANDARD IS TO BE USED PER MAINTENANCE REPORT)											
Run three tests using a standard solution. All three tests must be within +5% of the standard value and must have a spread of .005 or less. Mark the box corresponding to the standard solution being used.											
<input type="checkbox"/> 0.10% STANDARD - MUST READ BETWEEN 0.095% AND 0.105% INCLUSIVE											
<input checked="" type="checkbox"/> 0.08% STANDARD - MUST READ BETWEEN 0.076% AND 0.084% INCLUSIVE											
<input type="checkbox"/> 0.04% STANDARD - MUST READ BETWEEN 0.038% AND 0.042% INCLUSIVE											
TEST 1 = 0.080 g/210L		TEST 2 = 0.080 g/210L		TEST 3 = 0.080 g/210L							
INDICATE THE NUMBER OF BREATH TESTS IN THE FOLLOWING RANGES SINCE THE LAST MAINTENANCE REPORT:											
REFUSALS	1	0-.04	0	.05-.09	0	.10-.14	1	.15-.19	1	OVER .19	0
LIST ANY NEW PARTS AND DESCRIBE ANY ALTERATION OR MODIFICATION THAT WAS MADE TO RESTORE THE INSTRUMENT TO OPERATE SATISFACTORILY AND WITHIN ESTABLISHED LIMITS (USE OTHER SIDE IF NECESSARY).											
INSPECTING OFFICER											
SIGNATURE 			PRINT FULL NAME ST CLAIR, JOE								
TYPE II PERMIT NUMBER 290022		EXPIRATION DATE 01/25/2021		TELEPHONE NUMBER ( 636 ) 475-5301							
RETURN COMPLETED REPORT TO THE: Breath Alcohol Program, Missouri Department of Health and Senior Services, by mail, fax, or e-mail											

# Airgas

Airgas USA LLC (LAB)  
3500 Bernard Street  
St. Louis, Mo. 63103  
Ph: (314) 533-3100  
Fax: (314) 533-7328

## Certificate of Analysis

Customer Name  
Exclusive Supplier  
Intoximeters, Inc.  
2081 Craig Road  
St. Louis, Mo 63146

Test Date: 6-Dec-2019

**Lot # AG933901 Model 108cacc**

<u>Exp. Date</u>	<u>Cyl. Type</u>	<u>Component</u>	<u>Certified Concentration</u>
5-Dec-2021	108	Ethanol Nitrogen	0.080 ± 0.002 BrAC (218 ppm) Balance

Certification Traceable to N.I.S.T. RGM and to CRM Ethanol Standards:

<u>RGM Serial No.</u>	<u>Concentration</u>	<u>RGM Serial No.</u>	<u>Concentration</u>
EB0010581	392.1 ppm	EB0010603	393.0 ppm
EB0010570	259.8 ppm	EB0010559	258.2 ppm
EB0010285	208.0 ppm	EB0010595	208.3 ppm
EB0010561	103.6 ppm	EB0010562	104.2 ppm
EB0010681	52.12 ppm	EB0010579	52.81 ppm

  

<u>CRM Serial No.</u>	<u>Concentration</u>	<u>CRM Serial No.</u>	<u>Concentration</u>
CC434668	800.0 ppm	0056649	390.1 ppm
CC234503	253.0 ppm	0056662	150.2 ppm

Analytical Method: NDIR

Digitally signed by Quality Control  
Date: 2019.12.09 11:14:65 -06:00  
Reason: Dry gas standard certification of analyte  
Location: Airgas USA LLC (Lab)

Approved for Release: \_\_\_\_\_

*Rod Marsala*  
Rod Marsala

ISO 17025:2005 A2LA accredited. Certificate Number 3082.06  
ISO 17034:2016 A2LA accredited. Certificate Number 3082.07



STATE OF MISSOURI  
DEPARTMENT OF HEALTH AND SENIOR SERVICES  
BREATH ALCOHOL PROGRAM



**PERMIT**  
**TYPE II**  
**JOE L ST CLAIR**

is hereby authorized to instruct and supervise operators, train instructors, inspect, calibrate, perform field service and repairs, and operate the following breath analyzer(s):

**INTOX EC/IR II**

for the determination of the alcoholic content of blood from a sample of expired air. Permit issued under the provisions of sections 577.020 through 577.041, RSMo and 306.111 through 306.119 RSMo.

DATE 1/25/2019

NUMBER 290022

EXPIRES 1/25/2021

DIRECTOR OF STATE PUBLIC HEALTH LABORATORY

DIRECTOR OF DEPARTMENT OF HEALTH AND SENIOR SERVICES