



MISSOURI DEPARTMENT OF HEALTH AND SENIOR SERVICES
STATE PUBLIC HEALTH LABORATORY
BREATH ALCOHOL PROGRAM

INTOX EC/IR II MAINTENANCE REPORT

REPORT #3

RECEIVED
BY MISSOURI DEPT. OF HEALTH AND SENIOR SERVICES
6/11/20

Complete this report at the time of the regular monthly preventive maintenance check (not to exceed 35 days). Complete this report whenever the instrument is serviced or repaired and whenever it is placed into service. Retain the original and send a copy within 15 days to the Breath Alcohol Program, DHSS.

| | | |
|----------------------------|------------------------------|----------------------------------|
| INTOX EC/IR II SN 12946 | NAME OF AGENCY MOBERLY PD | DATE OF INSPECTION 05/22/2020 |
|----------------------------|------------------------------|----------------------------------|

| | |
|---|---------------------------------|
| LOCATION OF INSTRUMENT (STREET AND CITY) 300 N. CLARK STREET MOBERLY | TIME OF INSPECTION 09:55 CDT |
|---|---------------------------------|

CHECKLIST: Place a mark in the box by each item if found to be satisfactory or is operating within established limits. (Write in observed values where determined). Unmarked items must be corrected before using instrument.

DIAGNOSTIC RECORD

| | |
|---|--|
| <input checked="" type="checkbox"/> BLANK CHECK | <input checked="" type="checkbox"/> CO2 CHECK |
| <input checked="" type="checkbox"/> FC 1 TEMP | <input checked="" type="checkbox"/> FLOW CHECK |
| <input checked="" type="checkbox"/> SRC TEMP | <input checked="" type="checkbox"/> FCB CHECK |
| <input checked="" type="checkbox"/> DET TEMP | <input checked="" type="checkbox"/> CRC COMP CHECK |
| <input checked="" type="checkbox"/> BT TEMP | <input checked="" type="checkbox"/> CRC CAL CHECK |
| <input checked="" type="checkbox"/> STD 2 TEMP | <input checked="" type="checkbox"/> PRINT TEST |
| <input checked="" type="checkbox"/> ETH CHECK | |

BREATH ANALYZER ACCURACY STANDARDS

| | |
|---|--|
| <input type="checkbox"/> SIMULATOR SOLUTION | <input checked="" type="checkbox"/> COMPRESSED ETHANOL-GAS MIXTURE |
| <input checked="" type="checkbox"/> STANDARD SUPPLIER INTOXIMETERS | LOT# AG826202 |
| | EXP. DATE 09/19/2020 |
| <input type="checkbox"/> SIMULATOR TEMP (34°C ±0.2°C) | SIMULATOR S/N |
| | SIMULATOR EXP DATE |

CALIBRATION CHECK - (ONLY ONE STANDARD IS TO BE USED PER MAINTENANCE REPORT)

Run three tests using a standard solution. All three tests must be within +5% of the standard value and must have a spread of .005 or less. Mark the box corresponding to the standard solution being used. (PRINTOUT ATTACHED)

- 0.10% STANDARD - MUST READ BETWEEN 0.095% AND 0.105% INCLUSIVE
- 0.08% STANDARD - MUST READ BETWEEN 0.076% AND 0.084% INCLUSIVE
- 0.04% STANDARD - MUST READ BETWEEN 0.038% AND 0.042% INCLUSIVE

| | | |
|-------------------------------|-------------------------------|-------------------------------|
| TEST 1 \approx 0.100 g/210L | TEST 2 \approx 0.100 g/210L | TEST 3 \approx 0.100 g/210L |
|-------------------------------|-------------------------------|-------------------------------|

INDICATE THE NUMBER OF BREATH TESTS IN THE FOLLOWING RANGES SINCE THE LAST MAINTENANCE REPORT:

| | | | | | | | | | | | |
|----------|---|-------|---|---------|---|---------|---|---------|---|----------|---|
| REFUSALS | 0 | 0-.04 | 0 | .05-.09 | 0 | .10-.14 | 0 | .15-.19 | 0 | OVER .19 | 0 |
|----------|---|-------|---|---------|---|---------|---|---------|---|----------|---|

LIST ANY NEW PARTS AND DESCRIBE ANY ALTERATION OR MODIFICATION THAT WAS MADE TO RESTORE THE INSTRUMENT TO OPERATE SATISFACTORILY AND WITHIN ESTABLISHED LIMITS (USE OTHER SIDE IF NECESSARY).

INSPECTING OFFICER

| | |
|---------------------------------|--------------------------------------|
| SIGNATURE | PRINT FULL NAME bobby calvert |
| TYPE II PERMIT NUMBER 200078 | EXPIRATION DATE 02/26/2022 |
| | TELEPHONE NUMBER (660) 263-0346 |

RETURN COMPLETED REPORT TO THE:
Breath Alcohol Program, Missouri Department of Health and Senior Services,
Southeast District Office, 2875 James Blvd, Poplar Bluff, MO 63901



Airgas USA LLC (LAB)

3500 Bernard Street

St. Louis, Mo. 63103

Ph: (314) 533-3100

Fax: (314) 533-7328

Certificate of Analysis

Customer Name

Exclusive Supplier

Intoximeters, Inc.
2081 Craig Road
St. Louis, Mo 63146

Test Date: 19-Sep-2018

Lot # AG826202 Model 108cadd

| | | | |
|--|--------------------------------|--|---|
| <u>Exp. Date</u> 19-Sep-2020 | <u>Cyl. Type</u> 108 | <u>Component</u> Ethanol Nitrogen | <u>Certified Concentration</u> 0.100 ± 2% BrAC (272 ppm) Balance |
|--|--------------------------------|--|---|

Certification Traceable to N.I.S.T. RGM Ethanol Standards:

| <u>Serial No.</u> | <u>Concentration</u> | <u>Serial No.</u> | <u>Concentration</u> |
|--------------------------|-----------------------------|--------------------------|-----------------------------|
| EB0010581 | 392.1 ppm | EB0010603 | 393.0 ppm |
| EB0010570 | 259.8 ppm | EB0010559 | 258.2 ppm |
| EB0010285 | 208.0 ppm | EB0010595 | 208.3 ppm |
| EB0010561 | 103.6 ppm | EB0010562 | 104.2 ppm |
| EB0010681 | 52.12 ppm | EB0010579 | 52.81 ppm |

Analytical Method: NDIR

Digitally signed by Quality Control
Date: 2018.09.19 15:52:10 -05:00
Reason: Dry gas standard certification of analysis
Location: Airgas USA LLC (Lab)

Approved for Release:

Rod Marsala

ISO 17025:2005 A2LA accredited. Certificate Number 3082.06



STATE OF MISSOURI
 DEPARTMENT OF HEALTH AND SENIOR SERVICES
 BREATH ALCOHOL PROGRAM



PERMIT
TYPE II
BOBBY G CALVERT

Is hereby authorized to instruct and supervise operators, train instructors, inspect, calibrate, perform field service and repairs, and operate the following breath analyzer(s):

ALCO-SENSOR IV WITH PRINTER, INTOX EC/IR II

for the determination of the alcoholic content of blood from a sample of expired air. Permit issued under the provisions of sections 577.020 through 577.041, RSMo and 306.111 through 306.119 RSMo.

DATE 3/5/2018

NUMBER 280112

EXPIRES 3/5/2020

DIRECTOR OF STATE PUBLIC HEALTH LABORATORY

DIRECTOR OF DEPARTMENT OF HEALTH AND SENIOR SERVICES

MO 680 0771 (6-16)

LAB 4 (16-16)

STATE OF MISSOURI
 DEPARTMENT OF HEALTH AND SENIOR SERVICES
 BREATH ALCOHOL PROGRAM

INSTRUMENT OPERATOR CARD

The named cardholder is authorized to operate an evidential breath alcohol instrument for the determination of the alcoholic content in breath form of expired air in Missouri.

Operator: CALVERT, BOBBY
 Permit No: 280112
 Date Issued: 3/5/2018 Date Expires: 3/5/2020