



MISSOURI DEPARTMENT OF HEALTH AND SENIOR SERVICES
 STATE PUBLIC HEALTH LABORATORY
 BREATH ALCOHOL PROGRAM
INTOX DMT MAINTENANCE REPORT

RECEIVED
 By Tracy Crews at 10:32 am, Feb 10, 2020

REPORT #1

Complete this report at the time of the regular monthly preventive maintenance check (not to exceed 35 days).
 Complete this report whenever the instrument is serviced or repaired and whenever it is placed into service.
 Retain the original and send a copy within 15 days to the Breath Alcohol Program, DHSS.

INTOX DMT SN 500175	NAME OF AGENCY Missouri State Highway Patrol	DATE OF INSPECTION 02/01/2020
LOCATION OF INSTRUMENT (STREET AND CITY) 518 N. LINCOLN ST., KAHOKA CLARK COUNTY S.O.		TIME OF INSPECTION 23:14:51

CHECKLIST: Place a mark in the box by each item if found to be satisfactory or is operating within established limits. (Write in observed values where determined). Unmarked items must be corrected before using instrument.

DIAGNOSTIC RECORD

DATE AND TIME <u>02/01/2020 23:14:53</u>	<input checked="" type="checkbox"/> DETECTOR
<input checked="" type="checkbox"/> PROGRAM	<input checked="" type="checkbox"/> FILTER 1
<input checked="" type="checkbox"/> SAMPLE CHAMBER <u>48.8°C</u>	<input checked="" type="checkbox"/> FILTER 2
<input checked="" type="checkbox"/> BREATH TUBE <u>41.9°C</u>	<input checked="" type="checkbox"/> FILTER 3
<input checked="" type="checkbox"/> PUMP	<input checked="" type="checkbox"/> INTERNAL STANDARD

BREATH ANALYZER ACCURACY STANDARDS

SIMULATOR STANDARD COMPRESSED ETHANOL-GAS MIXTURE

STANDARD SUPPLIER INTOXIMETERS LOT # AG905201 EXP. DATE 02/21/2021

SIMULATOR TEMP (34°C ± 0.2°C) _____ SIMULATOR SN _____ SIMULATOR EXP DATE _____

CALIBRATION CHECK - (ONLY ONE STANDARD IS TO BE USED PER MAINTENANCE REPORT)
 Run three tests using a standard. All three tests must be within ±5% of the standard value and must have a spread of .005 or less. Mark the box corresponding to the standard being used.

0.10% STANDARD - MUST READ BETWEEN 0.095% AND 0.105% INCLUSIVE

0.08% STANDARD - MUST READ BETWEEN 0.076% AND 0.084% INCLUSIVE

0.04% STANDARD - MUST READ BETWEEN 0.038% AND 0.042% INCLUSIVE

TEST 1: 0.098 TEST 2: 0.098 TEST 3: 0.098

PERFORM R.F.I. TEST

INDICATE THE NUMBER OF BREATH TESTS IN THE FOLLOWING RANGES SINCE THE LAST MAINTENANCE REPORT:

REFUSALS: 0	0-04: 45	05-09: 0	10-14: 1	15-19: 0	OVER 19: 0
-------------	----------	----------	----------	----------	------------

LIST ANY NEW PARTS AND DESCRIBE ANY ALTERATION OR MODIFICATION THAT WAS MADE TO RESTORE THE INSTRUMENT TO OPERATE SATISFACTORILY AND WITHIN ESTABLISHED LIMITS (USE OTHER SIDE IF NECESSARY)

INSPECTING OFFICER

SIGNATURE 	PRINT FULL NAME BRETT D TAPPENDORF
---------------	---------------------------------------

TYPE II PERMIT NUMBER 280141	EXPIRATION DATE 03/12/2020	TELEPHONE NUMBER 660-385-2132
---------------------------------	-------------------------------	----------------------------------

RETURN COMPLETED REPORT TO THE Breath Alcohol Program, MO Department of Health and Senior Services
 Southeast District Office
 2875 James Blvd, Poplar Bluff, MO 63901



Airgas USA LLC (LAB)
 3500 Bernard Street
 St. Louis, Mo. 63103
 Ph: (314) 533-3100
 Fax: (314) 533-7328

Certificate of Analysis

Customer Name
Exclusive Supplier
 Intoximeters, Inc.
 2081 Craig Road
 St. Louis, Mo 63146

Test Date: 21-Feb-2019

Lot # AG905201 Model 108cadd

Exp. Date 21-Feb-2021	Cyl. Type 108	Component Ethanol Nitrogen	Certified Concentration 0.100 ± 2% BrAC (260 ppm) Balance
---------------------------------	-------------------------	---	--

Certification Traceable to N.I.S.T. RGM and to CRM Ethanol Standards:

<table border="0"> <tr><td><u>RGM Serial No.</u></td><td><u>Concentration</u></td></tr> <tr><td>EB0010581</td><td>392.1 ppm</td></tr> <tr><td>EB0010570</td><td>259.8 ppm</td></tr> <tr><td>EB0010285</td><td>208.0 ppm</td></tr> <tr><td>EB0010561</td><td>103.6 ppm</td></tr> <tr><td>EB0010681</td><td>52.12 ppm</td></tr> </table>	<u>RGM Serial No.</u>	<u>Concentration</u>	EB0010581	392.1 ppm	EB0010570	259.8 ppm	EB0010285	208.0 ppm	EB0010561	103.6 ppm	EB0010681	52.12 ppm	<table border="0"> <tr><td><u>RGM Serial No.</u></td><td><u>Concentration</u></td></tr> <tr><td>EB0010603</td><td>393.0 ppm</td></tr> <tr><td>EB0010559</td><td>258.2 ppm</td></tr> <tr><td>EB0010595</td><td>208.3 ppm</td></tr> <tr><td>EB0010562</td><td>104.2 ppm</td></tr> <tr><td>EB0010579</td><td>52.81 ppm</td></tr> </table>	<u>RGM Serial No.</u>	<u>Concentration</u>	EB0010603	393.0 ppm	EB0010559	258.2 ppm	EB0010595	208.3 ppm	EB0010562	104.2 ppm	EB0010579	52.81 ppm
<u>RGM Serial No.</u>	<u>Concentration</u>																								
EB0010581	392.1 ppm																								
EB0010570	259.8 ppm																								
EB0010285	208.0 ppm																								
EB0010561	103.6 ppm																								
EB0010681	52.12 ppm																								
<u>RGM Serial No.</u>	<u>Concentration</u>																								
EB0010603	393.0 ppm																								
EB0010559	258.2 ppm																								
EB0010595	208.3 ppm																								
EB0010562	104.2 ppm																								
EB0010579	52.81 ppm																								
<table border="0"> <tr><td><u>CRM Serial No.</u></td><td><u>Concentration</u></td></tr> <tr><td>CC434668</td><td>800.0 ppm</td></tr> <tr><td>CC234503</td><td>253.0 ppm</td></tr> </table>	<u>CRM Serial No.</u>	<u>Concentration</u>	CC434668	800.0 ppm	CC234503	253.0 ppm	<table border="0"> <tr><td><u>CRM Serial No.</u></td><td><u>Concentration</u></td></tr> <tr><td>0056649</td><td>390.1 ppm</td></tr> <tr><td>0056662</td><td>150.2 ppm</td></tr> </table>	<u>CRM Serial No.</u>	<u>Concentration</u>	0056649	390.1 ppm	0056662	150.2 ppm												
<u>CRM Serial No.</u>	<u>Concentration</u>																								
CC434668	800.0 ppm																								
CC234503	253.0 ppm																								
<u>CRM Serial No.</u>	<u>Concentration</u>																								
0056649	390.1 ppm																								
0056662	150.2 ppm																								

Analytical Method: NDIR

Digitally signed by Quality Control
 Date: 2019.02.21 15:31:31 -06:00
 Reason: Dry gas standard certification of analysis
 Location: Airgas USA LLC (Lab)

Approved for Release:

 Rod Marsala

ISO 17025:2005 A2LA accredited. Certificate Number 3082.06
ISO 17034:2016 A2LA accredited. Certificate Number 3082.07



STATE OF MISSOURI
 DEPARTMENT OF HEALTH AND SENIOR SERVICES
 BREATH ALCOHOL PROGRAM



PERMIT
TYPE II
BRETT D TAPPENDORF

is hereby authorized to instruct and supervise operators, train instructors, inspect, calibrate, perform field service and repairs, and operate the following breath analyzer(s):

INTOX DMT

for the determination of the alcoholic content of blood from a sample of expired air. Permit issued under the provisions of sections 577.020 through 577.041, RSMo and 306.111 through 306.119 RSMo.

DATE 3/12/2018

DIRECTOR OF STATE PUBLIC HEALTH LABORATORY

NUMBER 280141

EXPIRES 3/12/2020

DIRECTOR OF DEPARTMENT OF HEALTH AND SENIOR SERVICES

MO 580-0771 (6-10)

LAB-4 (R6-10)

STATE OF MISSOURI
 DEPARTMENT OF HEALTH AND SENIOR SERVICES
 BREATH ALCOHOL PROGRAM

INSTRUMENT OPERATOR CARD

The named cardholder is authorized to operate an evidential breath alcohol instrument for the determination of the alcoholic content in breath form of expired air in Missouri.

Operator TAPPENDORF, BRETT
 Permit No 280141
 Date Issued 3/12/2018 Date Expires 3/12/2020