



MISSOURI DEPARTMENT OF HEALTH AND SENIOR SERVICES  
STATE PUBLIC HEALTH LABORATORY  
BREATH ALCOHOL PROGRAM

**INTOX DMT MAINTENANCE REPORT**

**RECEIVED**

By Tracy Crews at 10:12 am, Nov 23, 2020

REPORT #1

Complete this report at the time of the regular monthly preventive maintenance check (not to exceed 35 days).  
Complete this report whenever the instrument is serviced or repaired and whenever it is placed into service.  
Retain the original and send a copy within 15 days to the Breath Alcohol Program, DHSS.

|   |  |   |
|---|--|---|
| INTOX DMT SN<br><b>500202</b>   | NAME OF AGENCY<br><b>Missouri State Highway Patrol</b> | DATE OF INSPECTION<br><b>11/20/2020</b> |
| LOCATION OF INSTRUMENT (STREET AND CITY)<br><b>Reynolds Co. Sheriff's Dept, Centerville, MO</b> |  | TIME OF INSPECTION<br><b>20:22:55</b>   |

CHECKLIST: Place a mark in the box by each item if found to be satisfactory or is operating within established limits. (Write in observed values where determined). Unmarked items must be corrected before using instrument.

|   |  |
|---|--|
| <input checked="" type="checkbox"/> <b>DIAGNOSTIC RECORD</b>            |  |
| DATE AND TIME <u>11/20/2020 20:22:57</u>                                | <input checked="" type="checkbox"/> <b>DETECTOR</b>          |
| <input checked="" type="checkbox"/> <b>PROGRAM</b>                      | <input checked="" type="checkbox"/> <b>FILTER 1</b>          |
| <input checked="" type="checkbox"/> <b>SAMPLE CHAMBER</b> <u>48.7°C</u> | <input checked="" type="checkbox"/> <b>FILTER 2</b>          |
| <input checked="" type="checkbox"/> <b>BREATH TUBE</b> <u>46.7°C</u>    | <input checked="" type="checkbox"/> <b>FILTER 3</b>          |
| <input checked="" type="checkbox"/> <b>PUMP</b>                         | <input checked="" type="checkbox"/> <b>INTERNAL STANDARD</b> |

|  |   |
|--|---|
| <b>BREATH ANALYZER ACCURACY STANDARDS</b>          |   |
| <input type="checkbox"/> <b>SIMULATOR STANDARD</b> | <input checked="" type="checkbox"/> <b>COMPRESSED ETHANOL-GAS MIXTURE</b> |

|  |                              |                                    |
|--|------------------------------|------------------------------------|
| <input checked="" type="checkbox"/> <b>STANDARD SUPPLIER</b> <u>INTOXIMETERS</u> | <b>LOT #</b> <u>AG004403</u> | <b>EXP. DATE</b> <u>02/13/2022</u> |
|--|------------------------------|------------------------------------|

|   |                      |                                 |
|---|----------------------|---------------------------------|
| <input type="checkbox"/> <b>SIMULATOR TEMP</b> (34°C ± 0.2°C) _____ | <b>SIM. SN</b> _____ | <b>SIM. NIST EXP DATE</b> _____ |
|---|----------------------|---------------------------------|

|  |  |
|--|--|
| <input checked="" type="checkbox"/> <b>CALIBRATION CHECK - (ONLY ONE STANDARD IS TO BE USED PER MAINTENANCE REPORT)</b><br>Run three tests using a standard. All three tests must be within ±5% of the standard value and must have a spread of .005 or less. Mark the box corresponding to the standard being used. |  |
| <input checked="" type="checkbox"/> <b>0.10% STANDARD - MUST READ BETWEEN 0.095% AND 0.105% INCLUSIVE</b>  |  |
| <input type="checkbox"/> <b>0.08% STANDARD - MUST READ BETWEEN 0.076% AND 0.084% INCLUSIVE</b>   |  |
| <input type="checkbox"/> <b>0.04% STANDARD - MUST READ BETWEEN 0.038% AND 0.042% INCLUSIVE</b>   |  |

|                             |                             |                             |
|-----------------------------|-----------------------------|-----------------------------|
| <b>TEST 1:</b> <u>0.100</u> | <b>TEST 2:</b> <u>0.100</u> | <b>TEST 3:</b> <u>0.100</u> |
|-----------------------------|-----------------------------|-----------------------------|

|  |
|--|
| <input checked="" type="checkbox"/> <b>PERFORM R.F.I. TEST</b> |
|--|

INDICATE THE NUMBER OF BREATH TESTS IN THE FOLLOWING RANGES SINCE THE LAST MAINTENANCE REPORT:

|                           |                        |                          |                          |                          |                           |
|---------------------------|------------------------|--------------------------|--------------------------|--------------------------|---------------------------|
| <b>REFUSALS:</b> <u>0</u> | <b>0-.04:</b> <u>0</u> | <b>.05-.09:</b> <u>0</u> | <b>.10-.14:</b> <u>0</u> | <b>.15-.19:</b> <u>0</u> | <b>OVER .19:</b> <u>0</u> |
|---------------------------|------------------------|--------------------------|--------------------------|--------------------------|---------------------------|

LIST ANY NEW PARTS AND DESCRIBE ANY ALTERATION OR MODIFICATION THAT WAS MADE TO RESTORE THE INSTRUMENT TO OPERATE SATISFACTORILY AND WITHIN ESTABLISHED LIMITS (USE OTHER SIDE IF NECESSARY)

Adjusted time.

**INSPECTING OFFICER**

|                      |   |
|----------------------|---|
| <b>SIGNATURE</b><br> | <b>PRINT FULL NAME</b><br><b>THOMAS W MEYER</b> |
|----------------------|---|

|   |   |  |
|---|---|--|
| <b>TYPE II PERMIT NUMBER</b><br><b>200244</b> | <b>EXPIRATION DATE</b><br><b>09/08/2022</b> | <b>TELEPHONE NUMBER</b><br><b>417-469-3121</b> |
|---|---|--|

RETURN COMPLETED REPORT TO THE **Breath Alcohol Program, Missouri Department of Health and Senior Services**  
by mail, fax, or email



STATE OF MISSOURI  
DEPARTMENT OF HEALTH AND SENIOR SERVICES  
BREATH ALCOHOL PROGRAM

**PERMIT  
TYPE II**

**THOMAS W. MEYER**

is hereby authorized to instruct and supervise operators, train instructors, inspect, calibrate, perform field sobriety tests and operate the following breath analyzer(s):

**ALCO-SENSOR IV WITH PRINTER, INTOX DMT**

for the determination of the alcoholic content of blood from a sample of expired air. Permit issued under the provisions of laws 577.020 through 577.041, RSMo and 306.111 through 306.119 RSMo.

DATE 9/8/2020

NUMBER 200244

EXPIRES 9/8/2022

MO-60 (7/11) (5-19)

*Thomas W. Meyer*  
DIRECTOR OF STATE POLICE DEPARTMENT

*[Signature]*  
DIRECTOR OF DEPARTMENT OF HEALTH AND SENIOR SERVICES

 STATE OF MISSOURI  
DEPARTMENT OF HEALTH AND SENIOR SERVICES  
BREATH ALCOHOL PROGRAM

**INSTRUMENT OPERATOR CARD**

*The named cardholder is authorized to operate an essential breath alcohol instrument for the determination of the alcoholic content in breath form of expiration in Missouri.*

Operator MEYER, THOMAS  
Permit No 200244  
Date Issued 9/8/2020 Date Expires 9/8/2022





KEYNOIDS

Airgas USA LLC (LAB)  
3500 Bernard Street  
St. Louis, Mo. 63103  
Ph: (314) 533-5100  
Fax: (314) 533-7400

### Certificate of Analysis

**Customer Name**  
Exclusive Supplier  
Intoximeters, Inc  
2081 Craig Road  
St. Louis, Mo 63146

**Test Date:** 12 Feb 2021

**Lot #** AG004403 **Model** 108caccd

| <u>Exp. Date</u> | <u>Cyl. Type</u> | <u>Component</u>    | <u>Certified Concentration</u>      |
|------------------|------------------|---------------------|-------------------------------------|
| 13-Feb-2022      | 108              | Ethanol<br>Nitrogen | 0.100 ± 2% BrAQ 1260 ppm<br>Balance |

Certification Traceable to N.I.S.T. RGM and to CRM Ethanol Standards:

| <u>RGM Serial No.</u> | <u>Concentration</u> | <u>RGM Serial No.</u> | <u>Concentration</u> |
|-----------------------|----------------------|-----------------------|----------------------|
| EB0010581             | 392.1 ppm            | EB0010603             | 393.0 ppm            |
| EB0010570             | 259.8 ppm            | EB0010559             | 258.2 ppm            |
| EB0010285             | 208.0 ppm            | EB0010595             | 208.3 ppm            |
| EB0010561             | 103.6 ppm            | EB0010562             | 104.2 ppm            |
| EB0010681             | 52.12 ppm            | EB0010579             | 52.81 ppm            |

  

| <u>CRM Serial No.</u> | <u>Concentration</u> | <u>CRM Serial No.</u> | <u>Concentration</u> |
|-----------------------|----------------------|-----------------------|----------------------|
| CC434668              | 800.0 ppm            | 0056649               | 390.1 ppm            |
| CC234503              | 253.0 ppm            | 0056662               | 150.2 ppm            |

**Analytical Method:** NDIR

Digitally signed by Quality Control  
Date: 2020.02.18 10:32:01 -06:00  
Reason: Dry gas standard certification of analysis  
Location: Airgas USA LLC (Lab)

Approved for Release:

Rod Marsala

ISO 17025:2005 A2LA accredited. Certificate Number 3082.06  
ISO 17034:2016 A2LA accredited. Certificate Number 3082.07