



Airgas USA LLC (LAB)
 3500 Bernard Street
 St. Louis, Mo. 63103
 Ph: (314) 533-3100
 Fax: (314) 533-7328

Certificate of Analysis

Customer Name
 Exclusive Supplier
 Intoximeters, Inc.
 2081 Craig Road
 St. Louis, Mo 63146

Test Date: 23-Apr-2020

Lot # AG011305 Model 108cadd

Exp. Date 22-Apr-2022	Cyl. Type 108	Component Ethanol Nitrogen	Certified Concentration 0.100 ± 2% BrAC (260 ppm) Balance
---------------------------------	-------------------------	---	--

Certification Traceable to N.I.S.T. RGM and to CRM Ethanol Standards:

<u>RGM Serial No.</u>	<u>Concentration</u>	<u>RGM Serial No.</u>	<u>Concentration</u>
EB0010581	392.1 ppm	EB0010603	393.0 ppm
EB0010570	259.8 ppm	EB0010559	258.2 ppm
EB0010285	208.0 ppm	EB0010595	208.3 ppm
EB0010561	103.6 ppm	EB0010562	104.2 ppm
EB0010681	52.12 ppm	EB0010579	52.81 ppm

<u>CRM Serial No.</u>	<u>Concentration</u>	<u>CRM Serial No.</u>	<u>Concentration</u>
CC434668	800.0 ppm	0056649	390.1 ppm
CC234503	253.0 ppm	0056662	150.2 ppm

Analytical Method: NDIR

Digitally signed by Quality Control
 Date: 2020.04.24 14:59:04 -05:00
 Reason: Dry gas standard certification of analysis
 Location: Airgas USA LLC (Lab)

Approved for Release: _____

Rod Marsala

 Rod Marsala

ISO 17025:2005 A2LA accredited. Certificate Number 3082.06
ISO 17034:2016 A2LA accredited. Certificate Number 3082.07



STATE OF MISSOURI
DEPARTMENT OF HEALTH AND SENIOR SERVICES
BREATH ALCOHOL PROGRAM

2

PERMIT
TYPE II
STEPHEN L. CONNER

is hereby authorized to instruct and supervise operators, train instructors, inspect, calibrate, perform field service and repairs and operate the following breath analyzer(s):

INTOX DMT

for the determination of the alcoholic content of blood from a sample of expired air. Permit issued under the provisions of sections 577.020 through 577.041, RSMo and 306.111 through 306.119 RSMo.

DATE 9/5/2018

NUMBER 280267

EXPIRES 9/5/2020

DIRECTOR OF STATE PUBLIC HEALTH LABORATORY

DIRECTOR OF DEPARTMENT OF HEALTH AND SENIOR SERVICES

MO 580-0771 (6-10)

LAB-4 (R6-10)



STATE OF MISSOURI
DEPARTMENT OF HEALTH AND SENIOR SERVICES
BREATH ALCOHOL PROGRAM

INSTRUMENT OPERATOR CARD

The named cardholder is authorized to operate an evidential breath alcohol instrument for the determination of the alcoholic content in breath form of expired air in Missouri.

Operator CONNER, STEPHEN
Permit No 280267
Date Issued 9/5/2018 **Date Expires** 9/5/2020





Chief of Police
Justin Arnold
Ozark Police Department
201 E. Brick Street
Ozark, Missouri 65721
Telephone (417) 581-6600
FAX (417) 581-8804



FACSIMILE COVER SHEET

THE INFORMATION CONTAINED IN THIS FAX IS PRIVILEGED AND CONFIDENTIAL; INTENDED FOR THE USE OF THE INDIVIDUAL(S) NAMED WITHIN. IF THE READER OF THIS MESSAGE IS NOT THE PERSON WHOM IT IS SENT (OR THE EMPLOYEE OR AGENT RESPONSIBLE TO DELIVER IT TO THAT PERSON) DO NOT CIRCULATE, DISCLOSE, COPY, REPLICATE OR DISTRIBUTE THIS MESSAGE. IF YOU HAVE RECEIVED THIS COMMUNICATION IN ERROR, PLEASE NOTIFY US BY PHONE AND RETURN THE ORIGINAL MESSAGE TO US AT THE ADDRESS ABOVE AT OUR EXPENSE.

DATE: 06/4/2020

TO: DHSS Breath Alcohol

ATTN:

FROM: Det. Stephen Conner

REFERENCE: DMT Maintenance

TOTAL PAGE(s) INCLUDING THIS SHEET: 5

MESSAGE: Let me know if you have any questions.
s.conner@ozarkpd.org

IF THIS TRANSMISSION IS NOT RECEIVED IN FULL PLEASE CONTACT ME AT 1-(417)-581-6600. THANK YOU.