

RECEIVED

By Tracy Crews at 11:48 am, Mar 16, 2020



MISSOURI DEPARTMENT OF HEALTH AND SENIOR SERVICES
STATE PUBLIC HEALTH LABORATORY
CMI INTOXILYZER 8000 MAINTENANCE REPORT

REPORT #2

Complete this report in duplicate at the time of the regular monthly preventive maintenance check, and whenever instrument is repaired. Send one copy to Department of Health and Senior Services, and retain one copy in department file.

INSTRUMENT SERIAL NUMBER 80-002092	LOCATION OF INSTRUMENT KCI POLICE DEPT.	DATE OF INSPECTION 03/15/2020	TIME OF INSPECTION 09:06
---------------------------------------	--	----------------------------------	-----------------------------

CALIBRATION CHECK RESULTS			CALIBRATION CHECK SUMMARY		
Test	g/210L	Time	STANDARD TYPE	STANDARD LOT #	STANDARD EXPIRATION DATE
Air Blank	0.000	09:07	DRY	17919080A1	08/05/2021
Cal Check	0.081	09:08	SIM TEMPERATURE	SIM SERIAL NUMBER	SIM CERTIFICATE EXPIRATION
Air Blank	0.000	09:08	N/A	N/A	N/A
Cal Check	0.080	09:08	STANDARD VALUE	STANDARD SUPPLIER	
Air Blank	0.000	09:09	0.080	CMI	
Cal Check	0.080	09:09	CALIBRATION CHECK RESULT 1		
Air Blank	0.000	09:10	0.081		
Cal Check	0.080	09:09	CALIBRATION CHECK RESULT 2		
Air Blank	0.000	09:10	0.080		
Cal Check	0.080	09:09	CALIBRATION CHECK RESULT 3		
Air Blank	0.000	09:10	0.080		
Pass			MAXIMUM DEVIATION (MUST BE WITHIN 5%)	SPREAD (MUST BE .005 OR LESS)	
			1.2%	0.001	

DIAGNOSTIC TEST RESULTS			RFI TEST RESULTS		
Test	Pass	Time	Test	g/210L	Time
Voltage/Current Test	Pass		Air Blank	RFI*	09:10
RAM Test	Pass		Air Blank	0.000	09:10
EEPROM Checksum Test	Pass		*RFI Detect		
Real Time Clock Test	Pass		Pass		
DSP Test	Pass		Pass		
Analytical Stability Test	Pass		Pass		
Modem Test	Pass		Pass		
Temperature Regulation Test	Pass		Pass		
Pass			Pass		

NUMBER OF REFUSALS AND SUBJECT BREATH TESTS IN EACH RANGE SINCE LAST MAINTENANCE REPORT

REFUSALS	.00-.04	.05-.09	.10-.14	.15-.19	OVER .19
0	0	0	0	0	0

List any new parts and describe any alteration or modification that was made to restore the instrument to operate satisfactorily and within established limits (use other side if necessary).

INSPECTING OFFICER

SIGNATURE 	PRINT NAME JOSHUA DEAL	
TYPE II PERMIT NUMBER 280220	EXPIRATION DATE 06/28/2020	TELEPHONE NUMBER 816-243-4000



7 Eastgate Dr. • P.O. Box 790 • Jacksonville, IL 62651-0790
217-245-2183 • Fax: 217-243-7634 • www.ilmoproducts.com

Certificate of Analysis

Certificate ID: 12203
Part #: BAC105L080T
Cylinder Size: 105L
Lot Number: 17919080A1
Expiration: 8/5/2021

0.080 BAC (For the calibration of instruments used to determine breath alcohol concentration)

Contents: 105 Liters @ 1000 psig 70°F (21°C)

Component:	Reported Concentration:	Analytical Accuracy (U, k=2):	Analytical Method:
Ethanol	208 ppm	+/- 0.002 BAC (6/21BL)	NDIR
Nitrogen	Balance	[5.2 ppm]	

Distributed by:
CMI Inc.
316 East Ninth Street
Owensboro, KY 42303
Phone 866-835-0690
www.alcoholtest.com

*Traceable to:
Certified Reference Material - 262.4 µmol/mol
Ethanol in Nitrogen - Serial No. GN0015026 Lot No. 050319E11

Store in dry area, away from sources of heat, ignition and direct sunlight. Do not allow storage area to exceed 52 °C (125 °F).

[Signature]
Specialty Gas Lab Tech

07-15-19
Issuance Date



The calibration results within this certificate were obtained using equipment and standards capable of producing analytical results traceable to NIST, and apply only to the items or material on this certificate. ILMO Products Company makes no warranty or representation as to the suitability of the use of any information provided for any particular purpose. The information use is at the sole discretion and risk of the user. Liability shall be limited to established replacement cost of this material or service.



STATE OF MISSOURI
 DEPARTMENT OF HEALTH AND SENIOR SERVICES
 BREATH ALCOHOL PROGRAM

2

PERMIT
TYPE II
JOSHUA J DEAL

is hereby authorized to instruct and supervise operators, train instructors, inspect, calibrate, perform field service and repairs, and operate the following breath analyzer(s):


INTOXILYZER 8000


for the determination of the alcoholic content of blood from a sample of expired air. Permit issued under the provisions of sections 577.020 through 577.041, RSMo and 306.111 through 306.119 RSMo.

DATE 6/28/2018

NUMBER 280220

EXPIRES 6/28/2020


 DIRECTOR OF STATE PUBLIC HEALTH LABORATORY


 DIRECTOR OF DEPARTMENT OF HEALTH AND SENIOR SERVICES

MO 580-0771 (6-10)

LAB-4 (R6-10)


 STATE OF MISSOURI
 DEPARTMENT OF HEALTH AND SENIOR SERVICES
 BREATH ALCOHOL PROGRAM

INSTRUMENT OPERATOR CARD

The named cardholder is authorized to operate an evidential breath alcohol instrument for the determination of the alcoholic content in breath form of expired air in Missouri.



Operator DEAL, JOSHUA
 Permit No 280220
 Date Issued 6/28/2018 Date Expires 6/28/2020