



MISSOURI DEPARTMENT OF HEALTH AND SENIOR SERVICES
STATE PUBLIC HEALTH LABORATORY
CMI INTOXILYZER 8000 MAINTENANCE REPORT

RECEIVED

By Tracy Crews at 11:29 am, Feb 04, 2020

REPORT #2

Complete this report in duplicate at the time of the regular monthly preventive maintenance check, and whenever instrument is repaired. Send one copy to Department of Health and Senior Services, and retain one copy in department file.

INSTRUMENT SERIAL NUMBER 80-005828	LOCATION OF INSTRUMENT WEBB CITY P. D.	DATE OF INSPECTION 02/01/2020	TIME OF INSPECTION 09:02
---------------------------------------	---	----------------------------------	-----------------------------

CALIBRATION CHECK RESULTS			CALIBRATION CHECK SUMMARY		
Test	g/210L	Time	STANDARD TYPE	STANDARD LOT #	STANDARD EXPIRATION DATE
Air Blank	0.000	09:03	DRY	AG907710	03/18/2021
Cal Check	0.099	09:03	SIM TEMPERATURE	SIM SERIAL NUMBER	SIM CERTIFICATE EXPIRATION
Air Blank	0.000	09:04	N/A	N/A	N/A
Cal Check	0.100	09:04	STANDARD VALUE	STANDARD SUPPLIER	
Air Blank	0.000	09:05	0.100	INTOXIMETERS	
Cal Check	0.100	09:05	CALIBRATION CHECK RESULT 1		
Air Blank	0.000	09:06	0.099		
Pass			CALIBRATION CHECK RESULT 2		
			0.100		
			CALIBRATION CHECK RESULT 3		
			0.100		
			MAXIMUM DEVIATION (MUST BE WITHIN 5%)	SPREAD (MUST BE .005 OR LESS)	
			1.0%	0.001	

DIAGNOSTIC TEST RESULTS			RFI TEST RESULTS		
			Test	g/210L	Time
Voltage/Current Test	Pass		Air Blank	0.000	09:06
RAM Test	Pass		Subject Test	RFI*	09:07
EEPROM Checksum Test	Pass		Air Blank	0.000	09:07
Real Time Clock Test	Pass		*RFI Detect		
DSP Test	Pass		Pass		
Analytical Stability Test	Pass				
Modem Test	Pass				
Temperature Regulation Test	Pass				

NUMBER OF REFUSALS AND SUBJECT BREATH TESTS IN EACH RANGE SINCE LAST MAINTENANCE REPORT							
REFUSALS	.00-.04	.05-.09	.10-.14	.15-.19	OVER .19		
0	0	0	0	1	0		

List any new parts and describe any alteration or modification that was made to restore the instrument to operate satisfactorily and within established limits (use other side if necessary).

INSPECTING OFFICER		
SIGNATURE 	PRINT NAME CHRISTOPHER SHONK	
TYPE II PERMIT NUMBER 290143	EXPIRATION DATE 06/27/2021	TELEPHONE NUMBER 4176731911



Airgas USA LLC (LAB)

3500 Bernard Street

St. Louis, Mo. 63103

Ph: (314) 533-3100

Fax: (314) 533-7328

Certificate of Analysis

Customer Name

Exclusive Supplier
Intoximeters, Inc.
2081 Craig Road
St. Louis, Mo 63146

Test Date: 20-Mar-2019

Lot # AG907710 Model 108cadd

<u>Exp. Date</u>	<u>Cyl. Type</u>	<u>Component</u>	<u>Certified Concentration</u>
18-Mar-2021	108	Ethanol Nitrogen	0.100 ± 2% BrAC (260 ppm) Balance

Certification Traceable to N.I.S.T. RGM and to CRM Ethanol Standards:

<u>RGM Serial No.</u>	<u>Concentration</u>	<u>RGM Serial No.</u>	<u>Concentration</u>
EB0010581	392.1 ppm	EB0010603	393.0 ppm
EB0010570	259.8 ppm	EB0010559	258.2 ppm
EB0010285	208.0 ppm	EB0010595	208.3 ppm
EB0010561	103.6 ppm	EB0010562	104.2 ppm
EB0010681	52.12 ppm	EB0010579	52.81 ppm

<u>CRM Serial No.</u>	<u>Concentration</u>	<u>CRM Serial No.</u>	<u>Concentration</u>
CC434668	800.0 ppm	0056649	390.1 ppm
CC234503	253.0 ppm	0056662	150.2 ppm

Analytical Method: NDIR

Digitally signed by Quality Control
Date: 2019.03.20 11:13:55 -05:00
Reason: Dry gas standard certification of analysis
Location: Airgas USA LLC (Lab)

Approved for Release: 
Rod Marsala

ISO 17025:2005 A2LA accredited. Certificate Number 3082.06
ISO 17034:2016 A2LA accredited. Certificate Number 3082.07



STATE OF MISSOURI
 DEPARTMENT OF HEALTH AND SENIOR SERVICES
 BREATH ALCOHOL PROGRAM

2

PERMIT
TYPE II
CHRISTOPHER SHONK

is hereby authorized to instruct and supervise operators, train instructors, inspect, calibrate, perform field service and repairs, and operate the following breath analyzer(s):

INTOX DMT, INTOXILYZER 8000, ALCO-SENSOR IV W/PRINTER

for the determination of the alcoholic content of blood from a sample of expired air. Permit issued under the provisions of sections 577.020 through 577.041, RSMo and 306.111 through 306.119 RSMo.

DATE 6/27/2019

NUMBER 290143

EXPIRES 6/27/2021

DIRECTOR OF STATE PUBLIC HEALTH LABORATORY

DIRECTOR OF DEPARTMENT OF HEALTH AND SENIOR SERVICES

MO 580 0771 (6-10)

LAB 4 (R6-10)

STATE OF MISSOURI
 DEPARTMENT OF HEALTH AND SENIOR SERVICES
 BREATH ALCOHOL PROGRAM

INSTRUMENT OPERATOR CARD

The named cardholder is authorized to operate an evidential breath alcohol instrument for the determination of the alcoholic content in breath form of expired air in Missouri.

Operator SHONK, CHRISTOPHER
 Permit No 290143
 Date Issued 6/27/2019 Date Expires 6/27/2021