



MISSOURI DEPARTMENT OF HEALTH AND SENIOR SERVICES  
STATE PUBLIC HEALTH LABORATORY  
**CMI INTOXILYZER 8000 MAINTENANCE REPORT**

**RECEIVED**

By Tracy Crews at 3:45 pm, Dec 02, 2020

REPORT #2

Complete this report in duplicate at the time of the regular monthly preventive maintenance check, and whenever instrument is repaired. Send one copy to Department of Health and Senior Services, and retain one copy in department file.

INSTRUMENT SERIAL NUMBER <b>80-005828</b>	LOCATION OF INSTRUMENT <b>WEBB CITY P. D.</b>	DATE OF INSPECTION <b>12/02/2020</b>	TIME OF INSPECTION <b>14:12</b>
--	--	---	------------------------------------

CALIBRATION CHECK RESULTS			CALIBRATION CHECK SUMMARY		
Test	g/210L	Time	STANDARD TYPE	STANDARD LOT #	STANDARD EXPIRATION DATE
Air Blank	0.000	14:14	DRY	AG907710	03/18/2021
Cal Check	0.099	14:14	SIM TEMPERATURE	SIM SERIAL NUMBER	SIM CERTIFICATE EXPIRATION
Air Blank	0.000	14:14	N/A	N/A	N/A
Cal Check	0.099	14:15	STANDARD VALUE	STANDARD SUPPLIER	
Air Blank	0.000	14:15	0.100	INTOXIMETERS	
Cal Check	0.099	14:16	CALIBRATION CHECK RESULT 1		
Air Blank	0.000	14:16	0.099		
Cal Check	0.099	14:16	CALIBRATION CHECK RESULT 2		
Air Blank	0.000	14:16	0.099		
<b>Pass</b>			CALIBRATION CHECK RESULT 3		
			0.099		
			MAXIMUM DEVIATION (MUST BE WITHIN 5%)	SPREAD (MUST BE .005 OR LESS)	
			1.0%	0.000	

DIAGNOSTIC TEST RESULTS			RFI TEST RESULTS		
			Test	g/210L	Time
Voltage/Current Test	Pass		Air Blank	0.000	14:17
RAM Test	Pass		Subject Test	RFI*	14:17
EEPROM Checksum Test	Pass		Air Blank	0.000	14:17
Real Time Clock Test	Pass		*RFI Detect		
DSP Test	Pass		<b>Pass</b>		
Analytical Stability Test	Pass				
Modem Test	Pass				
Temperature Regulation Test	Pass				

NUMBER OF REFUSALS AND SUBJECT BREATH TESTS IN EACH RANGE SINCE LAST MAINTENANCE REPORT							
REFUSALS	00-04	05-09	10-14	15-19	OVER 19		
0	0	0	0	2	0		

List any new parts and describe any alteration or modification that was made to restore the instrument to operate satisfactorily and within established limits (use other side if necessary).

INSPECTING OFFICER		
SIGNATURE 	PRINT NAME <b>CHRISTOPHER SHONK</b>	
TYPE II PERMIT NUMBER <b>290143</b>	EXPIRATION DATE <b>06/27/2021</b>	TELEPHONE NUMBER <b>4176731911</b>



Airgas USA LLC (LAB)  
 3500 Bernard Street  
 St. Louis, Mo. 63103  
 Ph: (314) 533-3100  
 Fax: (314) 533-7328

## Certificate of Analysis

Customer Name  
 Exclusive Supplier  
 Intoximeters, Inc.  
 2081 Craig Road  
 St. Louis, Mo 63146

Test Date: 20-Mar-2019

Lot # AG907710 Model 108cadd

<u>Exp. Date</u> 18-Mar-2021	<u>Cyl. Type</u> 108	<u>Component</u> Ethanol Nitrogen	<u>Certified Concentration</u> 0.100 ± 2% BrAC (260 ppm) Balance
---------------------------------	-------------------------	---	--

Certification Traceable to N.I.S.T. RGM and to CRM Ethanol Standards:

<u>RGM Serial No.</u>	<u>Concentration</u>	<u>RGM Serial No.</u>	<u>Concentration</u>
EB0010581	392.1 ppm	EB0010603	393.0 ppm
EB0010570	259.8 ppm	EB0010559	258.2 ppm
EB0010285	208.0 ppm	EB0010595	208.3 ppm
EB0010561	103.6 ppm	EB0010562	104.2 ppm
EB0010681	52.12 ppm	EB0010579	52.81 ppm

  

<u>CRM Serial No.</u>	<u>Concentration</u>	<u>CRM Serial No.</u>	<u>Concentration</u>
CC434668	800.0 ppm	0056649	390.1 ppm
CC234503	253.0 ppm	0056662	150.2 ppm

Analytical Method: NDIR

Digitally signed by Quality Control  
 Date: 2019.03.20 11:13:55 -05:00  
 Reason: Dry gas standard certification of analysis  
 Location: Airgas USA LLC (Lab)

Approved for Release:

\_\_\_\_\_  
 Rod Marsala

**ISO 17025:2005 A2LA accredited. Certificate Number 3082.06**  
**ISO 17034:2016 A2LA accredited. Certificate Number 3082.07**



STATE OF MISSOURI  
 DEPARTMENT OF HEALTH AND SENIOR SERVICES  
 BREATH ALCOHOL PROGRAM

2

**PERMIT**  
**TYPE II**  
**CHRISTOPHER SHONK**

is hereby authorized to instruct and supervise operators, train instructors, inspect, calibrate, perform field service and repairs, and operate the following breath analyzer(s):

**INTOX DMT, INTOXILYZER 8000, ALCO-SENSOR IV W/PRINTER**

for the determination of the alcoholic content of blood from a sample of expired air. Permit issued under the provisions of sections 577.020 through 577.041, RSMo and 306.111 through 306.119 RSMo.

DATE 6/27/2019

NUMBER 290143

EXPIRES 6/27/2021

DIRECTOR OF STATE PUBLIC HEALTH LABORATORY

DIRECTOR OF DEPARTMENT OF HEALTH AND SENIOR SERVICES

MOI 565.0771 (5-10)

LAB 4 (R16 10)

**STATE OF MISSOURI**  
 DEPARTMENT OF HEALTH AND SENIOR SERVICES  
 BREATH ALCOHOL PROGRAM

**INSTRUMENT OPERATOR CARD**

The named cardholder is authorized to operate an evidential breath alcohol instrument for the determination of the alcoholic content in breath from expired air in Missouri.

Operator SHONK, CHRISTOPHER  
 Permit No 290143  
 Date Issued 6/27/2019 Date Expires 6/27/2021