



MISSOURI DEPARTMENT OF HEALTH AND SENIOR SERVICES
 STATE PUBLIC HEALTH LABORATORY
CMI INTOXILYZER 8000 MAINTENANCE REPORT

REPORT #2

Complete this report in duplicate at the time of the regular monthly preventive maintenance check, and whenever instrument is repaired. Send one copy to Department of Health and Senior Services, and retain one copy in department file.

INSTRUMENT SERIAL NUMBER 80-005835	LOCATION OF INSTRUMENT IPD BAT VAN STATION1	DATE OF INSPECTION 08/11/2020	TIME OF INSPECTION 16:37
---------------------------------------	--	----------------------------------	-----------------------------

CALIBRATION CHECK RESULTS			CALIBRATION CHECK SUMMARY		
Test	g/210L	Time	STANDARD TYPE	STANDARD LOT #	STANDARD EXPIRATION DATE
Air Blank	0.000	16:40	DRY	AG928002	10/07/2021
Cal Check	0.078	16:41	SIM TEMPERATURE	SIM SERIAL NUMBER	SIM CERTIFICATE EXPIRATION
Air Blank	0.000	16:41	N/A	N/A	N/A
Cal Check	0.080	16:42	STANDARD VALUE	STANDARD SUPPLIER	
Air Blank	0.000	16:42	0.080	INTOXIMETERS, INC	
Cal Check	0.079	16:42	CALIBRATION CHECK RESULT 1		
Air Blank	0.000	16:43	0.078		
Pass			CALIBRATION CHECK RESULT 2		
			0.080		
			CALIBRATION CHECK RESULT 3		
			0.079		
			MAXIMUM DEVIATION (MUST BE WITHIN 5%)	SPREAD (MUST BE .005 OR LESS)	
			2.5%	0.002	

DIAGNOSTIC TEST RESULTS			RFI TEST RESULTS		
Test	Result	Time	Test	g/210L	Time
Voltage/Current Test	Pass		Air Blank	0.000	16:43
RAM Test	Pass		Subject Test	RFI*	16:44
EEPROM Checksum Test	Pass		Air Blank	0.000	16:44
Real Time Clock Test	Pass		*RFI Detect		
DSP Test	Pass		Pass		
Analytical Stability Test	Pass				
Modem Test	Pass				
Temperature Regulation Test	Pass				

NUMBER OF REFUSALS AND SUBJECT BREATH TESTS IN EACH RANGE SINCE LAST MAINTENANCE REPORT						
REFUSALS	.00-.04	.05-.09	.10-.14	.15-.19	OVER .19	
0	4	0	0	0	0	0

List any new parts and describe any alteration or modification that was made to restore the instrument to operate satisfactorily and within established limits (use other side if necessary).
INSTRUMENT WAS WITHIN STATE STANDARDS

INSPECTING OFFICER		
SIGNATURE <i>Ron Baltzer #1460</i>	PRINT NAME RON BALTZER	
TYPE II PERMIT NUMBER 290150	EXPIRATION DATE 07/17/2021	TELEPHONE NUMBER 816-325-7300



Airgas USA LLC (LAB)
 3500 Bernard Street
 St. Louis, Mo. 63103
 Ph: (314) 533-3100
 Fax: (314) 533-7328

Certificate of Analysis

Customer Name
 Exclusive Supplier
 Intoximeters, Inc.
 2081 Craig Road
 St. Louis, Mo 63146

Test Date: 7-Oct-2019

Lot # AG928002 Model 108caccd

<u>Exp. Date</u>	<u>Cyl. Type</u>	<u>Component</u>	<u>Certified Concentration</u>
7-Oct-2021	108	Ethanol	0.080 ± 0.002 BrAC (208 ppm)
		Nitrogen	Balance

Certification Traceable to N.I.S.T. RGM and to CRM Ethanol Standards:

<u>RGM Serial No.</u>	<u>Concentration</u>	<u>RGM Serial No.</u>	<u>Concentration</u>
EB0010581	392.1 ppm	EB0010603	393.0 ppm
EB0010570	259.8 ppm	EB0010559	258.2 ppm
EB0010285	208.0 ppm	EB0010595	208.3 ppm
EB0010561	103.6 ppm	EB0010562	104.2 ppm
EB0010681	52.12 ppm	EB0010579	52.81 ppm

<u>CRM Serial No.</u>	<u>Concentration</u>	<u>CRM Serial No.</u>	<u>Concentration</u>
CC434668	800.0 ppm	0056649	390.1 ppm
CC234503	253.0 ppm	0056662	150.2 ppm

Analytical Method: NDIR

Digitally signed by Quality Control
 Date: 2019.10.07 16:34:11 -05:00
 Reason: Dry gas standard certification of analysis
 Location: Airgas USA LLC (Lab)

Approved for Release: _____

Rod Marsala

 Rod Marsala

ISO 17025:2005 A2LA accredited. Certificate Number 3082.06
 ISO 17034:2016 A2LA accredited. Certificate Number 3082.07



STATE OF MISSOURI
DEPARTMENT OF HEALTH AND SENIOR SERVICES
BREATH ALCOHOL PROGRAM



PERMIT
TYPE II
RON BALTZER

is hereby authorized to instruct and supervise operators, train instructors, inspect, calibrate, perform field service and repairs, and operate the following breath analyzer(s):

INTOXILYZER 8000

for the determination of the alcoholic content of blood from a sample of expired air. Permit issued under the provisions of sections 577.020 through 577.041, RSMo and 306.111 through 306.119 RSMo.

DATE 7/17/2019

NUMBER 290150

EXPIRES 7/17/2021

DIRECTOR OF STATE PUBLIC HEALTH LABORATORY

DIRECTOR OF DEPARTMENT OF HEALTH AND SENIOR SERVICES

MO 580-0771 (6-10)

LAB-4 (R6-10)



STATE OF MISSOURI
DEPARTMENT OF HEALTH AND SENIOR SERVICES
BREATH ALCOHOL PROGRAM

INSTRUMENT OPERATOR CARD

The named cardholder is authorized to operate an evidential breath alcohol instrument for the determination of the alcoholic content in breath form of expired air in Missouri.

Operator BALTZER, RON
Permit No 290150
Date Issued 7/17/2019 Date Expires 7/17/2021





MISSOURI DEPARTMENT OF HEALTH AND SENIOR SERVICES
STATE PUBLIC HEALTH LABORATORY
CMI INTOXILYZER 8000 MAINTENANCE REPORT

REPORT #2

Complete this report in duplicate at the time of the regular monthly preventive maintenance check, and whenever instrument is repaired. Send one copy to Department of Health and Senior Services, and retain one copy in department file.

INSTRUMENT SERIAL NUMBER 80-005846	LOCATION OF INSTRUMENT IPD BATVAN STATION 2	DATE OF INSPECTION 08/11/2020	TIME OF INSPECTION 16:48
---------------------------------------	--	----------------------------------	-----------------------------

CALIBRATION CHECK RESULTS			CALIBRATION CHECK SUMMARY		
Test	g/210L	Time	STANDARD TYPE	STANDARD LOT #	STANDARD EXPIRATION DATE
Air Blank	0.000	16:50	DRY	AG928002	10/07/2021
Cal Check	0.079	16:50	SIM TEMPERATURE	SIM SERIAL NUMBER	SIM CERTIFICATE EXPIRATION
Air Blank	0.000	16:51	N/A	N/A	N/A
Cal Check	0.080	16:51	STANDARD VALUE	STANDARD SUPPLIER	
Air Blank	0.000	16:51	0.080	INTOXIMETERS, INC	
Cal Check	0.080	16:52	CALIBRATION CHECK RESULT 1		
Air Blank	0.000	16:52	0.079		
Cal Check	0.080	16:52	CALIBRATION CHECK RESULT 2		
Air Blank	0.000	16:52	0.080		
Pass			CALIBRATION CHECK RESULT 3		
			0.080		
			MAXIMUM DEVIATION (MUST BE WITHIN 5%)	SPREAD (MUST BE .005 OR LESS)	
			1.2%	0.001	

DIAGNOSTIC TEST RESULTS			RFI TEST RESULTS		
			Test	g/210L	Time
Voltage/Current Test	Pass		-----	-----	-----
RAM Test	Pass		Air Blank	0.000	16:53
EEPROM Checksum Test	Pass		Subject Test	RFI*	16:53
Real Time Clock Test	Pass		Air Blank	0.000	16:53
DSP Test	Pass		*RFI Detect		
Analytical Stability Test	Pass				
Modem Test	Pass				
Temperature Regulation Test	Pass				
Pass			Pass		

NUMBER OF REFUSALS AND SUBJECT BREATH TESTS IN EACH RANGE SINCE LAST MAINTENANCE REPORT						
REFUSALS	.00-.04	.05-.09	.10-.14	.15-.19	OVER .19	
0	0	0	0	0	0	0

List any new parts and describe any alteration or modification that was made to restore the instrument to operate satisfactorily and within established limits (use other side if necessary).
INSTRUMENT WITHIN STATE STANDARDS

INSPECTING OFFICER		
SIGNATURE <i>Ron Baltzer #1460</i>	PRINT NAME RON BALTZER	
TYPE II PERMIT NUMBER 290150	EXPIRATION DATE 07/17/2021	TELEPHONE NUMBER 816-325-7300



Airgas USA LLC (LAB)
 3500 Bernard Street
 St. Louis, Mo. 63103
 Ph: (314) 533-3100
 Fax: (314) 533-7328

Certificate of Analysis

Customer Name
 Exclusive Supplier
 Intoximeters, Inc.
 2081 Craig Road
 St. Louis, Mo 63146

Test Date: 7-Oct-2019

Lot # AG928002 Model 108caccd

<u>Exp. Date</u> 7-Oct-2021	<u>Cyl. Type</u> 108	<u>Component</u> Ethanol Nitrogen	<u>Certified Concentration</u> 0.080 ± 0.002 BrAC (208 ppm) Balance
--------------------------------	-------------------------	---	---

Certification Traceable to N.I.S.T. RGM and to CRM Ethanol Standards:

<u>RGM Serial No.</u> EB0010581	<u>Concentration</u> 392.1 ppm	<u>RGM Serial No.</u> EB0010603	<u>Concentration</u> 393.0 ppm
EB0010570	259.8 ppm	EB0010559	258.2 ppm
EB0010285	208.0 ppm	EB0010595	208.3 ppm
EB0010561	103.6 ppm	EB0010562	104.2 ppm
EB0010681	52.12 ppm	EB0010579	52.81 ppm
<u>CRM Serial No.</u> CC434668	<u>Concentration</u> 800.0 ppm	<u>CRM Serial No.</u> 0056649	<u>Concentration</u> 390.1 ppm
CC234503	253.0 ppm	0056662	150.2 ppm

Analytical Method: NDIR

Digitally signed by Quality Control
 Date: 2019.10.07 16:34:11 -05:00
 Reason: Dry gas standard certification of analysis
 Location: Airgas USA LLC (Lab)

Approved for Release:

 Rod Marsala

ISO 17025:2005 A2LA accredited. Certificate Number 3082.06
 ISO 17034:2016 A2LA accredited. Certificate Number 3082.07



STATE OF MISSOURI
 DEPARTMENT OF HEALTH AND SENIOR SERVICES
 BREATH ALCOHOL PROGRAM



PERMIT
TYPE II
RON BALTZER

is hereby authorized to instruct and supervise operators, train instructors, inspect, calibrate, perform field service and repairs, and operate the following breath analyzer(s):

INTOXILYZER 8000

for the determination of the alcoholic content of blood from a sample of expired air. Permit issued under the provisions of sections 577.020 through 577.041, RSMo and 306.111 through 306.119 RSMo.

DATE 7/17/2019

NUMBER 290150

EXPIRES 7/17/2021

DIRECTOR OF STATE PUBLIC HEALTH LABORATORY

DIRECTOR OF DEPARTMENT OF HEALTH AND SENIOR SERVICES

MO 580-0771 (6-10)

LAB-4 (R6-10)



STATE OF MISSOURI
 DEPARTMENT OF HEALTH AND SENIOR SERVICES
 BREATH ALCOHOL PROGRAM

INSTRUMENT OPERATOR CARD

The named cardholder is authorized to operate an evidential breath alcohol instrument for the determination of the alcoholic content in breath form of expired air in Missouri.

Operator **BALTZER, RON**
 Permit No **290150**
 Date Issued **7/17/2019** Date Expires **7/17/2021**





MISSOURI DEPARTMENT OF HEALTH AND SENIOR SERVICES
 STATE PUBLIC HEALTH LABORATORY
CMI INTOXILYZER 8000 MAINTENANCE REPORT

REPORT #2

Complete this report in duplicate at the time of the regular monthly preventive maintenance check, and whenever instrument is repaired. Send one copy to Department of Health and Senior Services, and retain one copy in department file.

INSTRUMENT SERIAL NUMBER 80-005847	LOCATION OF INSTRUMENT IPD DETENTION	DATE OF INSPECTION 08/11/2020	TIME OF INSPECTION 15:54
---------------------------------------	---	----------------------------------	-----------------------------

CALIBRATION CHECK RESULTS			CALIBRATION CHECK SUMMARY		
Test	g/210L	Time	STANDARD TYPE DRY	STANDARD LOT # AG928002	STANDARD EXPIRATION DATE 10/07/2021
Air Blank	0.000	15:55	SIM TEMPERATURE N/A	SIM SERIAL NUMBER N/A	SIM CERTIFICATE EXPIRATION N/A
Cal Check	0.079	15:56	STANDARD VALUE 0.080	STANDARD SUPPLIER INTOXIMETERS, INC	
Air Blank	0.000	15:56	CALIBRATION CHECK RESULT 1 0.079		
Cal Check	0.079	15:57	CALIBRATION CHECK RESULT 2 0.079		
Air Blank	0.000	15:57	CALIBRATION CHECK RESULT 3 0.079		
Cal Check	0.079	15:57	MAXIMUM DEVIATION (MUST BE WITHIN 5%) 1.2%		
Air Blank	0.000	15:58	SPREAD (MUST BE .005 OR LESS) 0.000		

Pass

DIAGNOSTIC TEST RESULTS			RFI TEST RESULTS		
Voltage/Current Test	Pass		Test	g/210L	Time
RAM Test	Pass		Air Blank	0.000	15:58
EEPROM Checksum Test	Pass		Subject Test	RFI*	15:59
Real Time Clock Test	Pass		Air Blank	0.000	15:59
DSP Test	Pass		*RFI Detect		
Analytical Stability Test	Pass				
Modem Test	Pass				
Temperature Regulation Test	Pass				

Pass

NUMBER OF REFUSALS AND SUBJECT BREATH TESTS IN EACH RANGE SINCE LAST MAINTENANCE REPORT						
REFUSALS	.00-.04	.05-.09	.10-.14	.15-.19	OVER .19	
0	9	0	1	1	1	

List any new parts and describe any alteration or modification that was made to restore the instrument to operate satisfactorily and within established limits (use other side if necessary).

INSTRUMENT WAS WITHIN STATE STANDARDS

INSPECTING OFFICER		
SIGNATURE <i>Ron Baltzer #1960</i>	PRINT NAME RON BALTZER	
TYPE II PERMIT NUMBER 290150	EXPIRATION DATE 07/17/2021	TELEPHONE NUMBER 816-325-7300



Airgas USA LLC (LAB)
 3500 Bernard Street
 St. Louis, Mo. 63103
 Ph: (314) 533-3100
 Fax: (314) 533-7328

Certificate of Analysis

Customer Name
 Exclusive Supplier
 Intoximeters, Inc.
 2081 Craig Road
 St. Louis, Mo 63146

Test Date: 7-Oct-2019

Lot # AG928002 Model 108caccd

<u>Exp. Date</u> 7-Oct-2021	<u>Cyl. Type</u> 108	<u>Component</u> Ethanol Nitrogen	<u>Certified Concentration</u> 0.080 ± 0.002 BrAC (208 ppm) Balance
--------------------------------	-------------------------	---	---

Certification Traceable to N.I.S.T. RGM and to CRM Ethanol Standards:

<table border="0"> <tr><td><u>RGM Serial No.</u></td><td><u>Concentration</u></td></tr> <tr><td>EB0010581</td><td>392.1 ppm</td></tr> <tr><td>EB0010570</td><td>259.8 ppm</td></tr> <tr><td>EB0010285</td><td>208.0 ppm</td></tr> <tr><td>EB0010561</td><td>103.6 ppm</td></tr> <tr><td>EB0010681</td><td>52.12 ppm</td></tr> </table>	<u>RGM Serial No.</u>	<u>Concentration</u>	EB0010581	392.1 ppm	EB0010570	259.8 ppm	EB0010285	208.0 ppm	EB0010561	103.6 ppm	EB0010681	52.12 ppm	<table border="0"> <tr><td><u>RGM Serial No.</u></td><td><u>Concentration</u></td></tr> <tr><td>EB0010603</td><td>393.0 ppm</td></tr> <tr><td>EB0010559</td><td>258.2 ppm</td></tr> <tr><td>EB0010595</td><td>208.3 ppm</td></tr> <tr><td>EB0010562</td><td>104.2 ppm</td></tr> <tr><td>EB0010579</td><td>52.81 ppm</td></tr> </table>	<u>RGM Serial No.</u>	<u>Concentration</u>	EB0010603	393.0 ppm	EB0010559	258.2 ppm	EB0010595	208.3 ppm	EB0010562	104.2 ppm	EB0010579	52.81 ppm
<u>RGM Serial No.</u>	<u>Concentration</u>																								
EB0010581	392.1 ppm																								
EB0010570	259.8 ppm																								
EB0010285	208.0 ppm																								
EB0010561	103.6 ppm																								
EB0010681	52.12 ppm																								
<u>RGM Serial No.</u>	<u>Concentration</u>																								
EB0010603	393.0 ppm																								
EB0010559	258.2 ppm																								
EB0010595	208.3 ppm																								
EB0010562	104.2 ppm																								
EB0010579	52.81 ppm																								
<table border="0"> <tr><td><u>CRM Serial No.</u></td><td><u>Concentration</u></td></tr> <tr><td>CC434668</td><td>800.0 ppm</td></tr> <tr><td>CC234503</td><td>253.0 ppm</td></tr> </table>	<u>CRM Serial No.</u>	<u>Concentration</u>	CC434668	800.0 ppm	CC234503	253.0 ppm	<table border="0"> <tr><td><u>CRM Serial No.</u></td><td><u>Concentration</u></td></tr> <tr><td>0056649</td><td>390.1 ppm</td></tr> <tr><td>0056662</td><td>150.2 ppm</td></tr> </table>	<u>CRM Serial No.</u>	<u>Concentration</u>	0056649	390.1 ppm	0056662	150.2 ppm												
<u>CRM Serial No.</u>	<u>Concentration</u>																								
CC434668	800.0 ppm																								
CC234503	253.0 ppm																								
<u>CRM Serial No.</u>	<u>Concentration</u>																								
0056649	390.1 ppm																								
0056662	150.2 ppm																								

Analytical Method: NDIR

Digitally signed by Quality Control
 Date: 2019.10.07 16:34:11 -05:00
 Reason: Dry gas standard certification of analysis
 Location: Airgas USA LLC (Lab)

Approved for Release:

 Rod Marsala

ISO 17025:2005 A2LA accredited. Certificate Number 3082.06
 ISO 17034:2016 A2LA accredited. Certificate Number 3082.07



STATE OF MISSOURI
 DEPARTMENT OF HEALTH AND SENIOR SERVICES
 BREATH ALCOHOL PROGRAM



PERMIT
TYPE II
RON BALTZER

is hereby authorized to instruct and supervise operators, train instructors, inspect, calibrate, perform field service and repairs, and operate the following breath analyzer(s):

INTOXILYZER 8000

for the determination of the alcoholic content of blood from a sample of expired air. Permit issued under the provisions of sections 577.020 through 577.041, RSMo and 306.111 through 306.119 RSMo.

DATE 7/17/2019

NUMBER 290150

EXPIRES 7/17/2021

DIRECTOR OF STATE PUBLIC HEALTH LABORATORY

DIRECTOR OF DEPARTMENT OF HEALTH AND SENIOR SERVICES

MO 680-0771 (6-10)

LAB-4 (R6-10)

STATE OF MISSOURI
 DEPARTMENT OF HEALTH AND SENIOR SERVICES
 BREATH ALCOHOL PROGRAM

INSTRUMENT OPERATOR CARD

The named cardholder is authorized to operate an evidential breath alcohol instrument for the determination of the alcoholic content in breath form of expired air in Missouri.

Operator **BALTZER, RON**
 Permit No **290150**
 Date issued **7/17/2019** Date Expires **7/17/2021**