

RECEIVED

By Tracy Crews at 8:38 am, Mar 19, 2020



MISSOURI DEPARTMENT OF HEALTH AND SENIOR SERVICES
STATE PUBLIC HEALTH LABORATORY
CMI INTOXILYZER 8000 MAINTENANCE REPORT

REPORT #2

Complete this report in duplicate at the time of the regular monthly preventive maintenance check, and whenever instrument is repaired. Send one copy to Department of Health and Senior Services, and retain one copy in department file.

INSTRUMENT SERIAL NUMBER 80-005837	LOCATION OF INSTRUMENT KANSAS CITY PD	DATE OF INSPECTION 03/05/2020	TIME OF INSPECTION 22:26
---------------------------------------	------------------------------------------	----------------------------------	-----------------------------

CALIBRATION CHECK RESULTS			CALIBRATION CHECK SUMMARY		
Test	g/210L	Time	STANDARD TYPE DRY	STANDARD LOT # 00919080A1	STANDARD EXPIRATION DATE 03/05/2021
Air Blank	0.000	22:28	SIM TEMPERATURE N/A	SIM SERIAL NUMBER N/A	SIM CERTIFICATE EXPIRATION N/A
Cal Check	0.084	22:28	STANDARD VALUE 0.080	STANDARD SUPPLIER CMI	
Air Blank	0.000	22:29	CALIBRATION CHECK RESULT 1 0.084		
Cal Check	0.083	22:29	CALIBRATION CHECK RESULT 2 0.083		
Air Blank	0.000	22:30	CALIBRATION CHECK RESULT 3 0.083		
Cal Check	0.083	22:30	MAXIMUM DEVIATION (MUST BE WITHIN 5%) 5.0%		
Air Blank	0.000	22:31	SPREAD (MUST BE .005 OR LESS) 0.001		
Pass					


DIAGNOSTIC TEST RESULTS			RFI TEST RESULTS		
Voltage/Current Test	Pass		Test	g/210L	Time
RAM Test	Pass		Air Blank	RFI*	22:31
EEPROM Checksum Test	Pass		Air Blank	0.000	22:31
Real Time Clock Test	Pass		*RFI Detect		
DSP Test	Pass				
Analytical Stability Test	Pass				
Modem Test	Pass				
Temperature Regulation Test	Pass				
Pass			Pass		

NUMBER OF REFUSALS AND SUBJECT BREATH TESTS IN EACH RANGE SINCE LAST MAINTENANCE REPORT						
REFUSALS	.00-.04	.05-.09	.10-.14	.15-.19	OVER .19	
0	0	0	0	0	0	0

List any new parts and describe any alteration or modification that was made to restore the instrument to operate satisfactorily and within established limits (use other side if necessary).

INSTRUMENT TESTED AND CERTIFIED

INSPECTING OFFICER		
SIGNATURE 	PRINT NAME JORDAN INFRANCA	
TYPE II PERMIT NUMBER 290130	EXPIRATION DATE 06/21/2021	TELEPHONE NUMBER 8163825897



7 Eastgate Dr • P.O. Box 790 • Jackson, MO 64301-0790
217-245-2103 • Fax: 217-245-7654 • www.wilmcoproducts.com

Certificate of Analysis

Certificate ID: 11721
Part #: BAC10510801
Cylinder Size: 105L
Lot Number: 00919080A1
Expiration: 3/5/2021

0.080 BAC (for the calibration of instruments used to determine breath alcohol concentration)

Contents: 105 Liters @ 1000 psig 70°F (21°C)
Component: Reported Analytical Accuracy
Concentration: (U, L=2); Analytical Method
Ethanol: 208 ppm (1/2 sec, BAC/2000) NDIR
Nitrogen: 5.2 ppm

Distributed by: CMI Inc.
316 East Ninth Street,
Owensboro, KY 42303
Phone 866-635-0690
www.alcoholtest.com

NIIST Traceable Reference Material
Cylinder No. CCI2/5533 / Job No. 09160304
Certified 365.2 umol/mol Ethanol in Nitrogen

Store in dry area, away from sources of heat,
ignition and direct sunlight. Do not allow storage
area to exceed 52 °C (125 °F).

Specialty Gas Lab Tech

Date: 01-24-19



ISO/IEC 17025:2005 Accredited Laboratory

The calibration results which this certificate were obtained using equipment and methods capable of producing results which are traceable to NIST and only on the item recorded on this certificate. LENO Product Company makes no warranty or representation of any kind regarding the accuracy of the information provided for any particular purpose. The information on this certificate is based on the data of the test, which shall be limited to a standard of measurement and of the material or service.



STATE OF MISSOURI
DEPARTMENT OF HEALTH AND SENIOR SERVICES
BREATH ALCOHOL PROGRAM

PERMIT
TYPE II

JORDAN INFRANCA

I hereby authorize to install, and supervise, operate, train, instructors, inspect, calibrate, perform field service and repairs, and operate the following breath analyzer(s):
ALCO-SENSOR IV WITH PRINTER, INTOXILYZER 8000
for the determination of the alcohol content of blood from a sample of expired air. Permit issued under the provisions of sections 577.080 through 577.094, RSMo and 608.111 through 608.115 RSMo.

DATE: 6/21/2019
NUMBER: 290130
EXPIRES: 6/21/2021

DIRECTOR OF STATE PUBLIC HEALTH LABORATORY

DIRECTOR OF DEPARTMENT OF HEALTH AND SENIOR SERVICES
L&H 60001

STATE OF MISSOURI
DEPARTMENT OF HEALTH AND SENIOR SERVICES
BREATH ALCOHOL PROGRAM
INSTRUMENT OPERATOR CARD
The instrument operator is authorized to operate an approved breath alcohol instrument for the determination of the alcohol content in breath from an individual in Missouri.
Operator: INFRANCA, JORDAN
Permit No: 290130
Data Expires: 6/21/2021