

**RECEIVED**

By Tracy Crews at 11:33 am, Nov 09, 2020



MISSOURI DEPARTMENT OF HEALTH AND SENIOR SERVICES  
STATE PUBLIC HEALTH LABORATORY  
**CMI INTOXILYZER 8000 MAINTENANCE REPORT**

EPO

REPORT #2

Complete this report in duplicate at the time of the regular monthly preventive maintenance check, and whenever instrument is repaired. Send one copy to Department of Health and Senior Services, and retain one copy in department file.

INSTRUMENT SERIAL NUMBER 80-005837	LOCATION OF INSTRUMENT KANSAS CITY PD	DATE OF INSPECTION 11/03/2020	TIME OF INSPECTION 00:44
---------------------------------------	--	----------------------------------	-----------------------------

CALIBRATION CHECK RESULTS			CALIBRATION CHECK SUMMARY		
Test	g/210L	Time	STANDARD TYPE	STANDARD LOT #	STANDARD EXPIRATION DATE
Air Blank	0.000	00:45	DRY	14020080A2	07/05/2022
Cal Check	0.082	00:46	SIM TEMPERATURE	SIM SERIAL NUMBER	SIM CERTIFICATE EXPIRATION
Air Blank	0.000	00:46	N/A	N/A	N/A
Cal Check	0.082	00:47	STANDARD VALUE	STANDARD SUPPLIER	
Air Blank	0.000	00:47	0.080	CMI	
Cal Check	0.081	00:47	CALIBRATION CHECK RESULT 1		
Air Blank	0.000	00:48	0.082		
			CALIBRATION CHECK RESULT 2		
			0.082		
			CALIBRATION CHECK RESULT 3		
			0.081		
			MAXIMUM DEVIATION (MUST BE WITHIN 5%)	SPREAD (MUST BE .005 OR LESS)	
			2.5%	0.001	

**Pass**

DIAGNOSTIC TEST RESULTS			RFI TEST RESULTS		
Voltage/Current Test	Pass		Test	g/210L	Time
RAM Test	Pass		Air Blank	RFI*	00:48
EEPROM Checksum Test	Pass		Air Blank	0.000	00:49
Real Time Clock Test	Pass		*RFI Detect		
DSP Test	Pass				
Analytical Stability Test	Pass				
Modem Test	Pass				
Temperature Regulation Test	Pass				

**Pass****Pass**

NUMBER OF REFUSALS AND SUBJECT BREATH TESTS IN EACH RANGE SINCE LAST MAINTENANCE REPORT						
REFUSALS	.00-.04	.05-.09	.10-.14	.15-.19	OVER .19	
1	1	1	2	0	2	

List any new parts and describe any alteration or modification that was made to restore the instrument to operate satisfactorily and within established limits (use other side if necessary).

TESTED AND CERTIFIED 11-03-2020

**INSPECTING OFFICER**

SIGNATURE 	PRINT NAME JORDAN INFRANCA	
TYPE II PERMIT NUMBER 290130	EXPIRATION DATE 06/21/2021	TELEPHONE NUMBER 816-482-8195

STATE OF MISSOURI  
DEPARTMENT OF HEALTH AND SENIOR SERVICES  
PERMIT ALCOHOL PROGRAM

2


PERMIT  
TYPE II

JORDAN INFRANCA

is hereby authorized to utilize and operate the equipment, when installed, internal, external, combined, perform field service and repair, and operate the following: ALCO-SHNSOR IV; WEH-PRINTER; INTOXILYZER 8000

On the date indicated this alcoholic content of blood/urine sample of expired air Permit issued under the provisions of sections 57.020 through 57.044 F.S.M.S. and 57.045 through 57.050, F.S.M.S.

DATE: 02/12/20  
NUMBER: 2000130  
EXPIRES: 02/12/21  
OPERATOR: [Signature]  
SIGNATURE OF OPERATOR OF EQUIPMENT: [Signature]  
DATE: 02/12/20



7 Eastgate Dr. • P.O. Box 790 • Jacksonville, IL 62651-0790  
217-245-2183 • Fax: 217-243-7634 • www.ilmoproducts.com

Certificate of Analysis

Certificate ID: 13021  
Part #: BAC105L080T  
Cylinder Size: 105L  
Lot Number: 14020080A2  
Expiration: 7/5/2022

0.080 BAC (For the calibration of instruments used to determine breath alcohol concentration)

Contents: 105 Liters @ 1000 psig 70°F (21°C)

Component	Reported Concentration:	Accuracy (U, k=2):	Analytical Method:	Distributed by:
Ethanol	288 ppm	+/- 0.002 BAC (6/218L)	NDIR	CMI Inc
Nitrogen	Balance	[5.2 ppm]		316 East Ninth Street Owensboro, KY 42303 Phone 866-835-0690 www.alcoholtest.com

\*Traceable to: Certified Reference Material - 262.4 µmol/mol  
Ethanol in Nitrogen - Serial No. GN0015026 Lot No. 050319E11  
Store in dry area, away from sources of heat, ignition and direct sunlight. Do not allow storage area to exceed 52 °C (125 °F).

Specialty Gas Lab Tech  
[Signature]  
[Signature]

06-15-2020  
Issuance Date



ISO/IEC 17025:2017 Accredited Laboratory

The calibration results within this certificate were obtained using equipment and standards capable of producing analytical results traceable to NIST, and apply only to the items contained on this certificate. ILMO Products Company makes no warranty or representation as to the suitability of the use of any information provided for any particular purpose. The information use is at the sole discretion and risk of the user. Liability shall be limited to established replacement cost of the material or service.