

RECEIVED

By Tracy Crews at 8:38 am, Mar 19, 2020



MISSOURI DEPARTMENT OF HEALTH AND SENIOR SERVICES
STATE PUBLIC HEALTH LABORATORY
CMI INTOXILYZER 8000 MAINTENANCE REPORT

REPORT #2

Complete this report in duplicate at the time of the regular monthly preventive maintenance check, and whenever instrument is repaired. Send one copy to Department of Health and Senior Services, and retain one copy in department file.

INSTRUMENT SERIAL NUMBER 80-005838	LOCATION OF INSTRUMENT KCMO POLICE DEPT.	DATE OF INSPECTION 03/05/2020	TIME OF INSPECTION 22:47
---------------------------------------	---	----------------------------------	-----------------------------

CALIBRATION CHECK RESULTS

Test	g/210L	Time
Air Blank	0.000	22:48
Cal Check	0.082	22:49
Air Blank	0.000	22:49
Cal Check	0.081	22:50
Air Blank	0.000	22:50
Cal Check	0.081	22:50
Air Blank	0.000	22:51

Pass

CALIBRATION CHECK SUMMARY

STANDARD TYPE DRY	STANDARD LOT # 00919080A1	STANDARD EXPIRATION DATE 03/05/2021
SIM TEMPERATURE N/A	SIM SERIAL NUMBER N/A	SIM CERTIFICATE EXPIRATION N/A
STANDARD VALUE 0.080	STANDARD SUPPLIER CMI	
CALIBRATION CHECK RESULT 1 0.082		
CALIBRATION CHECK RESULT 2 0.081		
CALIBRATION CHECK RESULT 3 0.081		
MAXIMUM DEVIATION (MUST BE WITHIN 5%) 2.5%		SPREAD (MUST BE .005 OR LESS) 0.001

DIAGNOSTIC TEST RESULTS

Voltage/Current Test	Pass
RAM Test	Pass
EEPROM Checksum Test	Pass
Real Time Clock Test	Pass
DSP Test	Pass
Analytical Stability Test	Pass
Modem Test	Pass
Temperature Regulation Test	Pass

Pass

RFI TEST RESULTS

Test	g/210L	Time
Air Blank	RFI*	22:51
Air Blank	0.000	22:52
*RFI Detect		

Pass

NUMBER OF REFUSALS AND SUBJECT BREATH TESTS IN EACH RANGE SINCE LAST MAINTENANCE REPORT

REFUSALS	.00-.04	.05-.09	.10-.14	.15-.19	OVER .19
0	54	0	0	0	0

List any new parts and describe any alteration or modification that was made to restore the instrument to operate satisfactorily and within established limits (use other side if necessary).

TESTED AND CERTIFIED

INSPECTING OFFICER

SIGNATURE 	PRINT NAME JORDAN INFRANCA
---------------	-------------------------------

TYPE II PERMIT NUMBER 290130	EXPIRATION DATE 06/21/2021	TELEPHONE NUMBER 8163825897
---------------------------------	-------------------------------	--------------------------------

