

5/1/20

RECEIVED

By Tracy Crews at 12:12 pm, Aug 14, 2020



MISSOURI DEPARTMENT OF HEALTH AND SENIOR SERVICES
STATE PUBLIC HEALTH LABORATORY

CMI INTOXILYZER 8000 MAINTENANCE REPORT

REPORT #2

Complete this report in duplicate at the time of the regular monthly preventive maintenance check, and whenever instrument is repaired. Send one copy to Department of Health and Senior Services, and retain one copy in department file.

INSTRUMENT SERIAL NUMBER 80-005838	LOCATION OF INSTRUMENT KCMO POLICE DEPT.	DATE OF INSPECTION 08/06/2020	TIME OF INSPECTION 21:55
---------------------------------------	---	----------------------------------	-----------------------------

CALIBRATION CHECK RESULTS			CALIBRATION CHECK SUMMARY		
Test	g/210L	Time	STANDARD TYPE	STANDARD LOT #	STANDARD EXPIRATION DATE
Air Blank	0.000	21:57	DRY	00919080A1	03/05/2021
Cal Check	0.081	21:57	SIM TEMPERATURE	SIM SERIAL NUMBER	SIM CERTIFICATE EXPIRATION
Air Blank	0.000	21:58	N/A	N/A	N/A
Cal Check	0.080	21:58	STANDARD VALUE	STANDARD SUPPLIER	
Air Blank	0.000	21:58	0.080	CMI	
Cal Check	0.079	21:59	CALIBRATION CHECK RESULT 1		
Air Blank	0.000	21:59	0.081		
Cal Check	0.079	21:59	CALIBRATION CHECK RESULT 2		
Air Blank	0.000	21:59	0.080		
Pass			CALIBRATION CHECK RESULT 3		
			0.079		
			MAXIMUM DEVIATION (MUST BE WITHIN 5%)	SPREAD (MUST BE .005 OR LESS)	
			1.2%	0.002	

DIAGNOSTIC TEST RESULTS			RFI TEST RESULTS		
Test	Pass/Fail		Test	g/210L	Time
Voltage/Current Test	Pass		Air Blank	0.000	22:00
RAM Test	Pass		Subject Test	RFI*	22:00
EEPROM Checksum Test	Pass		Air Blank	0.000	22:01
Real Time Clock Test	Pass		*RFI Detect		
DSP Test	Pass		Pass		
Analytical Stability Test	Pass				
Modem Test	Pass				
Temperature Regulation Test	Pass				
Pass					

NUMBER OF REFUSALS AND SUBJECT BREATH TESTS IN EACH RANGE SINCE LAST MAINTENANCE REPORT						
REFUSALS	.00-.04	.05-.09	.10-.14	.15-.19	OVER .19	
0	2	0	1	1	1	

List any new parts and describe any alteration or modification that was made to restore the instrument to operate satisfactorily and within established limits (use other side if necessary).

TESTED AND CERTIFIED

INSPECTING OFFICER

SIGNATURE 	PRINT NAME JEREMY WHITE
---------------	----------------------------

TYPE II PERMIT NUMBER 280275	EXPIRATION DATE 09/05/2020	TELEPHONE NUMBER 8163825881
---------------------------------	-------------------------------	--------------------------------



STATE OF MISSOURI
DEPARTMENT OF HEALTH AND SENIOR SERVICES
BREATH ALCOHOL PROGRAM



**PERMIT
TYPE II**

JEREMY A. WHITE

is hereby authorized to instruct and supervise operators, train instructors, inspect, calibrate, perform field service and repairs, and operate the following breath analyzer(s):

ALCO-SENSOR IV WITH PRINTER, INTOXILYZER 8000

for the determination of the alcoholic content of blood from a sample of expired air. Permit issued under the provisions of sections 577.020 through 577.041, RSMo and 306.111 through 306.119 RSMo.

DATE 9/5/2018
NUMBER 280275
EXPIRES 9/5/2020
0.884-5771 (8-10)

W.A. White
DIRECTOR OF STATE PUBLIC HEALTH LABORATORY

Jeremy A. White
DIRECTOR OF DEPARTMENT OF HEALTH AND SENIOR SERVICES
LAB-4 (RIS-10)

STATE OF MISSOURI SENIOR SERVICES
DEPARTMENT OF HEALTH AND SENIOR SERVICES
BREATH ALCOHOL PROGRAM

INSTRUMENT OPERATOR CARD

The named operator is authorized to operate an evidential breath alcohol analyzer for the determination of the alcoholic content in breath form of expired air in Missouri.

Operator: WHITE, JEREMY
Permit No: 280275
Date Issued: 9/5/2018 Date Expires: 9/5/2020



7 Eastgate Dr. • P.O. Box 790 • Jacksonville, IL 62651-0790
217-243-2183 • Fax: 217-243-7634 • www.ilmoproducts.com

Certificate of Analysis

Certificate ID: 11721
Part #: BAC105L080T
Cylinder Size: 105L
Lot Number: 00919080A1
Expiration: 3/5/2021

0.080 BAC (For the calibration of instruments used to determine breath alcohol concentration)

Contents: 105 Liters @ 1000 psig 70°F (21°C)

Component	Reported Concentration:	Accuracy (U, I=2):	Analytical Method:
Ethanol	208 ppm	+/- 0.002 DIC(G/20A)	NOZTR
Nitrogen	Balance	[± 0.2 ppm]	

Distributed by:
CMA Inc.
316 East Ninth Street
Cynthiana, KY 42303
Phone 866-835-0690
www.cmafirstbest.com

Store in dry area, away from sources of heat, ignition and direct sunlight. Do not allow storage area to exceed 52 °C (125 °F).

Jeremy A. White
Specialty Gas Lab Tech
Date 01-24-19



The calibration results within this certificate were obtained using equipment and standards capable of producing analytical results traceable to NIST, and apply only to the hours contained on this certificate. ILMO Products Company makes no warranty or representation as to the reliability of the use of any information provided for any particular purpose. This information use for the sole discretion and risk of the user. Liability shall be limited to established replacement cost of this material or service.

ISO/IEC 17025:2005 Accredited Laboratory