

RECEIVED

By Tracy Crews at 11:33 am, Nov 09, 2020



MISSOURI DEPARTMENT OF HEALTH AND SENIOR SERVICES
STATE PUBLIC HEALTH LABORATORY
CMI INTOXILYZER 8000 MAINTENANCE REPORT

REPORT #2

Complete this report in duplicate at the time of the regular monthly preventive maintenance check, and whenever instrument is repaired. Send one copy to Department of Health and Senior Services, and retain one copy in department file.

INSTRUMENT SERIAL NUMBER 80-005838		LOCATION OF INSTRUMENT KCMO POLICE DEPT.		DATE OF INSPECTION 11/02/2020		TIME OF INSPECTION 23:26	
CALIBRATION CHECK RESULTS				CALIBRATION CHECK SUMMARY			
Test	g/210L	Time		STANDARD TYPE DRY	STANDARD LOT # 14020080A2	STANDARD EXPIRATION DATE 07/05/2022	
Air Blank	0.000	23:34		SIM TEMPERATURE N/A	SIM SERIAL NUMBER N/A	SIM CERTIFICATE EXPIRATION N/A	
Cal Check	0.081	23:34		STANDARD VALUE 0.080	STANDARD SUPPLIER CMI		
Air Blank	0.000	23:35		CALIBRATION CHECK RESULT 1 0.081			
Cal Check	0.080	23:35		CALIBRATION CHECK RESULT 2 0.080			
Air Blank	0.000	23:35		CALIBRATION CHECK RESULT 3 0.081			
Cal Check	0.081	23:36		MAXIMUM DEVIATION (MUST BE WITHIN 5%) 1.2%		SPREAD (MUST BE .005 OR LESS) 0.001	
Air Blank	0.000	23:36		Pass			
DIAGNOSTIC TEST RESULTS				RFI TEST RESULTS			
Voltage/Current Test		Pass		Test	g/210L	Time	
RAM Test		Pass		-----			
EEPROM Checksum Test		Pass		Air Blank	RFI*	23:36	
Real Time Clock Test		Pass		Air Blank	0.000	23:37	
DSP Test		Pass		*RFI Detect			
Analytical Stability Test		Pass		Pass			
Modem Test		Pass					
Temperature Regulation Test		Pass					
NUMBER OF REFUSALS AND SUBJECT BREATH TESTS IN EACH RANGE SINCE LAST MAINTENANCE REPORT							
REFUSALS	.00-.04	.05-.09	.10-.14	.15-.19	OVER .19		
0	0	0	0	0	0	0	
List any new parts and describe any alteration or modification that was made to restore the instrument to operate satisfactorily and within established limits (use other side if necessary). TESTED AND CERTIFIED 11-02-2020							
INSPECTING OFFICER							
SIGNATURE <i>P.O. Infranca</i>				PRINT NAME JORDAN INFRANCA			
TYPE II PERMIT NUMBER 290130		EXPIRATION DATE 06/21/2021		TELEPHONE NUMBER 816-482-8195			

STATE OF MISSOURI
DEPARTMENT OF HEALTH AND SENIOR SERVICES
BREATH ALCOHOL PROGRAM

2

PERMIT
TYPE II

JORDAN INERANCA

is hereby authorized to utilize and operate the following breathalyzer:


ALCO SENSOR IV WITH PRINTER INTOXILYZER 8000
for the determination of the alcoholic content of blood from a sample of expired air. Permit issued under the provisions of sections 577.020 through 577.041, RSMo, and Chapter 1 through 600, IFRSMo.

DATE 04/21/21
NUMBER 390130
EXPIRES 04/21/21
OPERATOR

W. W. S. S.

MISSOURI DEPARTMENT OF HEALTH AND SENIOR SERVICES

MISSOURI DEPARTMENT OF HEALTH AND SENIOR SERVICES



Certificate of Analysis



7 Eastgate Dr. • P.O. Box 790 • Jacksonville, IL 62651-0790
217-245-2183 • Fax: 217-243-7634 • www.ilmo-products.com

Certificate ID: 13021
Part #: BAC105L080T
Cylinder Size: 105L
Lot Number: 14020080A2
Expiration: 7/5/2022

0.080 BAC (For the calibration of instruments used to determine breath alcohol concentration)

Contents: 105 Liters @ 1000 psig 70°F (21°C)

Component	Reported Concentration:	Analytical Accuracy (U, k=2):	Analytical Method:	Distributed by:
Ethanol	208 ppm	±0.002 BAC (6/210L) NDIR	NDIR	CMI Inc.
Nitrogen	Balance	[5.2 ppm]		316 East Ninth Street Owensboro, KY 42303 Phone 866-835-0690 www.alcoholtest.com

*Traceable to:
Certified Reference Material - 262.4 µmol/mol
Ethanol in Nitrogen - Serial No. GN0015026 Lot No. 050319E11

Store in dry area, away from sources of heat, ignition and direct sunlight. Do not allow storage area to exceed 52 °C (125 °F).

Specialty Gas Lab Tech
[Signature]
06-15-2020
Issuance Date



The calibration results within this certificate were obtained using equipment and standards capable of producing analytical results traceable to NIST, and apply only to the items contained on this certificate. ILPO Products Company makes no warranty or representation as to the suitability of the use of any information provided for any particular purpose. The information is for the sole direction and risk of the user. Liability shall be limited to established replacement cost of data material or service.

ISO/IEC 17025:2017 Accredited Laboratory