

130

RECEIVED
By Tracy Crews at 12:12 pm, Aug 14, 2020



MISSOURI DEPARTMENT OF HEALTH AND SENIOR SERVICES
STATE PUBLIC HEALTH LABORATORY

CMI INTOXILYZER 8000 MAINTENANCE REPORT

REPORT #2

Complete this report in duplicate at the time of the regular monthly preventive maintenance check, and whenever instrument is repaired. Send one copy to Department of Health and Senior Services, and retain one copy in department file.

INSTRUMENT SERIAL NUMBER 80-005839	LOCATION OF INSTRUMENT KCMO POLICE DEPT	DATE OF INSPECTION 08/06/2020	TIME OF INSPECTION 22:38
---------------------------------------	--	----------------------------------	-----------------------------

CALIBRATION CHECK RESULTS			CALIBRATION CHECK SUMMARY		
Test	g/210L	Time	STANDARD TYPE	STANDARD LOT #	STANDARD EXPIRATION DATE
Air Blank	0.000	22:43	DRY	00919080A1	03/05/2021
Cal Check	0.080	22:43	SIM TEMPERATURE	SIM SERIAL NUMBER	SIM CERTIFICATE EXPIRATION
Air Blank	0.000	22:44	N/A	N/A	N/A
Cal Check	0.079	22:44	STANDARD VALUE	STANDARD SUPPLIER	
Air Blank	0.000	22:45	0.080	CMI	
Cal Check	0.079	22:45	CALIBRATION CHECK RESULT 1		
Air Blank	0.000	22:46	0.080		
Cal Check	0.079	22:46	CALIBRATION CHECK RESULT 2		
Air Blank	0.000	22:46	0.079		
Cal Check	0.079	22:46	CALIBRATION CHECK RESULT 3		
Air Blank	0.000	22:46	0.079		
Pass			MAXIMUM DEVIATION (MUST BE WITHIN 5%)	SPREAD (MUST BE .005 OR LESS)	
			1.2%	0.001	

DIAGNOSTIC TEST RESULTS			RFI TEST RESULTS		
			Test	g/210L	Time
Voltage/Current Test	Pass		Air Blank	0.000	22:46
RAM Test	Pass		Subject Test	RFI*	22:47
EEPROM Checksum Test	Pass		Air Blank	0.000	22:47
Real Time Clock Test	Pass		*RFI Detect		
DSP Test	Pass				
Analytical Stability Test	Pass				
Modem Test	Pass				
Temperature Regulation Test	Pass				
Pass			Pass		

NUMBER OF REFUSALS AND SUBJECT BREATH TESTS IN EACH RANGE SINCE LAST MAINTENANCE REPORT						
REFUSALS	.00-.04	.05-.09	.10-.14	.15-.19	OVER .19	
0	0	0	0	0	0	0

List any new parts and describe any alteration or modification that was made to restore the instrument to operate satisfactorily and within established limits (use other side if necessary).

TESTED AND CERTIFIED

INSPECTING OFFICER			
SIGNATURE 		PRINT NAME JEREMY WHITE	
TYPE ID PERMIT NUMBER 280275		EXPIRATION DATE 09/05/2020	
		TELEPHONE NUMBER 8163825881	



STATE OF MISSOURI
DEPARTMENT OF HEALTH AND SENIOR SERVICES
BREATH ALCOHOL PROGRAM



PERMIT
TYPE II

JEREMY A. WHITE

is hereby authorized to instruct and supervise operators, train instructors, inspect, calibrate, perform field service and repairs, and operate the following breath analyzer(s):

ALCO-SENSOR IV WITH PRINTER, INTOXILYZER 8000

for the determination of the alcoholic content of blood from a sample of expired air. Permit issued under the provisions of sections 577.020 through 577.041, RSMo and 306.111 through 306.119 RSMo.

DATE 9/5/2018

NUMBER 280275

EXPIRES 9/5/2020

MD 580-071 (6-10)

DIRECTOR OF STATE PUBLIC HEALTH LABORATORY

[Signature]

DIRECTOR OF DEPARTMENT OF HEALTH AND SENIOR SERVICES
LAD-69510

STATE OF MISSOURI
DEPARTMENT OF HEALTH AND SENIOR SERVICES
BREATH ALCOHOL PROGRAM

INSTRUMENT OPERATOR CARD

The named operator is authorized to operate the instrument for the determination of the alcoholic content of breath from a subject in Missouri.

Operator: **WHITE, JEREMY**
Permit No: **280275**
Data Issued: **9/5/2018**
Data Expires: **9/5/2020**



7 Eastgate Dr. • P.O. Box 790 • Jacksonville, IL 62651-0790
217-243-2183 • Fax: 217-243-7634 • www.ilmgo.com

Certificate of Analysis

Certificate ID: 11721
Part #: BACC105L080T
Cylinder Size: 105L
Lot Number: 00919080A1
Expiration: 3/5/2021

0.080 BAC (For the calibration of instruments used to determine breath alcohol concentration)

Contents: 105 Liters @ 1000 psig 70°F (21°C)

Component	Reported Concentration: (U, L ²)	Accuracy (1 ² , L ²)	Analytical Method:
Ethanol	208 ppm	+/- 0.002 BAC(0.200) (1% Error)	NDIR
Nitrogen	Balance		

Distributed by:
CMI Inc.
316 East Ninth Street
Owensboro, KY 42303
Phone 866-835-0690
www.ilmgo.com

⁴NIST Traceable Reference Material
Cylinder No. CCG74523 / Job No. 09160306
Certified 362.2 ppm/100 Ethanol in Nitrogen

Store in dry area, away from sources of heat, ignition and direct sunlight. Do not allow storage area to exceed 52 °C (125 °F).

[Signature]
Specialty Gas Lab Tech

Date: 01-24-19



PILA
Calibration and Testing
Administrative 60180

The calibration results within this certificate were obtained using equipment and standards capable of producing analytical results traceable to NIST, and apply only to the items covered on this certificate. ILMGO Products Company makes no warranty or representation as to the suitability of the use of any information provided for any particular purpose. The information used in this data sheet is not the property of ILMGO and shall be returned to ILMGO upon request.

ISO/IEC 17025:2005 Accredited Laboratory