



MISSOURI DEPARTMENT OF HEALTH AND SENIOR SERVICES
STATE PUBLIC HEALTH LABORATORY

CMI INTOXILYZER 8000 MAINTENANCE REPORT

RECEIVED

By Tracy Crews at 10:58 am, May 18, 2020

MDEC

REPORT #2

Complete this report in duplicate at the time of the regular monthly preventive maintenance check, and whenever instrument is repaired. Send one copy to Department of Health and Senior Services, and retain one copy in department file.

INSTRUMENT SERIAL NUMBER 80-005840	LOCATION OF INSTRUMENT KCMO POLICE DEPT.	DATE OF INSPECTION 05/06/2020	TIME OF INSPECTION 03:13
---------------------------------------	---	----------------------------------	-----------------------------

CALIBRATION CHECK RESULTS			CALIBRATION CHECK SUMMARY		
Test	g/210L	Time	STANDARD TYPE DRY	STANDARD LOT # 00919080A1	STANDARD EXPIRATION DATE 03/05/2021
Air Blank	0.000	03:15	SIM TEMPERATURE N/A	SIM SERIAL NUMBER N/A	SIM CERTIFICATE EXPIRATION N/A
Cal Check	0.083	03:15	STANDARD VALUE 0.080	STANDARD SUPPLIER CMI	
Air Blank	0.000	03:16	CALIBRATION CHECK RESULT 1 0.083		
Cal Check	0.082	03:16	CALIBRATION CHECK RESULT 2 0.082		
Air Blank	0.000	03:16	CALIBRATION CHECK RESULT 3 0.083		
Cal Check	0.083	03:17	MAXIMUM DEVIATION (MUST BE WITHIN 5%) 3.7%		
Air Blank	0.000	03:17	SPREAD (MUST BE .005 OR LESS) 0.001		

Pass

DIAGNOSTIC TEST RESULTS			RFI TEST RESULTS		
Voltage/Current Test	Pass		Test	g/210L	Time
RAM Test	Pass		Air Blank	0.000	03:18
EEPROM Checksum Test	Pass		Subject Test	RFI*	03:18
Real Time Clock Test	Pass		Air Blank	0.000	03:19
DSP Test	Pass		*RFI Detect		
Analytical Stability Test	Pass				
Modem Test	Pass				
Temperature Regulation Test	Pass				

Pass

Pass

NUMBER OF REFUSALS AND SUBJECT BREATH TESTS IN EACH RANGE SINCE LAST MAINTENANCE REPORT							
REFUSALS	.00-.04	.05-.09	.10-.14	.15-.19	OVER .19		
0	0	0	0	0	0	0	0

List any new parts and describe any alteration or modification that was made to restore the instrument to operate satisfactorily and within established limits (use other side if necessary).

TESTED AND CERTIFIED.

INSPECTING OFFICER

SIGNATURE <i>PC R 5396</i>	PRINT NAME SHAWN DAVIS
-------------------------------	---------------------------

TYPE II PERMIT NUMBER 290088	EXPIRATION DATE 04/22/2021	TELEPHONE NUMBER 816-234-5000
---------------------------------	-------------------------------	----------------------------------



STATE OF MISSOURI
DEPARTMENT OF HEALTH AND SENIOR SERVICES
BREATH ALCOHOL PROGRAM

PERMIT
TYPE II

SHAWN DAVIS

is hereby authorized to instruct and supervise operators, train instructors, inspect, calibrate, perform field service and repairs, and operate the following breath analyzer(s):

ALCO-SENSOR IV WITH PRINTER, INTOXILYZER 8000

for the determination of the alcoholic content of blood from a sample of expired air. Permit issued under the provisions of sections 577.020 through 577.041, RSMo and 306.111 through 306.119 RSMo.

DATE 4/22/2019

NUMBER 290088

EXPIRES 4/22/2021

MO 580-0771 (6-10)

DIRECTOR OF STATE PUBLIC HEALTH LABORATORY

DIRECTOR OF DEPARTMENT OF HEALTH AND SENIOR SERVICES

LAB-4 (RS-10)

STATE OF MISSOURI
DEPARTMENT OF HEALTH AND SENIOR SERVICES
BREATH ALCOHOL PROGRAM

INSTRUMENT OPERATOR CARD

The name of the operator is authorized to operate an evidential breath alcohol instrument for the determination of the alcoholic content in breath from all expired air in Missouri.

Operator Name: **DAVIS, SHAWN**
Permit No. **42202019** Date Expires **4/22/2021**

2



7 Eastgate Dr. • P.O. Box 790 • Jackson, IL 62651-0790
217-245-2183 • Fax: 217-243-7634 • www.ilmo-products.com

Certificate of Analysis

Certificate ID: 11721

Part #: BAC105L088T

Cylinder Size: 105L

Lot Number: 88919889A1

Expiration: 3/5/2021

0.080 BAC (For the calibration of instruments used to determine breath alcohol concentration)

Contents: 105 Liters @ 1000 psig 70°F (21°C)

Component	Reported Concentration:	Accuracy (U, I=2):	Analytical Method:	Distributed by:
Ethanol	288 ppm	+/- 0.232 BAC (0.210)	NDIR	CMI Inc
Nitrogen	Balance	(5.2 ppm)		316 East Ninth Street Owensboro, KY 42303 Phone 866-835-0690 www.alcoholtest.com

*NIST Traceable Reference Material
Cylinder No. CCZ74523 / Job No. 09160306
Certified 362.7 µmol/mol Ethanol in Nitrogen

Store in dry area, away from sources of heat, ignition and direct sunlight. Do not allow storage area to exceed 52 °C (125 °F).

[Signature]
Specialty Gas Lab Tech

01-24-19
Date



The calibration results within this certificate were obtained using equipment and methods capable of producing analytical results traceable to NIST, and apply only to the items contained on this certificate. ILMO Products Company neither warrants nor represents to the user that the information provided for any particular purpose. The information use is at the sole discretion and risk of the user. Liability shall be limited to established replacement costs of job material or service.

ISO/IEC 17025:2005 Accredited Laboratory