

RECEIVED

By Tracy Crews at 11:33 am, Nov 09, 2020



MISSOURI DEPARTMENT OF HEALTH AND SENIOR SERVICES
STATE PUBLIC HEALTH LABORATORY
CMI INTOXILYZER 8000 MAINTENANCE REPORT

REPORT #2

Complete this report in duplicate at the time of the regular monthly preventive maintenance check, and whenever instrument is repaired. Send one copy to Department of Health and Senior Services, and retain one copy in department file.

INSTRUMENT SERIAL NUMBER 80-005841	LOCATION OF INSTRUMENT KCMO POLICE DEPT.	DATE OF INSPECTION 11/02/2020	TIME OF INSPECTION 23:41
---------------------------------------	---	----------------------------------	-----------------------------

CALIBRATION CHECK RESULTS			CALIBRATION CHECK SUMMARY		
Test	g/210L	Time	STANDARD TYPE	STANDARD LOT #	STANDARD EXPIRATION DATE
Air Blank	0.000	23:43	DRY	14020080A2	07/05/2022
Cal Check	0.081	23:44	SIM TEMPERATURE	SIM SERIAL NUMBER	SIM CERTIFICATE EXPIRATION
Air Blank	0.000	23:44	N/A	N/A	N/A
Cal Check	0.081	23:44	STANDARD VALUE	STANDARD SUPPLIER	
Air Blank	0.000	23:45	0.080	CMI	
Cal Check	0.081	23:45	CALIBRATION CHECK RESULT 1		
Air Blank	0.000	23:45	0.081		
Cal Check	0.081	23:45	CALIBRATION CHECK RESULT 2		
Air Blank	0.000	23:46	0.081		
Pass			CALIBRATION CHECK RESULT 3		
			0.081		
			MAXIMUM DEVIATION (MUST BE WITHIN 5%)	SPREAD (MUST BE .005 OR LESS)	
			1.2%	0.000	

DIAGNOSTIC TEST RESULTS			RFI TEST RESULTS		
Test	Pass/Fail	Time	Test	g/210L	Time
Voltage/Current Test	Pass		Air Blank	RFI*	23:46
RAM Test	Pass		Air Blank	0.000	23:46
EEPROM Checksum Test	Pass		*RFI Detect		
Real Time Clock Test	Pass		Pass		
DSP Test	Pass				
Analytical Stability Test	Pass				
Modem Test	Pass				
Temperature Regulation Test	Pass				

NUMBER OF REFUSALS AND SUBJECT BREATH TESTS IN EACH RANGE SINCE LAST MAINTENANCE REPORT							
REFUSALS	.00-.04	.05-.09	.10-.14	.15-.19	OVER .19		
0	0	0	0	0	0		0

List any new parts and describe any alteration or modification that was made to restore the instrument to operate satisfactorily and within established limits (use other side if necessary).
TESTED AND CERTIFIED 11-02-2020

INSPECTING OFFICER

SIGNATURE <i>P.A. J...</i>	PRINT NAME JORDAN INFRANCA	
TYPE II PERMIT NUMBER 290130	EXPIRATION DATE 06/21/2021	TELEPHONE NUMBER 816-482-8195

STATE OF MISSOURI
DEPARTMENT OF HEALTH AND SENIOR SERVICES
BREATH ALCOHOL PROGRAM

2

PERMIT
TYPE II

JORDAN INFRANCA

is hereby authorized to install and operate, maintain, repair, calibrate, and perform field service and repairs, and operate the following instrument(s):

ALCO-SENSOR IV WITH PERMUTER, INTOXILYZER 8000

for the determination of the alcoholic content of blood from samples of expired air. Permit issued under the provisions of sections 577.020 through 577.045, RSMo and 601-011 through 601-013, RSMo.

DATE: 6/15/2020
NUMBER: 200150
Expires: 6/15/2021
Signature: [Signature]
Commissioner of Department of Health and Senior Services
Date Issued: [Date]



7 Eastgate Dr. • P.O. Box 790 • Jacksonville, IL 62651-0790
217-245-2183 • Fax: 217-243-7634 • www.ilmo-products.com

Certificate of Analysis

Certificate ID: 13021
Part #: BAC105L080T
Vial Size: 105L
Lot Number: 14023080A2
Expiration: 7/5/2022

0.080 BAC (For the calibration of instruments used to determine breath alcohol concentration)

Contents: 105 Liters @ 1000 psig 70°F (21°C)

Component	Reported Concentration:	Analytical Accuracy (U, %±):	Analytical Method:	Distributed by:
Ethanol	208 ppm	+/-0.002 BAC(6/210L)	NDIR	CMJ Inc.
Nitrogen	Balance	[5.2 ppm]		316 East Ninth Street Owensboro, KY 42303 Phone 866-835-0690 www.alcoholtest.com

*Traceable to:
Certified Reference Material - 262.4 µmol/mol
Ethanol in Nitrogen - Serial No. GN0015026 Lot No. 050319E11

Store in dry area, away from sources of heat, ignition and direct sunlight. Do not allow storage area to exceed 52 °C (125 °F).

Specialty Gas Lab Tech
[Signature]
[Signature]

06-15-2020
Issuance Date



The calibration results within this certificate were obtained using equipment and standards capable of producing analytical results traceable to NIST, and apply only to the items contained on this certificate. ILMO Products Company makes no warranty or representation as to the suitability of the use of any information provided for any particular purpose. The information use is at the sole discretion and risk of the user. Liability shall be limited to established replacement cost of this material or service.

ISO/IEC 17025:2017 Accredited Laboratory