

RECEIVED

By Tracy Crews at 8:14 am, Aug 27, 2020

MISSOURI DEPARTMENT OF HEALTH AND SENIOR SERVICES
STATE PUBLIC HEALTH LABORATORY**CMI INTOXILYZER 8000 MAINTENANCE REPORT**

REPORT #2

Complete this report in duplicate at the time of the regular monthly preventive maintenance check, and whenever instrument is repaired. Send one copy to Department of Health and Senior Services, and retain one copy in department file.

INSTRUMENT SERIAL NUMBER 80-005843		LOCATION OF INSTRUMENT KANSAS CITY P.D.		DATE OF INSPECTION 08/11/2020		TIME OF INSPECTION 19:08				
CALIBRATION CHECK RESULTS				CALIBRATION CHECK SUMMARY						
Test ----- Air Blank Cal Check Air Blank Cal Check Air Blank Cal Check Air Blank	g/210L 0.000 0.076 0.000 0.077 0.000 0.077 0.000	Time 19:10 19:10 19:11 19:11 19:11 19:12 19:12	STANDARD TYPE DRY		STANDARD LOT # 14020080A2		STANDARD EXPIRATION DATE 07/05/2022			
			SIM TEMPERATURE N/A		SIM SERIAL NUMBER N/A		SIM CERTIFICATE EXPIRATION N/A			
			STANDARD VALUE 0.080		STANDARD SUPPLIER CMI					
			CALIBRATION CHECK RESULT 1 0.076							
			CALIBRATION CHECK RESULT 2 0.077							
			CALIBRATION CHECK RESULT 3 0.077							
			MAXIMUM DEVIATION (MUST BE WITHIN 5%) 5.0%				SPREAD (MUST BE .005 OR LESS) 0.001			
			Pass							
DIAGNOSTIC TEST RESULTS				RFI TEST RESULTS						
Voltage/Current Test Pass				Test						
RAM Test Pass				g/210L						
EEPROM Checksum Test Pass				Time						
Real Time Clock Test Pass				-----						
DSP Test Pass				Air Blank 0.000 19:13						
Analytical Stability Test Pass				Subject Test RFI* 19:13						
Modem Test Pass				Air Blank 0.000 19:13						
Temperature Regulation Test Pass				*RFI Detect						
Pass				Pass						
NUMBER OF REFUSALS AND SUBJECT BREATH TESTS IN EACH RANGE SINCE LAST MAINTENANCE REPORT										
REFUSALS	.00-.04	.05-.09	.10-.14	.15-.19	OVER .19					
0	0	0	0	0	0	0				
List any new parts and describe any alteration or modification that was made to restore the instrument to operate satisfactorily and within established limits (use other side if necessary). TESTED AND CERTIFIED										
INSPECTING OFFICER										
SIGNATURE 				PRINT NAME NATHAN MAGERS						
TYPE II PERMIT NUMBER 290118		EXPIRATION DATE 05/29/2021		TELEPHONE NUMBER 8162345000						

