



MISSOURI DEPARTMENT OF HEALTH AND SENIOR SERVICES  
STATE PUBLIC HEALTH LABORATORY

**CMI INTOXILYZER 8000 MAINTENANCE REPORT**

**RECEIVED**

By Tracy Crews at 8:03 am, Jul 22, 2020

REPORT #2

Complete this report in duplicate at the time of the regular monthly preventive maintenance check, and whenever instrument is repaired. Send one copy to Department of Health and Senior Services, and retain one copy in department file.

INSTRUMENT SERIAL NUMBER 80-005845	LOCATION OF INSTRUMENT KANSAS CITY PD	DATE OF INSPECTION 07/10/2020	TIME OF INSPECTION 19:28
---------------------------------------	--	----------------------------------	-----------------------------

CALIBRATION CHECK RESULTS			CALIBRATION CHECK SUMMARY		
Test	g/210L	Time	STANDARD TYPE DRY	STANDARD LOT # 00919080A1	STANDARD EXPIRATION DATE 03/05/2021
Air Blank	0.000	19:29	SIM TEMPERATURE N/A	SIM SERIAL NUMBER N/A	SIM CERTIFICATE EXPIRATION N/A
Cal Check	0.080	19:30	STANDARD VALUE 0.080	STANDARD SUPPLIER CMI	
Air Blank	0.000	19:30	CALIBRATION CHECK RESULT 1 0.080		
Cal Check	0.079	19:31	CALIBRATION CHECK RESULT 2 0.079		
Air Blank	0.000	19:31	CALIBRATION CHECK RESULT 3 0.080		
Cal Check	0.080	19:31	MAXIMUM DEVIATION (MUST BE WITHIN 5%) 1.2%		
Air Blank	0.000	19:32	SPREAD (MUST BE .005 OR LESS) 0.001		

**Pass**

DIAGNOSTIC TEST RESULTS			RFI TEST RESULTS		
Voltage/Current Test	Pass		Test	g/210L	Time
RAM Test	Pass		-----		
EEPROM Checksum Test	Pass		Air Blank	0.000	19:33
Real Time Clock Test	Pass		Subject Test	RFI*	19:33
DSP Test	Pass		Air Blank	0.000	19:33
Analytical Stability Test	Pass		*RFI Detect		
Modem Test	Pass		<b>Pass</b>		
Temperature Regulation Test	Pass				

**Pass**

NUMBER OF REFUSALS AND SUBJECT BREATH TESTS IN EACH RANGE SINCE LAST MAINTENANCE REPORT						
REFUSALS	.00-.04	.05-.09	.10-.14	.15-.19	OVER .19	
0	0	0	0	0	0	0

List any new parts and describe any alteration or modification that was made to restore the instrument to operate satisfactorily and within established limits (use other side if necessary).

TESTED AND CERTIFIED

<b>INSPECTING OFFICER</b>		
SIGNATURE 	PRINT NAME DOUGLAS DAVIDSON	
TYPE II PERMIT NUMBER 290087	EXPIRATION DATE 04/22/2021	TELEPHONE NUMBER 8162345000



STATE OF MISSOURI  
DEPARTMENT OF HEALTH AND SENIOR SERVICES  
BREATH ALCOHOL PROGRAM

2

PERMIT  
TYPE II

DOUGLAS DAVIDSON

is hereby authorized to instruct and supervise operators, train instructors, inspect, calibrate, perform field service and repairs, and operate the following breath analyzer(s):

**ALCO-SENSOR IV WITH PRINTER, INTOXILYZER 8000**  
for the determination of the alcoholic content of blood from a sample of expired air. Permit issued under the provisions of sections 577.020 through 577.041, RSMo and 306.111 through 306.118 RSMo.

DATE 4/22/2019

NUMBER 290087

EXPIRES 4/22/2021

VO 89-071 (6-19)

*WSD*

DIRECTOR OF STATE PUBLIC HEALTH LABORATORY

*[Signature]*

DIRECTOR OF DEPARTMENT OF HEALTH AND SENIOR SERVICES


LOG-2 (6-19)

STATE OF MISSOURI  
DEPARTMENT OF HEALTH AND SENIOR SERVICES  
BREATH ALCOHOL PROGRAM

**INSTRUMENT OPERATOR CARD**

The named operator is authorized to operate an alcohol breath alcohol instrument for the determination of the alcoholic content in breath form of expired air.

Operator: **DAVIDSON, DOUGLAS**  
Permit No. **290087**  
Date Issued **4/22/2019** Date Expires **4/22/2021**



7 Eastgate Dr. • P.O. Box 790 • Jacksonville, IL 62651-0790  
217-245-2183 • Fax: 217-243-7634 • www.ilmo-products.com

specialty gases

Certificate of Analysis

Certificate ID: 11721

Part #: BAC105L080T

Cylinder Size: 105L

Lot Number: 00919080A1

Expiration: 3/5/2021

0.080 BAC (for the calibration of instruments used to determine breath alcohol concentration)

Contents: 105 Liters @ 1000 psig 70°F (21°C)

Analytical

Component:	Reported Concentration:	Accuracy (U, L±2):	Analytical Method:
Ethanol	208 ppm	+/- 0.002 BAC (6/200)	NDIR
Nitrogen	Balance	[5.2 ppm]	

Distributed by:

CMJ Inc.  
316 East Ninth Street  
Owensboro, KY 42303  
Phone 866-835-0690  
www.alcoholtest.com

\*NIST Traceable Reference Material  
Cylinder No. CCZ74523 / Job No. 09160306  
Certified 362.2 micromol Ethanol in Nitrogen

Store in dry area, away from sources of heat, ignition and direct sunlight. Do not allow storage area to exceed 52 °C (125 °F).

*[Signature]*  
Specialty Gas Lab Tech

01-24-19

Date



PJLA  
Calibration and Testing  
Accreditation 61185

The calibration results within this certificate were obtained using equipment and standards capable of producing analytical results traceable to NIST, and apply only to the items contained on this certificate. ILMO Products Company makes no warranty or representation as to the suitability of the use of our products for any particular purpose. The information use is at the sole discretion and risk of the user. Liability shall be limited to established replacement costs of this material or device.

ISO/IEC 17025:2005 Accredited Laboratory