

RECEIVED

By Tracy Crews at 11:33 am, Nov 09, 2020

MISSOURI DEPARTMENT OF HEALTH AND SENIOR SERVICES
STATE PUBLIC HEALTH LABORATORY**CMI INTOXILYZER 8000 MAINTENANCE REPORT**

REPORT #2

Complete this report in duplicate at the time of the regular monthly preventive maintenance check, and whenever instrument is repaired. Send one copy to Department of Health and Senior Services, and retain one copy in department file.

INSTRUMENT SERIAL NUMBER 80-005845	LOCATION OF INSTRUMENT KANSAS CITY PD	DATE OF INSPECTION 11/02/2020	TIME OF INSPECTION 23:53
---------------------------------------	--	----------------------------------	-----------------------------

CALIBRATION CHECK RESULTS			CALIBRATION CHECK SUMMARY		
Test	g/210L	Time	STANDARD TYPE	STANDARD LOT #	STANDARD EXPIRATION DATE
Air Blank	0.000	23:55	DRY	14020080A2	07/05/2022
Cal Check	0.081	23:55	SIM TEMPERATURE	SIM SERIAL NUMBER	SIM CERTIFICATE EXPIRATION
Air Blank	0.000	23:56	N/A	N/A	N/A
Cal Check	0.081	23:56	STANDARD VALUE	STANDARD SUPPLIER	
Air Blank	0.000	23:57	0.080	CMI	
Cal Check	0.081	23:57	CALIBRATION CHECK RESULT 1		
Air Blank	0.000	23:58	0.081		
Cal Check	0.081	23:58	CALIBRATION CHECK RESULT 2		
Air Blank	0.000	23:58	0.081		
Cal Check	0.081	23:58	CALIBRATION CHECK RESULT 3		
Air Blank	0.000	23:58	0.081		
Pass			MAXIMUM DEVIATION (MUST BE WITHIN 5%)	SPREAD (MUST BE .005 OR LESS)	
			1.2%	0.000	

DIAGNOSTIC TEST RESULTS			RFI TEST RESULTS		
Test	Pass		Test	g/210L	Time
Voltage/Current Test	Pass		Air Blank	RFI*	23:58
RAM Test	Pass		Air Blank	0.000	23:58
EEPROM Checksum Test	Pass		*RFI Detect		
Real Time Clock Test	Pass				
DSP Test	Pass				
Analytical Stability Test	Pass				
Modem Test	Pass				
Temperature Regulation Test	Pass				
Pass			Pass		

NUMBER OF REFUSALS AND SUBJECT BREATH TESTS IN EACH RANGE SINCE LAST MAINTENANCE REPORT						
REFUSALS	.00-.04	.05-.09	.10-.14	.15-.19	OVER .19	
0	3	2	1	3	6	

List any new parts and describe any alteration or modification that was made to restore the instrument to operate satisfactorily and within established limits (use other side if necessary).

TESTED AND CERTIFIED

INSPECTING OFFICER		
SIGNATURE <i>P.O. J. Franca</i>	PRINT NAME JORDAN INFRANCA	
TYPE II PERMIT NUMBER 290130	EXPIRATION DATE 06/21/2021	TELEPHONE NUMBER 816-482-8195

STATE OF MISSOURI
DEPARTMENT OF HEALTH AND SENIOR SERVICES
BRENTH ALCOHOL PROGRAM

PERMIT
TYPE II

JORDAN INFRANCA

is hereby authorized to install, use, maintain, operate, repair, test, inspect, calibrate, perform field service and repair, and operate in following jurisdiction:

ALCO-SENSOR IV WITH PERMETER, INTOXILYZER 8000

(for use on public places and alcoholic content of blood from a sample of expired air. Permit issued under the provisions of sections 577.025 through 577.041, RSMo and Chapter 110, RSMo.)

DATE: 06/17/2019

NUMBER: 2019130


EXPIRES: 06/17/2021

OPERATOR:

LAK

Division of Environmental Health Department

Division of Environmental Health Department
L-10-1000



7 Eastgate Dr. • P.O. Box 790 • Jackson, IL 62651-0790
217-245-2183 • Fax: 217-249-7634 • www.ilmo-products.com

Certificate of Analysis

Certificate ID: 13021
Part #: BAC105L080T
Cylinder Size: 105L
Lot Number: 14020808A2
Expiration: 7/5/2022

0.080 BAC (For the calibration of instruments used to determine breath alcohol concentration)

Contents: 105 Liters @ 1000 psig 70°F (21°C)

Component	Reported Concentration	Analytical Accuracy (U, k=2)	Analytical Method
Ethanol	298 ppm	+/-0.002 BAC (G/218L) [5.2 ppm]	NDIR
Nitrogen	Balance		

Distributed by:
CMI Inc
316 East Ninth Street
Owensboro, KY 42303
Phone 866-835-0690
www.alcoholtest.com

*Traceable to:
Certified Reference Material - 2624 µmol/mol
Ethanol in Nitrogen - Serial No. GN0015026 Lot No. 050319E11
Store in dry area, away from sources of heat, ignition and direct sunlight. Do not allow storage area to exceed 52 °C (125 °F).

Specby: Gas Lab Tech
[Signature]
Issuance Date: 06-15-2020



The calibration results with this certificate were obtained using equipment and standards capable of producing analytical results traceable to NIST, and apply only to the items contained on this certificate. LMO Products Company makes no warranty or representation as to the suitability of the use of any information provided for any particular purpose. This information is at the sole discretion and risk of the user. Liability shall be limited to established replacement cost of this material or service.

ISO/IEC 17025:2017 Accredited Laboratory