



MISSOURI DEPARTMENT OF HEALTH AND SENIOR SERVICES
STATE PUBLIC HEALTH LABORATORY
CMI INTOXILYZER 8000 MAINTENANCE REPORT

REPORT #2

Complete this report in duplicate at the time of the regular monthly preventive maintenance check, and whenever instrument is repaired. Send one copy to Department of Health and Senior Services, and retain one copy in department file.

INSTRUMENT SERIAL NUMBER 80-005850	LOCATION OF INSTRUMENT LAKE WINNEBAGO PD	DATE OF INSPECTION 05/16/2020	TIME OF INSPECTION 16:30
--	--	---	------------------------------------

CALIBRATION CHECK RESULTS			CALIBRATION CHECK SUMMARY		
Test	g/210L	Time	STANDARD TYPE	STANDARD LOT #	STANDARD EXPIRATION DATE
Air Blank	0.000	16:32	DRY	AG000606	01/06/2022
Cal Check	0.082	16:33	SIM TEMPERATURE	SIM SERIAL NUMBER	SIM CERTIFICATE EXPIRATION
Air Blank	0.000	16:33	N/A	N/A	N/A
Cal Check	0.081	16:33	STANDARD VALUE	STANDARD SUPPLIER	
Air Blank	0.000	16:34	0.080	INTOXIMETERS	
Cal Check	0.081	16:34	CALIBRATION CHECK RESULT 1		
Air Blank	0.000	16:34	0.082		
Cal Check	0.081	16:34	CALIBRATION CHECK RESULT 2		
Air Blank	0.000	16:35	0.081		
Pass			CALIBRATION CHECK RESULT 3		
			0.081		
			MAXIMUM DEVIATION (MUST BE WITHIN 5%)	SPREAD (MUST BE .005 OR LESS)	
			2.5%	0.001	

DIAGNOSTIC TEST RESULTS			RFI TEST RESULTS		
			Test	g/210L	Time
Voltage/Current Test	Pass		Air Blank	RFI*	16:35
RAM Test	Pass		Air Blank	0.000	16:36
EEPROM Checksum Test	Pass		*RFI Detect		
Real Time Clock Test	Pass		Pass		
DSP Test	Pass				
Analytical Stability Test	Pass				
Modem Test	Pass				
Temperature Regulation Test	Pass				

NUMBER OF REFUSALS AND SUBJECT BREATH TESTS IN EACH RANGE SINCE LAST MAINTENANCE REPORT						
REFUSALS	.00-.04	.05-.09	.10-.14	.15-.19	OVER .19	
0	0	0	0	0	0	0

List any new parts and describe any alteration or modification that was made to restore the instrument to operate satisfactorily and within established limits (use other side if necessary).

INSPECTING OFFICER

SIGNATURE 	PRINT NAME MCGINNESS, JASON
TYPE II PERMIT NUMBER 290184	EXPIRATION DATE 08/19/2021
	TELEPHONE NUMBER 816-537-7900



STATE OF MISSOURI
DEPARTMENT OF HEALTH AND SENIOR SERVICES
BREATH ALCOHOL PROGRAM

2

PERMIT
TYPE II
JASON M MCGINNESS

is hereby authorized to instruct and supervise operators, train instructors, inspect, calibrate, perform field service and repairs, and operate the following breath analyzer(s):

INTOXILYZER 8000

for the determination of the alcoholic content of blood from a sample of expired air. Permit issued under the provisions of sections 577.020 through 577.041, RSMo and 306.111 through 306.119 RSMo.

DATE 8/19/2019

NUMBER 290184

EXPIRES 8/19/2021

DIRECTOR OF STATE PUBLIC HEALTH LABORATORY

DIRECTOR OF DEPARTMENT OF HEALTH AND SENIOR SERVICES

MO 580-0771 (6-10)

LAB-4 (R6-10)



STATE OF MISSOURI
DEPARTMENT OF HEALTH AND SENIOR SERVICES
BREATH ALCOHOL PROGRAM

INSTRUMENT OPERATOR CARD

The named cardholder is authorized to operate an evidential breath alcohol instrument for the determination of the alcoholic content in breath form of expired air in Missouri.

Operator MCGINNESS, JASON
Permit No 290184
Date Issued 8/19/2019 **Date Expires** 8/19/2021

